



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4285
CALIBRATION DATE: 22-Feb-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -4.722137e-05
a1 = 2.801537e-04
a2 = -2.678737e-06
a3 = 1.618354e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	736364.3	1.0000	-0.0000
4.5000	630344.8	4.5000	0.0000
15.0000	403082.3	15.0000	-0.0000
18.5000	349396.1	18.5000	-0.0000
24.0000	280744.2	24.0001	0.0001
29.0000	231493.3	29.0000	-0.0000
32.5000	202914.5	32.5000	-0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln(n)]^2 + a_3[\ln(n)]^3\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

