



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4285
CALIBRATION DATE: 31-Mar-26

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 6.210360e-06
a1 = 2.678931e-04
a2 = -1.743056e-06
a3 = 1.380907e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
0.9999	736312.6	0.9999	-0.0000
4.5000	630306.9	4.5000	0.0000
15.0000	403074.1	15.0000	0.0000
18.5000	349393.7	18.5000	-0.0000
24.0000	280745.3	24.0000	0.0000
29.0000	231493.3	29.0000	-0.0000
32.5000	202912.5	32.5000	0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

