

ECO Calibration and Repairs

Diagnosis: Evaluated instrument (**FLSB-2694**). It was found that the unit required calibration.

Repairs and Modifications: The unit was calibrated, and standard service was performed. Unit returned with customer settings. [4-2 LP]

Comments: Added new 9v batteries, faceplate and shutter. New device files and characterization sheets included.

ECO Standard Service:

The instrument's bulkhead connector, pressure housing and optical head are inspected for damage. The optics are inspected for any crack and/ or delamination. The instrument is also evaluated to determine proper functionality.

If applicable, the head is inspected for cracks in motor bores. Case seals, shaft, shaft seal, faceplate, wiper, desiccant pack, O-rings and batteries (if equipped) are replaced.

Incoming settings and memory are collected if incoming conditions allow. When requested, a pre-service characterization can be performed. Then, the final calibration and characterization are conducted. Noise, stability, and live pressure test are also performed before service is completed.

A device file, repair sheet, and new characterization sheets are provided to customers via USB.