



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4425
CALIBRATION DATE: 10-Apr-26

SBE 16plus CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.058887e+00
h = 1.445821e-01
i = -2.662955e-04
j = 4.208359e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2710.12	0.0000	0.00000
1.0000	34.6290	2.96148	5277.48	2.9615	0.00001
4.5001	34.6089	3.26709	5473.71	3.2671	-0.00003
14.9999	34.5666	4.24420	6057.89	4.2443	0.00006
18.5000	34.5574	4.58771	6249.95	4.5877	-0.00003
24.0000	34.5475	5.14304	6548.30	5.1430	-0.00002
29.0000	34.5420	5.66243	6815.17	5.6624	-0.00001
32.5000	34.5390	6.03308	6999.21	6.0331	0.00001

f = Instrument Output (Hz) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = (g + h * f + i * f + j * f) / (1 + δ * t + ϵ * p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

