



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4425
CALIBRATION DATE: 10-Apr-26

SBE 16plus TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 1.239921e-03
a1 = 2.602973e-04
a2 = -6.204088e-08
a3 = 1.473382e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	686173.600	1.0002	0.0002
4.5001	611978.200	4.4997	-0.0004
14.9999	425202.400	15.0004	0.0005
18.5000	374461.300	18.4997	-0.0003
24.0000	305173.500	23.9999	-0.0001
29.0000	252155.600	28.9999	-0.0001
32.5000	220021.700	32.5001	0.0001

n = Instrument Output (counts)

MV = (n - 524288) / 1.6e+007

R = (MV * 2.900e+009 + 1.024e+008) / (2.048e+004 - MV * 2.0e+005)

Temperature ITS-90 (°C) = 1 / {a0 + a1[ln(R)] + a2[ln(R)] + a3[ln(R)]} - 273.15

Residual (°C) = instrument temperature - bath temperature

