



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4426
CALIBRATION DATE: 14-Apr-26

SBE 16plus CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.026676e+00
h = 1.447403e-01
i = -3.163562e-04
j = 4.474544e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2668.16	0.0000	0.00000
0.9999	34.6728	2.96486	5259.16	2.9649	0.00000
4.5000	34.6533	3.27086	5456.44	3.2709	-0.00000
15.0000	34.6123	4.24923	6043.37	4.2492	-0.00000
18.4999	34.6038	4.59320	6236.33	4.5932	0.00001
24.0000	34.5945	5.14927	6535.90	5.1493	-0.00000
29.0000	34.5896	5.66935	6803.82	5.6694	0.00000
32.5001	34.5866	6.04046	6988.49	6.0404	-0.00004

f = Instrument Output (Hz) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = (g + h * f + i * f + j * f) / (1 + δ * t + ϵ * p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

