



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4426
CALIBRATION DATE: 14-Apr-26

SBE 16plus TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 1.244795e-03
a1 = 2.599558e-04
a2 = 1.403585e-07
a3 = 1.392223e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
0.9999	666338.600	0.9999	0.0000
4.5000	593536.300	4.4999	-0.0001
15.0000	411090.900	15.0001	0.0001
18.4999	361697.600	18.4999	0.0000
24.0000	294378.727	23.9998	-0.0002
29.0000	242940.300	29.0002	0.0002
32.5001	211799.636	32.5001	-0.0000

n = Instrument Output (counts)

MV = (n - 524288) / 1.6e+007

R = (MV * 2.900e+009 + 1.024e+008) / (2.048e+004 - MV * 2.0e+005)

Temperature ITS-90 (°C) = 1 / {a0 + a1[ln(R)] + a2[ln(R)] + a3[ln(R)]} - 273.15

Residual (°C) = instrument temperature - bath temperature

