



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 50217
CALIBRATION DATE: 10-Apr-26

SBE 16plus V2 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 1.239731e-03
a1 = 2.804317e-04
a2 = -1.768045e-06
a3 = 2.050119e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	554698.200	1.0000	0.0000
4.5001	489416.600	4.5001	-0.0000
14.9999	330180.400	15.0000	0.0001
18.5000	288101.300	18.4999	-0.0001
24.0000	231436.300	23.9999	-0.0001
29.0000	188659.900	29.0001	0.0001
32.5000	162975.700	32.4999	-0.0001

n = Instrument Output (counts)

MV = (n - 524288) / 1.6e+007

R = (MV * 2.900e+009 + 1.024e+008) / (2.048e+004 - MV * 2.0e+005)

Temperature ITS-90 (°C) = 1 / {a0 + a1[ln(R)] + a2[ln(R)] + a3[ln(R)]} - 273.15

Residual (°C) = instrument temperature - bath temperature

