



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 50236
CALIBRATION DATE: 10-Apr-26

SBE 16plus V2 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.918258e-01
h = 1.256802e-01
i = -1.006536e-04
j = 2.377152e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2810.27	0.0000	0.00000
1.0000	34.6290	2.96148	5604.45	2.9615	0.00002
4.5001	34.6089	3.26709	5816.21	3.2671	-0.00001
14.9999	34.5666	4.24420	6446.00	4.2442	-0.00001
18.5000	34.5574	4.58771	6652.98	4.5877	0.00000
24.0000	34.5475	5.14304	6974.29	5.1431	0.00001
29.0000	34.5420	5.66243	7261.61	5.6624	0.00002
32.5000	34.5390	6.03308	7459.68	6.0331	-0.00001

f = Instrument Output (Hz) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = (g + h * f + i * f + j * f) / (1 + δ * t + ϵ * p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

