



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 50236
CALIBRATION DATE: 10-Apr-26

SBE 16plus V2 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 1.245303e-03
a1 = 2.774405e-04
a2 = -1.434905e-06
a3 = 1.954029e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	557096.800	1.0000	0.0000
4.5001	491774.000	4.5001	-0.0000
14.9999	332260.500	15.0000	0.0001
18.5000	290059.700	18.4999	-0.0001
24.0000	233193.300	24.0000	-0.0000
29.0000	190234.200	29.0001	0.0001
32.5000	164425.300	32.5000	-0.0000

n = Instrument Output (counts)

MV = (n - 524288) / 1.6e+007

R = (MV * 2.900e+009 + 1.024e+008) / (2.048e+004 - MV * 2.0e+005)

Temperature ITS-90 (°C) = 1 / {a0 + a1[ln(R)] + a2[ln(R)] + a3[ln(R)]} - 273.15

Residual (°C) = instrument temperature - bath temperature

