



SKYE INSTRUMENTS LTD.
21, DDOLE ENTERPRISE PARK,
LLANDRINDOD WELLS,
POWYS. LD1 6DF. U.K.
TEL: +44 (0) 1597 824811
email: skyemail@skyeinstruments.com
website: www.skyeinstruments.com

CALIBRATION CERTIFICATE No : TQUA/037/0121

UNIT TYPE :-	TAG-PARQ
OUTPUT :-	0.01433 μA / $\mu\text{mol m}^{-2} \text{s}^{-1}$
DATE OF CALIBRATION :-	08/01/2021
LAMP REFERENCE :-	SK4
A/D UNIT :-	1239702

Calibrated against a National Physical Laboratory UK reference standard lamp.
Uncertainty 5% (typically $< \pm 3\%$) based on an estimated confidence of not less than 95%

Calibrated By :-	G. Sims
Checked By :-	
Issue Date :-	11/01/2021

THIS UNIT IS DUE FOR
RECALIBRATION WITHIN 2 YEARS OF
THE ABOVE CALIBRATION DATE

