Mediciones Con Tacometro												Coef. de Var. de las Muestras		
Referencia (rpm)	Medicion 1	Medicion 2	Medicion 3	Medicion 4	Medicion 5	Medicion 6	Medicion 7	Medicion 8	Medicion 9	Medicion 10	CV	Referencia (rpr	n) CV	
300	295	295	294	294	296	297	295	294	295	295	0,30%	300	0,30%	
450	415	412	413	415	416	413	413	414	414	416	0,45%	450	0,45%	
200	180	178	177	177	178	179	178	179	179	178	0,50%	200	0,50%	
250	237	236	236	235	237	236	237	237	238	236	0,31%	250	0,31%	
											0,39%	Media	0,39%	
Mediciones De Ti	empo De Ciclo													
Medicion	Valor (us)													
1	52											Error de	Error de las muestras	
2	52											Referencia (rpr	n) Error (rpr	
3	52											300	-5	
4	52											450	-35,9	
5	52											200	-21,7	
6	52											250	-13,5	
7	52											Media	-19,025	
8	56													
9	52													
10	52													
11	56													
12	56													
13	52													
14	52													
15	52													
16	56													
17	52													
18	56													
19	52													
20	52													
Media	53													