Monobind Inc.
Lake Forest, CA 92630, USA

AccuBind
ELISA Microwells

AccuBind™ Estradiol (E2) ELISA
Product Code: 4925-300

Intended Use: The Quantitative Determination of Estradiol Concentration in Human Serum or Plasma by a Microplate Enzyme Immunoassay

SUMMARY AND EXPLANATION OF THE TEST

Measurement of estradiol in serum or plasma is considered to be the most reliable way to assess its rate of production.

Estradiol (17β-estradiol) is a steroid hormone (molecular weight of 272.3 daltons), which circulates predominantly protein-bound. In addition to estradiol, other natural steroidal estrogens include estrone, estriol and their metabolites. Natural estrogens are hormones secreted principally by the ovarian follicles and also by the adrenals, corpus luteum, and placenta and, in males, by the testes. Exogenous estrogens (natural or synthetic) elicit, to varying degrees, all the pharmacologic responses usually produced by endogenous estrogens.

Estrogenic hormones are secreted at varying rates during the menstrual cycle throughout the period of ovarian activity. During pregnancy, the placenta becomes the main source of estrogens. At menopause, ovarian secretion of estrogens declines at varying rates. The gonadotropins of the anterior pituitary regulate secretion of the ovarian hormones, estradiol and progesterone; hypothalamic control of pituitary gonadotropin production is in turn regulated by plasma concentrations of the estrogens and progesterone. This complex feedback system results in the cyclic phenomenon of ovulation and menstruation.

Estradiol determinations have proved of value in a variety of contexts, including the investigation of precocious puberty in girls and gynecomastia in men. Its principal uses have been in the differential diagnosis of amenorrhea and in the monitoring of ovulation induction.

This kit uses a specific anti-estradiol antibody, and does not require prior sample extraction of serum or plasma. Cross-reactivity to other naturally occurring and structurally related steroids is low.

The employment of several serum references of known estradiol concentration permits construction of a graph of activity and concentration. From comparison to the dose response curve, an unknown specimen's activity can be correlated with estradiol concentration.

PRINCIPLE

Delayed Competitive Enzyme Immunoassay (TYPE 9):

The essential reagents required for a enzyme immunoassay include antibody, enzyme-antigen conjugate and native antigen. Upon mixing the biotinylated antibody with a serum containing the antigen, a reaction results between the antigen and the antibody. The interaction is illustrated by the following equation:

$$Ag + Ab_{Btn} \rightleftharpoons AgAb_{Btn}$$

Ab_{Btn} = Biotinylated antibody
Ag = Antigen (Variable Quantity) $AgAb_{Btn}$ = Immune Complex

After a short incubation, the enzyme conjugate is added (This delayed addition permits an increase in sensitivity for low concentration samples). Upon the addition of the enzyme conjugate, competition reaction results between the enzyme analog and the antigen in the sample for a limited number of antibody binging sites (not consumed in the first incubation).

$$^{Enz}Ag + Ag + rAb_{Btn} \underset{k_{-a}}{\overset{k_a}{\rightleftharpoons}} AgAb_{Btn} + {}^{Enz}AgAb_{Btn}$$

^{Enz}Ag = Enzyme-antigen Conjugate (Constant Quantity)

$^{Enz}Ag\ Ab_{Btn}$ = Enzyme-antigen Conjugate -Antibody Complex

rAb_{Btn} = Biotinylated antibody not reacted in first incubation

k_a = Rate Constant of Association

k_{-a} = Rate Constant of Disassociation

$K = k_a / k_{-a}$ = Equilibrium Constant

A simultaneous reaction between the biotin attached to the antibody and the streptavidin immobilized on the microwell occurs. This effects the separation of the antibody bound fraction after decantation or aspiration.

$$AgAb_{Btn} + {}^{Enz}AgAb_{Btn} + \underline{Streptavidin_{CW}} \Rightarrow \underline{\text{immobilized complex}}$$

$\underline{Streptavidin_{CW}}$ = Streptavidin immobilized on well

Immobilized complex = sandwich complex bound to the solid surface

The enzyme activity in the antibody bound fraction is inversely proportional to the native antigen concentration. By utilizing several different serum references of known antigen con-centration, a dose response curve can be generated from which the antigen concentration of an unknown can be ascertained.

REAGENTS MATERIALS PROVIDED:

A. Estradiol Calibrators – 1ml/vial - Icons A-G
Seven (7) vials of serum reference for estradiol at concentrations of 0 (**A**), 20 (**B**), 100 (**C**), 250 (**D**), 500 (**E**), 1500 (**F**) and 3000 (**G**) in pg/ml. Store at 2-8°C. A preservative has been added. The calibrators can be expressed in molar concentrations (nM/L) by multiplying by 2.72.

For example: 1pg/ml x 3.67= 3.67 pM/L

B. Estradiol Enzyme Reagent – 6.0 ml/vial
One (1) vial of Estradiol (Analog)-horseradish peroxides (HRP) conjugate in a protein-stabilizing matrix red with dye. Store at 2-8°C.

C. Estradiol Biotin Reagent – 6.0 ml - Icon ∇
One (1) bottle of reagent contains anti-estradiol biotinylated purified rabbit IgG conjugate in buffer, green dye and preservative. Store at 2-8°C.

D. Streptavidin Coated Plate – 96 wells –Icon ⇓
One 96-well microplate coated with 1.0 µg/ml streptavidin and packaged in an aluminum bag with a drying agent. Store at 2-8°C.

E. Wash Solution Concentrate – 20ml - Icon
One (1) vial contains a surfactant in buffered saline. A preservative has been added. Store at 2-30°C.

F. Substrate Reagent – 12ml/vial - Icon S N
One (1) bottle contains tetramethylbenzidine (TMB) and hydrogen peroxide (H_2O_2) in buffer. Store at 2-8°C.

G. Stop Solution – 8ml/vial - Icon STOP
One (1) vial contains a strong acid(H_2SO_4). Store at 2-30°C.

H. Product Instructions.

Note 1: Do not use reagents beyond the kit expiration date.

Note 2: Opened reagents are stable for sixty (60) days when stored at 2-8°C.

Note 3: Above reagents are for a single 96-well microplate.

Required But Not Provided:

1. Pipette capable of delivering 25 µl and 50 µl with a precision of better than 1.5%.
2. Dispenser(s) for repetitive deliveries of 0.100ml and 0.350ml volumes with a precision of better than 1.5%.
3. Adjustable volume (200-1000µl) dispenser(s) for conjugate.
4. Microplate washer or a squeeze bottle (optional).
5. Microplate Reader with 450nm and 620nm wavelength absorbance capability.
6 Absorbent Paper for blotting the microplate wells.
7 Plastic wrap or microplate cover for incubation steps.
8 Vacuum aspirator (optional) for wash steps.
9 Timer.
10 Quality control materials.

PRECAUTIONS

For In Vitro Diagnostic Use
Not for Internal or External Use in Humans or Animals

All products that contain human serum have been found to be non-reactive for Hepatitis B Surface Antigen, HIV 1&2 and HCV Antibodies by FDA required tests. Since no known test can offer complete assurance that infectious agents are absent, all human serum products should be handled as potentially hazardous and capable of transmitting disease. Good laboratory procedures for handling blood products can be found in the Center for Disease Control / National Institute of Health, "Biosafety in Microbiological and Biomedical Laboratories," 2nd Edition, 1988, HHS Publication No. (CDC) 88-8395.

SPECIMEN COLLECTION AND PREPARATION

The specimens shall be blood; serum or heparanised plasma in type and taken with the usual precautions in the collection of venipuncture samples. For accurate comparison to establish normal values, a fasting morning serum sample should be obtained. The blood should be collected in a redtop (with or without gel additives) venipuncture tube(s) or for plasma use evacuated tube(s) containing heparin.. Allow the blood to clot for serum samples. Centrifuge the specimen to separate the serum or plasma from the cells.

Samples may be refrigerated at 2-8°C for a maximum period of five (5) days. If the specimen(s) cannot be assayed within this time, the sample(s) may be stored at temperatures of -20°C for up to 30 days. Avoid repetitive freezing and thawing. When assayed in duplicate, 0.050ml of the specimen is required.

REAGENT PREPARATION

1. **Wash Buffer**
Dilute contents of wash solution to 1000ml with distilled or deionized water in a suitable storage container. Diluted buffer can be stored at room temperature (20-27°C) for up to 60 days.

Note: Do not use the working substrate if it looks blue.

TEST PROCEDURE

Before proceeding with the assay, bring all reagents, serum references and controls to room temperature (20 - 27°C).

1. Format the microplates' wells for each serum reference, control and patient specimen to be assayed in duplicate. **Replace any unused microwell strips back into the aluminum bag, seal and store at 2-8ºC.**
2. Pipette 0.025 ml (25 µL) of the appropriate serum reference, control or specimen into the assigned well.
3. Add 0.050 ml (50µl) of the Estradiol Biotin Reagent to all wells.
4. Swirl the microplate gently for 20-30 seconds to mix.
5. Cover and incubate for 30 minutes at room temperature.
6. Add 0.050 ml (50µl) of Estradiol Enzyme Reagent to all wells.

Add directly on top the reagents dispensed in the wells.

7. Swirl the microplate gently for 20-30 seconds to mix.
8. Cover and incubate for 90 minutes at room temperature.
9. Discard the contents of the microplate by decantation or aspiration. If decanting, blot the plate dry with absorbent paper.
10. Add 350µl of wash buffer (see Reagent Preparation Section), decant (tap and blot) or aspirate. Repeat two (2) additional times for a total of three (3) washes. **An automatic or manual plate washer can be used. Follow the manufacturer's instruction for proper usage. If a squeeze bottle is employed, fill each well by depressing the container (avoiding air bubbles) to dispense the wash. Decant the wash and repeat two (2) additional times.**
11. Add 0.100 ml (100µl) of substrate solution to all wells (see Reagent Preparation Section). **Always add reagents in the same order to minimize reaction time differences between wells.**
DO NOT SHAKE THE PLATE AFTER SUBSTRATE ADDITION
12. Incubate at room temperature for twenty (20) minutes.
13. Add 0.050ml (50µl) of stop solution to each well and gently mix for 15-20 seconds. **Always add reagents in the same order to minimize reaction time differences between wells.**
14. Read the absorbance in each well at 450nm (using a reference wavelength of 620-630nm.. **The results should be read within thirty (30) minutes of adding the stop solution.**

Note: Dilute the samples suspected of concentrations higher than 3000pg/ml 1:5 and 1:10 with estradiol '0' pg/ml calibrator or male patient serum pools with a known low value for estradiol.

QUALITY CONTROL

Each laboratory should assay controls at levels in the low, normal and high range for monitoring assay performance. These controls should be treated as unknowns and values determined in every test procedure performed. Quality control charts should be maintained to follow the performance of the supplied reagents. Pertinent statistical methods should be employed to ascertain trends. The individual laboratory should set acceptable assay performance limits. In addition, maximum absorbance should be consistent with past experience. Significant deviation from established performance can indicate unnoticed change in experimental conditions or degradation of kit reagents. Fresh reagents should be used to determine the reason for the variations.

CALCULATION OF RESULTS

A dose response curve is used to ascertain the concentration of estradiol in unknown specimens.

1. Record the absorbance obtained from the printout of the microplate reader as outlined in Example 1.
2. Plot the absorbance for each duplicate serum reference versus the corresponding estradiol concentration in pg/ml on linear graph paper (do not average the duplicates of the serum references before plotting).
3. Connect the points with a best-fit curve.
4. To determine the concentration of estradiol for an unknown, locate the average absorbance of the duplicates for each unknown on the vertical axis of the graph, find the intersecting point on the curve, and read the concentration (in pg/ml) from the horizontal axis of the graph (the duplicates of the unknown may be averaged as indicated). In the following example, the average absorbance (1.202) intersects the dose response curve at (160pg/ml) estradiol concentration (See Figure 1).

EXAMPLE 1

Sample I.D.	Well Number	Abs (A)	Mean Abs (B)	Value (pg/ml)
Cal A	A1	2.268	2.256	0
	B1	2.244		
Cal B	C1	1.839	1.849	20
	D1	1.860		
Cal C	E1	1.409	1.426	100
	F1	1.443		
Cal D	G1	1.017	1.003	250
	H1	0.989		
Cal E	A2	0.698	0.723	500
	B2	0.748		
Cal F	C2	0.480	0.487	1500
	D2	0.493		
Cal G	E2	0.390	0.388	3000
	F2	0.385		
Pat# 1	G2	1.202	1.202	**160**
	H2	1.203		

*The above data and table below is for example only. Do not use it for calculating your results.

Figure 1



Note: Multiply the horizontal values by 1000 to convert into pg/ml.

Q.C. PARAMETERS

In order for the assay results to be considered valid the following criteria should be met:

1. The absorbance (OD) of calibrator 0 pg/ml should be ≥ 1.3.
2. Four out of six quality control pools should be within the established ranges.

RISK ANALYSIS

A. Assay Performance

1. It is important that the time of reaction in each well is held constant for reproducible results.
2. Pipetting of samples should not extend beyond ten (10) minutes to avoid assay drift.
3. If more than one (1) plate is used, it is recommended to repeat the dose response curve.
4. Addition of the substrate solution initiates a kinetic reaction, which is terminated by the addition of the stop solution. Therefore, the addition of the substrate and the stopping solution should be added in the same sequence to eliminate any time-deviation during reaction.
5. Plate readers measure vertically. Do not touch the bottom of the wells.
6. Failure to remove adhering solution adequately in the aspiration or decantation wash step(s) may result in poor replication and spurious results.
7. Use components from the same lot. No intermixing of reagents from different batches.
8. Accurate and precise pipetting, as well as following the exact time and temperature requirements prescribed are essential. Any deviation from this may yield inaccurate results.
9. It is important to calibrate all the equipment e.g. Pipettes, Readers, Washers and/or the automated instruments used for using this device.
10. Risk Analysis- as required by CE Mark IVD Directive 98/79/EC - for this and other devices, made by Monobind, can be requested via email from Monobind@monobind.com.

B. Interpretation

1. If computer controlled data reduction is used to interpret the results of the test, it is imperative that the predicted values for the calibrators fall within 10% of the assigned concentrations.

EXPECTED RANGES OF VALUES

In agreement with established reference intervals for a "normal" adult population and females during gestation the expected ranges for the Estradiol AccuBind™ ELISA Test System are detailed in Table 1.

TABLE 1
Expected Values for the Estradiol Test System

	Median	Range
Females	-	-
Follicular Phase	48	9-175
Luteal Phase	103	44-196
Periovulatory	209	107-281
Treated Menopausal	122	42-289
Untreated Menopausal	7.3	ND-20
Oral Contraceptives	13	ND-103
Males	19	4-94

During pregnancy the Estradiol serum levels rise rapidly till the end of third trimester [17].

It is important to keep in mind that establishment of a range of values which can be expected to be found by a given method for a population of "normal" persons is dependent upon a multiplicity of factors: the specificity of the method, the population tested and the precision of the method in the hands of the analyst. For these reasons each laboratory should depend upon the range of expected values established by the manufacturer only until an in-house range can be determined by the analysts using the method with a population indigenous to the area in which the laboratory is located.

PERFORMANCE CHARACTERISTICS

A. Precision

The within and between assay precision of the estradiol AccuBind™ Microplate EIA Test System were determined by analyses on three different levels of pool control sera. The number, mean values, standard deviation and coefficient of variation for each of these control sera are presented in Table 2 and Table 3.

TABLE 2
Within Assay Precision (Values in pg/ml)

Sample	N	X	σ	C.V.
Low	20	85.9	7.6	8.8%
Normal	20	260.5	20.3	7.8%
High	20	495.3	33.7	6.8%

TABLE 3
Between Assay Precision (Values in pg/ml)

Sample	N	X	σ	C.V.
Low	10	89.3	8.2	9.2%
Normal	10	245.5	23.7	9.7%
High	10	467.2	38.3	8.2%

*As measured in ten experiments in duplicate over a ten day period.

B. Accuracy

The Estradiol AccuBind™ Microplate ELISA Test System was compared with a chemiluminescence immunoassay method. Biological specimens from low, normal and relatively high estradiol level populations were used (The values ranged from 10 pg/ml – 4300 pg/ml). The total number of such specimens was 65. The least square regression equation and the correlation coefficient were computed for this estradiol EIA in comparison with the reference method. The data obtained is displayed in Table 4.

TABLE 4

Method	Mean (x)	Least Square Regression Analysis	Correlation Coefficient
This Method (Y)	139	Y= 10+0.985*(X)	0.979
Reference (X)	148		

Only slight amounts of bias between this method and the reference method are indicated by the closeness of the mean values. The least square regression equation and correlation coefficient indicates excellent method agreement.

C. Sensitivity

The estradiol AccuBind™ Microplate EIA Test System has a sensitivity of 6.5 pg/ml. The sensitivity was ascertained by determining the variability of the 0 pg/ml serum calibrator and using the 2σ (95% certainty) statistic to calculate the minimum dose.

D. Specificity

The % cross reactivity of the estradiol antibody to selected substances was evaluated by adding the interfering substance to a serum matrix at various concentrations. The cross-reactivity was calculated by deriving a ratio between dose of interfering substance to dose of estradiol needed to displace the same amount of labeled analog.

Substance	Cross Reactivity
Androstenedione	0.0003
Dihydotestosterone	0.0008
Cortisone	<0.0001
Corticosterone	<0.0001
Cortisol	0.0004
Estriol	<0.0001
DHEA sulfate	<0.0001
Estradiol	<0.0001
Estrone	<0.0001
Testosterone	<0.0001

REFERENCES

1. Abraham GE. The application of natural steroid radioimmunoassay to gynecologic endocrinology. In: Abraham GE, editor. Radioassay Systems in Clinical Endocrinology, Basel: Marcel Dekker,: 475-529 (1981).
2. Batzer F,"Hormonal evaluation of early pregnancy", *Fertility Sterility*, **34**:1-13 (1980).
3. Bauman J, "Basal body temperature: unreliable method of ovulation detection", *Fertility Sterility*, **36:729**-33, (1981).
4. Bergquist C, Nillius SJ, Wide L "Human gonadotropin therapy: Serum estradiol and progesterone patterns during conceptual cycles", *Fertility Sterility*; **39:**761-65 (1983).
5. Gautray JP, et al, "Clinical investigation of the menstrual cycle: clinical, endometrial and endocrine aspects of luteal defects", *Fertility Sterility*, **35**:296-303 (1981).
6. Hensleigh PA, Fainstat T, "Corpus luteum dysfunction: serum progesterone levels in diagnosis and assessment of therapy for recurrent and threatened abortion", *Fertility Sterility*, **32**:396-9. (1979)
7. Hernandez JL, et al,"Direct evidence of luteal insufficiency in women with habitual abortion", *Obstetric Gynecology* , 49:705-8.(1977).
8. Goldstein D, et al, "Correlation between Estradiol and Progesterone in cycles with luteal phase deficiency", *Fertility Sterility*, **37:**348-54 (1982).
9. Klopper A, Fuchs F. Progestagens. In: Fuchs F, Klopper A, editors. Endocrinology of Pregnancy. Hagerstown: Harper & Row, 99-122 (1977).
10. Lehmann F, Bettendorf G,"The endocrine shift from a normal cycle to anovulation", Insler V, Bettendorf G, editors. Advances in Diagnosis and Treatment of Infertility. Amsterdam: Elseveir/North Holland, 105 -13 (1981).
11. March CM. Luteal phase defects. In: Mishell DR, Davajan V, editors. Reproductive Endocrinology, Infertility and Contraception. Philadelphia: F. A. Davis Company, 1979: 469-76.
12. March CM, Goebelsmann U, Nakamura RM, Mishell Dr. Roles of estradiol and progesterone in eliciting the midcycle luteinizing hormone and follicle stimulating hormone surges. *J Clin Endocrinol Metab*, **49**:507-13 (1979).
13. BIO-ED slide/seminar educational program. Rochester: Bioeducational Publications, 1981.
14. Radwanska E, et al,"Plasma progesterone levels in normal and abnormal early human pregnancy", *Fertility Sterility* 30:398-402 (1978).
15. Tietz, Textbook of clinical chemistry,2nd ed. Philadelphia: W.B. Saunders, (1994).

Revision: 1 **Date: 070709** **DCO: 0153**
Cat #: 4925-*300*

Size		96(A)	192(B)
Reagent (fill)	A)	1ml set	1ml set
	B)	1 (6ml)	2 (6ml)
	C)	1 (6ml)	2 (6ml)
	D)	1 plate	2 plates
	E)	1 (20ml)	1 (20ml)
	F)	1 (12ml)	2 (12ml)
	G)	1 (8ml)	2 (8ml)

For Orders and Inquiries, please contact

Monobind Inc.
100 North Pointe Drive
Lake Forest, CA 92630 USA

Tel: 949-951-2665 Email: info@monobind.com
Fax: 949-951-3539 Web: www.monobind.com

Please visit our website to learn more about our other interesting products and services.

IVD CE 2° 8°

EC REP **CEpartner4U**, 3951 DB; 13.NL
Tel: +31 (0) 6-516.536.26

Instruments & Applications

Monobind's immunoassay products are designed to work in both manual and automated lab environments. AccuBind™ and AccuLite™ are compatible with any open-ended instrumentation, including chemistry analyzers, microplate readers and microplate washers. There may or may not be an application developed for your particular instrument, please visit the instrument section of our website, or contact techsupport@monobind.com

Monobind offers several instruments, including the Impulse 2 Luminometer CLIA Plate Reader designed hand-in-hand with our products and capable of 2-point calibration. Visit our website for more information.