SpectralForge Studio: A Physically Comprehensive Multi-Excitation Fluorescence Hyperspectral Simulator and Experiment-Design Toolkit for Layered Tissue

Narek Meloyan and Narine Sarvazyan

***Abstract*—Multi-excitation fluorescence hyperspectral imaging (ME-HSI) records an excitation–emission matrix at every pixel, but the signal of a buried fluorophore is distorted by the tissue above it, and real acquisitions carry no ground truth. We present SpectralForge Studio, an open simulator and experiment-design environment for ME-HSI of layered tissue. It combines a 3D scene editor, a forward model (ray optics, dye photophysics, a layered-block 3D Monte Carlo with voxel inclusions, a tissue-optics library, a tunable-filter camera chain) and a research toolkit: sensitivity analysis, unmixing, maximum-likelihood retrieval, Fisher-information analysis with a model-error term and channel selection, and calibration of a digital twin. The transport engine reproduces the MCML benchmark to $0.1\,\%$, conserves energy to 10^{-9} and matches an independent Monte Carlo for radial reflectance to $1\,\%$. A calibrated twin predicts held-out excitation lines of measured collagen spectra with a mean spectral angle of 8.4° (nominal instrument 14.2°), though not better than a copy of a measured spectrum (7.7°). In simulations of collagen-covered tissue, the buried signal decays with an effective depth of 0.07–0.36 mm for 340–430 nm excitation; cover parameters carry a first-order share of 0.66 of the optical redox-ratio variance against 0.16 for NADH and FAD; about 30 of 248 channels come within $10\,\%$ of the full-protocol Cramér–Rao bound, yet thickness retrieval is limited by model error (RMSE 160–260 µm), not photon noise; unmixing with a thin-cover library fails from 0.1 mm depth; an absorbing vessel under a dense collagen cover is detectable per pixel only to 0.24–0.44 mm at 340 nm and an endogenous-contrast nodule not at all; and the engine's per-pixel noise follows $N^{-1/2}$ with bit-identical output across thread counts. The software is open source.**

***Index Terms*—Fluorescence hyperspectral imaging, Monte Carlo light transport, tissue optics, simulation, experiment design, Fisher information, sensitivity analysis, digital twin.**

I. Introduction

MULTI-EXCITATION fluorescence hyperspectral imaging (ME-HSI) illuminates a sample at several excitation wavelengths and records, for each, an emission spectrum in many bands. Every pixel thus carries an excitation–emission matrix (EEM) and an image is a four-dimensional array over excitation, emission, height and width. The joint domain separates fluorophores that a single excitation cannot, which is why endogenous autofluorescence from NADH, FAD, collagen and elastin has been used for label-free imaging of tissue state [1]–[4], including cardiac ablation lesions [5]–[7] and tissue classification with ME-HSI [8]–[10]. A particularly important derived quantity is the optical redox ratio, built from the FAD and NADH signals, which tracks cellular metabolism [2], [11].

Tissue is rarely imaged bare. In many applications the fluorophore of interest sits under a cover that scatters, absorbs and fluoresces itself: a collagen-rich layer, an epithelium, a blood-perfused dermis. The cover attenuates the excitation on the way in and the emission on the way out, adds its own autofluorescence, and, because haemoglobin absorbs strongly in the violet and green, *distorts the emission spectrum* rather than merely scaling it [12], [13]. A ratio such as FAD/(FAD+NADH) is therefore not a pure property of the metabolic state. Quantifying how much of its variance is caused by the cover, and designing acquisitions and estimators that are robust to it, requires ground truth that real measurements cannot supply: the true thickness, concentrations and optical properties of every pixel.

Simulation is the natural remedy, and it serves three purposes. (i) It supplies *ground truth* for developing and benchmarking processing methods. (ii) It supports *experiment design*, i.e. choosing excitation lines, emission bands, exposure and geometry before instrument time is spent; each added excitation multiplies acquisition time and photobleaches the sample. (iii) It enables *inverse problems*: parameter retrieval, unmixing and uncertainty quantification with a forward model that is faithful enough to be inverted.

A. Gaps

Photon-transport codes are mature [14]–[17], but they are solvers: they return fluence, reflectance or absorption for a given set of optical properties. A user who wants to simulate an *imaging experiment* must still add the excitation source and its line shape, dye photophysics (inner filter, quenching, FRET, bleaching, Raman), the emission transport, the tunable-filter bands and the detector noise, and then wrap everything in a design-of-experiments and inference pipeline. Hyperspectral scene simulators for remote sensing [18] cover the camera and scene but not fluorescence transport in turbid media. Our previous simulator [19] generated labelled 4D fluorescence cubes with a one-dimensional transport model; it could not

N. Meloyan is with the Akian College of Science and Engineering and the Engineering Research Center, American University of Armenia, Yerevan, Armenia, and the L.A. Orbeli Institute of Physiology NAS RA, Yerevan, Armenia (e-mail: narek.h.meloyan@gmail.com).

N. Sarvazyan is with the Akian College of Science and Engineering, American University of Armenia, the L.A. Orbeli Institute of Physiology NAS RA, Yerevan, Armenia, and George Washington University, Washington, DC, USA (e-mail: phynas@gwu.edu).

represent a cover whose thickness varies laterally, light spreading sideways, or buried inclusions, and it offered no study or inference tools.

B. Contributions

1) **A schema-first, end-to-end ME-HSI simulator** with a 3D scene editor, a headless Python API and a command-line interface that share one typed experiment contract (Section III).
2) **A physically comprehensive forward model**: ray optics with Fresnel/Snell and microfacet glint, dye photophysics, white-Monte-Carlo transport kernels with reciprocity, and a layered-block 3D Monte Carlo with layer-conforming tallies, a peel-off estimator with multiple-importance weights, voxelised inclusions, and a literature tissue-optics model (Section IV).
3) **A research toolkit**: a study engine with provenance, Sobol'/Morris sensitivity, unmixing, surrogate maximum-likelihood retrieval, Fisher-information/CRLB analysis with D-optimal channel selection, and calibration of a digital twin (Section V).
4) **A validation suite** against analytic and published references, including a check of the voxel inclusions against an independent Monte Carlo, and a sim-to-real study with measured collagen and droplet data (Section VI).
5) **A set of simulation experiments** on collagen-covered tissue that quantify attenuation and spectral distortion, confounding of the redox ratio, information-optimal acquisition and its limits under model error, thickness retrieval, depth-dependent unmixing failure and label-free segmentation, together with the depth to which absorbing and fluorescent inclusions can be detected under the cover (E4) and the convergence, parallel scaling and cost of the engine (E10) (Section VII).

II. Related Work

Tissue-optics Monte Carlo. MCML [14] established the standard for multilayered tissue with implicit capture, Henyey–Greenstein scattering [20] and Fresnel boundaries; the same ideas are reviewed in [21], [22], and voxelised variants such as mcxyz extend them to heterogeneous media. GPU implementations followed, for layered media [23] and for voxel media [15]. Complex boundaries are handled by mesh-based codes [16], [17], [24], [25] and hybrid mesh–voxel methods [26]. Our engine is closest to MCML in its formulation (Section IV-F) but is extended to tilted planar interfaces, voxel inclusions and an excitation/emission two-pass structure with a next-event estimator; it is not a general-purpose replacement for mesh-based codes.

Fluorescence Monte Carlo. Fluorescence propagation was modelled by Welch *et al.* [27] and Jacques [28]; accelerated schemes for layered tissue [29] and time-resolved fluorescence [30] reduce the cost by reusing the excitation transport. We follow the same decomposition, but add quantitative dye photophysics and the full camera chain, and we resolve photobleaching per voxel across an excitation sequence.

Hyperspectral scene simulators. DIRSIG [18] is the best-known physics-based image simulator for remote sensing, with a rich sensor model but without fluorescence in turbid media. Domain randomisation [31] uses synthetic data for sim-to-real transfer in robotics; we instead calibrate a physical model to measured data.

Fluorescence EEM analysis and unmixing. EEM data are classically analysed with multi-way models such as PARAFAC [32], [33]; in imaging, linear unmixing of spectral images is established [34]–[36], with constrained solvers [37], [38] and endmember extraction by NMF [39] or vertex component analysis [40]. Tunable-filter spectral imaging is reviewed in [41]–[43].

Intrinsic fluorescence in turbid media. Wu *et al.* [12] and Müller *et al.* [13] showed how scattering and absorption distort measured fluorescence relative to the intrinsic spectrum, and Richards-Kortum and Sevick-Muraca [44] reviewed quantitative tissue spectroscopy. Our simulator reproduces such distortions from first principles, so that correction strategies can be tested against known ground truth.

Global sensitivity analysis. Variance-based Sobol' indices [45]–[47] and Morris screening [48], [49] are standard in computational modelling. We expose them as first-class operations on simulated imaging experiments and report the Monte Carlo noise floor of the indices.

Experiment design and band selection. Optimal experimental design maximises functionals of the Fisher information matrix [50]; greedy maximisation of monotone submodular criteria has a $1-1/e$ guarantee [51]. In hyperspectral imaging, band selection is usually data-driven [52]–[55]. Here the forward model is known, so channel selection can be posed as a model-based design problem and compared with data-driven selectors.

Digital twins. Digital twins couple a model with measurements of a specific physical asset [56], [57]. We use the term in this restricted sense: a forward model whose instrument and fluorophore parameters are calibrated to laboratory data, with explicit statements of what is and is not identifiable.

III. System

SpectralForge Studio consists of a typed experiment contract, a Python engine and a web application (Fig. 1).

Schema-first contract. An experiment is a single Pydantic object holding fluorophores, materials (each a uniform optical medium), objects, a background, light sources with excitation sequences, cameras and acquisition settings including the physics toggles. The schema is exported to JSON Schema and TypeScript, so the browser editor builds only experiments that validate in the engine. The experiment is embedded verbatim in every exported cube, together with a provenance record (package and library versions, git commit and dirty flag, seed and the SHA-256 hash of the canonical experiment JSON), so any result can be traced to its configuration.

Engine. The engine is written in Python with numba-compiled kernels. A capture first traces camera rays through analytic solids into a wavelength-independent hit buffer, then assembles the spectral radiance for each excitation step from




Fig. 1. Architecture of SpectralForge Studio. The experiment contract is defined once in Python (Pydantic) and exported as JSON Schema and TypeScript types for the browser, so the editor and the engine cannot diverge. The server exposes the engine, the study engine, the inverse tools and calibration to the editor and to headless clients; module names refer to the `spectraforge.studio` package.

per-material optics and the transport kernel, projects it onto the camera bands and applies the detector model. Layered blocks bypass the 1D kernel and are rendered by the 3D Monte Carlo of Section IV-F. Every random draw has its own seeded stream, and results are bit-identical for a given seed and independent of the thread count.

Web editor and headless access. The browser application provides the 3D scene editor, a live preview, an *Explain pixel* tool that decomposes the radiance of one pixel into named components (dye, secondary, elastic, glint, Raman, leakage, background), and an analysis workspace. The same functionality is available through a headless Python API and a command-line interface, with study endpoints for design of experiments and sensitivity analysis. A built-in mock engine keeps the browser usable when no engine is reachable.

IV. Physics Model

Table I lists the notation. Lengths are in mm, wavelengths in nm, powers in mW and concentrations in µM unless stated.

A. Radiometric formulation

For an excitation irradiance E on a surface, the fluorescence radiance leaving towards the camera is

$$L(\lambda_{\mathrm{em}}) = E \sum_k \langle \mu_a^{(k)} \rangle_{\mathrm{line}} \, \Phi_k \, e_k(\lambda_{\mathrm{em}}) \, G(\lambda_{\mathrm{em}}), \qquad (1)$$

where $\langle \mu_a^{(k)} \rangle_{\mathrm{line}}$ is the absorption of dye k averaged over the excitation line shape and

$$\mu_a^{(k)}(\lambda) = \frac{\ln 10}{10} \, \varepsilon_k(\lambda) \, c_k \quad [\mathrm{mm}^{-1}], \qquad (2)$$

with c_k in M and ε_k in $\mathrm{M}^{-1}\mathrm{cm}^{-1}$. The kernel G absorbs everything else: Fresnel transmission in and out, refraction, the $1/n^2$ basic-radiance factor, the $1/4\pi$ of isotropic emission, depth integration, excitation attenuation and emission reabsorption. Inner-filter and reabsorption effects are therefore not tunable parameters; they emerge from G when μ_a is large.

TABLE I
Notation.

Symbol	Meaning
$\lambda_{\mathrm{ex}}, \lambda_{\mathrm{em}}$	excitation / emission wavelength (nm)
$\mu_a, \mu_s, \mu_s', \mu_t$	absorption, scattering, reduced scattering, transport ($\mu_a + \mu_s$) coefficients (mm^{-1})
g, n	Henyey–Greenstein anisotropy, refractive index
ε_k, c_k	molar extinction ($\mathrm{M}^{-1}\mathrm{cm}^{-1}$), concentration of dye k
Φ_k, $e_k(\lambda)$	quantum yield, unit-integral emission density (nm^{-1})
E	excitation irradiance at the surface ($\mathrm{mW\,mm}^{-2}$)
G	transport kernel ($\mathrm{mm\,sr}^{-1}$)
L	radiance ($\mathrm{mW\,mm}^{-2}\mathrm{sr}^{-1}\mathrm{nm}^{-1}$)
w, Φ_v	photon-packet weight (mW); track-length fluence in voxel v
B, S	blood volume fraction, haemoglobin oxygen saturation
c	index of a channel (λ_{ex}, band); C the number of channels
$\boldsymbol{\theta}$	parameter vector (thickness, concentration, …)
ϵ	relative systematic model error of a channel
$\mu_c(\boldsymbol{\theta})$, σ_c^2	expected signal and variance of channel c (e^-, e^{-2})
$\mathbf{F}$, $\mathbf{J}$	Fisher information matrix, Jacobian $\partial\mu_c/\partial\theta_i$
S_i, S_{T_i}	first-order and total Sobol' indices
$\mathbf{E}$, $\mathbf{a}$	endmember matrix, abundance vector

B. Ray optics and material optics

Camera rays are traced through analytic solids (box, sphere, cylinder, droplet, vessels by CSG, and layered blocks); a ray crosses a chain of clear media (glass, liquid) by Snell refraction with unpolarised Fresnel transmission until it meets a turbid medium. Shadow rays from every entry point to the source aperture collect Fresnel factors and Beer–Lambert attenuation of the clear media in front. Specular glint at the first interface follows a microfacet BRDF [58] with Smith shadowing [59] and unpolarised Fresnel reflectance. For side-

lit clear media (a laser crossing a cuvette) 16 Gauss–Legendre nodes along the refracted view chord, each with its own light-entry face, replace the slab assumption.

Dyes also experience Stern–Volmer self-quenching,

$$\Phi_{\mathrm{eff},k} = \frac{\Phi_k}{1 + K_{\mathrm{sv}} \sum_j c_j}, \tag{3}$$

and FRET pairs (D, A, k) transfer with a concentration-dependent efficiency

$$E_{\mathrm{FRET}} = \frac{k\, c_A}{1 + k\, c_A}, \tag{4}$$

which scales the donor emission by $1 - E_{\mathrm{FRET}}$ and adds a sensitised acceptor term (the efficiency is a lumped parameter; there is no spectral-overlap integral). Water Raman scattering at a fixed wavenumber shift of 3400 cm^{-1} follows the excitation line, $1/\lambda_R = 1/\lambda_{\mathrm{ex}} - 3400 \cdot 10^{-7}$ with λ in nm, with a Gaussian band of 480 cm^{-1} FWHM. Fluorescence that is re-absorbed on the emission path ℓ is re-emitted once and isotropically by the absorbing dye,

$$\begin{aligned} \mathrm{sec}_k(\lambda) = \Big[\sum_{\lambda'} L(\lambda') \big(e^{\tau} - 1\big) \tfrac{\mu_a^{(k)}}{\mu_a^{\mathrm{tot}}} \Big] \\ \times \Phi_k\, b_k\, \tfrac{\lambda_{\mathrm{abs}}}{\lambda_{\mathrm{em}}}\, e_k(\lambda) e^{-\tau(\lambda)}, \end{aligned} \tag{5}$$

with $\tau(\lambda) = \mu_a^{\mathrm{tot}}(\lambda)\ell$ and b_k the bleaching survival.

C. Photobleaching

Each pixel and dye carries a survival fraction s. After an excitation step the absorbed photon dose per molecule is $\sigma\, \Phi_{\mathrm{ph}}\, t_{\mathrm{tot}}$ with $\sigma = \ln(10)\, 10^3 \varepsilon / N_A$ (cm^2) and photon flux Φ_{ph} at the illuminated surface (with a build-up factor $\mathrm{clip}(1 + 2\pi\rho, 1, 6)$ for turbid media, ρ the elastic remission), and

$$x = \Phi_b\, \sigma\, \Phi_{\mathrm{ph}}\, t_{\mathrm{tot}}, \qquad s \leftarrow s\, e^{-x}, \tag{6}$$

where Φ_b is the bleaching quantum yield; fluorescence within the step uses the mean survival $s(1 - e^{-x})/x$. Acquisition order therefore matters, as in measurements of fluorescein [60], [61]. In the block engine the dose is resolved per voxel.

D. One-dimensional transport kernels

Clear media. For a non-scattering slab,

$$G = \frac{T_{\mathrm{in}} T_{\mathrm{out}}}{4\pi n^2 \cos\theta_i'} \, \frac{1 - e^{-\kappa L}}{\kappa}, \quad \kappa = \mu_t^{\mathrm{ex}} \frac{\cos\theta_o'}{\cos\theta_i'} + \mu_t^{\mathrm{em}}, \tag{7}$$

with L the view chord, θ_i', θ_o' the refracted excitation and view angles, and T Fresnel transmittances. The limits are the dilute result $G \to T_{\mathrm{in}} T_{\mathrm{out}} L / (4\pi n^2 \cos\theta_i')$ and the front-face plateau $1/\kappa$.

Turbid media: white Monte Carlo and reciprocity. Helmholtz reciprocity equates the radiance escaping at cosine μ_o from an isotropic emitter at depth z with the fluence produced at z by a collimated beam entering at μ_o. Both legs therefore use one family of collimated-incidence fluence profiles f:

$$G(\lambda_{\mathrm{em}}) = \frac{T_{\mathrm{in}} T_{\mathrm{out}}}{4\pi n^2} \int_0^d f_{\mathrm{ex}}(z)\, f_{\mathrm{em}}(z)\, \mathrm{d}z. \tag{8}$$

The white Monte Carlo [62] walks photons with $\mu_s = 1$ and no absorption and histograms the path length s; absorption enters afterwards as the weight $\exp[-(\mu_a/\mu_s)s]$. The baked tables therefore depend only on (g, n, n_{out}) and dimensionless variables, not on wavelength, dye or concentration, and are cached on disk. Per material and excitation, a small table $J(\mu_i', \mu_o', d, \lambda_{\mathrm{em}})$ is assembled by matrix products over path length, and the per-pixel kernel is a fused numba kernel. The same tables yield the elastic remission $\rho(\theta_i, \theta_o)$ (a Lambertian reflector of reflectance R has $\rho = R/\pi$), the mean fluorescence origin depth used by Eq. (5) and a radial subsurface-spread profile. Below $\tau = \mu_s d = 0.02$ a single-scattering thin-slab profile makes the kernel tend smoothly to Eq. (7).

E. Tissue optics

Tissue absorption follows Jacques [63] (eq. 1):

$$\begin{aligned} \mu_a(\lambda) = B\, S\, \mu_a^{\mathrm{HbO_2}} + B(1 - S)\, \mu_a^{\mathrm{Hb}} + W \mu_a^{\mathrm{water}} \\ + F \mu_a^{\mathrm{lipid}} + M \mu_a^{\mathrm{mel}} + \sum_k \mu_a^{(k)}, \end{aligned} \tag{9}$$

with B the blood volume fraction, S the oxygen saturation, W the water, F the lipid and M the melanosome volume fraction. The haemoglobin terms are $\mu_a^{\mathrm{HbO_2}} = \ln(10)\, \varepsilon_{\mathrm{HbO_2}} C_{\mathrm{Hb}} / 64500$ for whole blood ($C_{\mathrm{Hb}} = 150$ g/L) and the melanosome term is $\mu_a^{\mathrm{mel}} = 519 (\lambda / 500\,\mathrm{nm})^{-3.5}$ cm^{-1}. Optional bilirubin and β-carotene terms and the dyes of Eq. (2) are added. The tabulated spectra are literature data stored as CSV with citation headers: haemoglobin (Prahl compilation, 250–1000 nm), water (Pope and Fry [64] for 380–727 nm, Segelstein and Hale and Querry [65] elsewhere), lipid [66], and PhotochemCAD pigments (organic-solvent spectra, whose peaks in tissue can be shifted by 10–20 nm). Reduced scattering uses the Mie–Rayleigh form of Jacques (eq. 2),

$$\mu_s'(\lambda) = a' \Big[f_{\mathrm{ray}} \Big(\tfrac{\lambda}{500} \Big)^{-4} + (1 - f_{\mathrm{ray}}) \Big(\tfrac{\lambda}{500} \Big)^{-b_{\mathrm{Mie}}} \Big], \tag{10}$$

with $\mu_s = \mu_s'/(1 - g)$. Intralipid phantoms use the measured $\mu_s(\lambda)$ and $g(\lambda)$ of van Staveren *et al.* [67] and fat emulsions are compared with [68]. The library provides 18 materials (three epidermis melanin levels, papillary and reticular dermis, subcutis, muscle, oral and colon mucosa, a dysplastic nodule, liver, breast, brain grey matter, collagen gel and dense collagen, whole blood, an Intralipid phantom and an India-ink absorber) and three tissue presets (a three-layer skin, mucosa with a buried nodule, and a vessel under collagen). Fig. 2 shows the library.

F. Layered-block 3D Monte Carlo

A *layered block* is a stack of layers whose interfaces are exact planes $y = a_i + b_i x + c_i z$ (uniform or linear-wedge thickness) bounded by four vertical faces and the bottom $y = 0$, with air outside. Packets carry power w (mW) and, as in MCML [14], undergo implicit capture $w \leftarrow w \mu_s / \mu_t$, Henyey–Greenstein scattering [20], Russian roulette and unpolarised Fresnel reflection or refraction at the true plane normal, so a wedge steers light correctly and total internal reflection is




Fig. 2. Literature tissue-optics library: (a) absorption and (b) reduced scattering of the 18 materials, 380–900 nm.

included. The bottom face is a Fresnel interface to air: $w(1-R)$ leaves and is lost to the stage, wR stays in the block. Absorption is tallied on a *layer-conforming* voxel grid with lateral cells of size Δ and a fixed number of sublayers across the local thickness $t_l(x,z)$, so a voxel volume is $V_v = \Delta^2 t_l/S$ (exact because t_l is linear) and a zero-thickness wedge end is represented without staircasing.

Excitation pass. Once per excitation line, packets are launched from a jittered grid on the faces that face the source with weight $S\cos\theta_{\mathrm{inc}}\,dA\,T_{\mathrm{other}}$ per mW. The specular part of the entry Fresnel split is not transported (the tracer renders it as glint). The pass tallies the track-length fluence

$$\Phi_v = \frac{1}{V_v}\sum_{\text{segments}} w\,\ell, \qquad P_{\mathrm{abs},v} = \mu_a \Phi_v V_v, \tag{11}$$

and the energy balance $R + T + A + \text{specular} + \text{lost} = 1$. The absorbed power of a layer is split between its dyes in proportion $\mu_a^{(d)}/\mu_a^{\mathrm{tot}}$; each channel (excitation, layer, dye) becomes an emission source $A_d\,\Phi_d^{\mathrm{eff}}\,s\,(1-E_{\mathrm{FRET}})$, plus a Raman channel in aqueous layers.

Emission pass. Once per emission wavelength node, source voxels are drawn by stratified inverse-CDF sampling on the total source $q(v) = \sum_c \sigma_{c,v}$, uniformly inside the voxel and with isotropic direction. A packet carries $\mathrm{coef}_c = \sigma_{c,v}\,Q/(q(v)N)$, so a single run serves all channels of all excitations; the radiance tally is then multiplied by the area-normalised emission density $e_c(\lambda_{\mathrm{em}})$. Nodes are placed adaptively (dense where the transport factor bends) and the factor is interpolated linearly between them, since the sharp structure lives in e_c.

Cone estimator and peel-off. Counting only packets that leave inside the camera cone of half-angle θ_a gives the unbiased estimate

$$L_{\mathrm{cell}} = \frac{\sum_{\mathrm{cone}} w}{A_{\mathrm{cell}}\,(\mathbf{n}\cdot\mathbf{c})\,\pi\sin^2\theta_a}, \tag{12}$$

which is noisy for the small cells needed to resolve a wedge. The *peel-off* (next-event) estimator instead adds, at every scattering event and at the launch of an emission packet, the expected contribution of the straight line of sight to the camera. The internal direction $\mathbf{d}$ follows from inverting Snell's law through the interface chain, and

$$\begin{aligned}\delta L_{\mathrm{cell}} &= \frac{w\,p_{\mathrm{HG}}(\mathbf{u}\cdot\mathbf{d})\,T\,e^{-\tau}J}{A_{\mathrm{cell}}(\mathbf{n}\cdot\mathbf{c})}\,\omega_{\mathrm{MIS}},\\ J &= \prod_j \frac{n_{\mathrm{up},j}^2\cos\vartheta_{\mathrm{up},j}}{n_{\mathrm{in},j}^2\cos\vartheta_{\mathrm{in},j}},\end{aligned} \tag{13}$$

where τ is the optical depth of the straight segments, T the product of Fresnel transmittances, J the étendue factor that converts solid angle inside to solid angle in air, and $\omega_{\mathrm{MIS}} = 1/(1+\Omega_c\,p\,J)$ a balance-heuristic multiple-importance weight [69] that bounds the forward-peaked tail of the phase function (Ω_c the cone solid angle). A flag marks the paths the peel-off does not cover (the launch and the path after a Fresnel reflection), which the cone estimator still scores, so nothing is counted twice; side faces use the cone estimator. The peel-off needs about five times fewer packets (photon counts are scaled by 0.2), and the top-face tally is smoothed with a sub-cell Gaussian of $\sigma = 0.35$ mm.

Determinism. Photons are cut into fixed batches that re-seed from (seed, batch) and own private tallies summed in batch order, so results are bit-identical for any thread count.

G. Inclusions

A layered block may embed spheres, ellipsoids, cylinders and boxes (vessels are cylinders). The layer planes stay exact; the inclusions live on a regular Cartesian voxel map over the bounding box of the block, with voxel size h (default 0.2 mm), $2\times2\times2$ analytic supersamples per voxel and an integer region id (0 for the layer material, $r \geq 1$ for inclusion r). A later inclusion wins overlaps; a mixed voxel takes the majority id among its in-block samples (ties go to the inclusion) and is optically homogeneous; inclusions are index-matched to the host, so there is no Fresnel boundary at their surface. The photon step is limited by the layer planes and, inside 4^3-voxel coarse zones that contain an inclusion voxel, by voxel boundaries; the dimensionless free path is carried across each partial move, so a change of μ_t at a boundary is exact. Voxel crossings consume no random number and the roulette is not rolled after them, so an inclusion with the host's optics follows the host-only random walk. Host and inclusion tallies are kept separate and together conserve the absorbed power. Inclusions can carry dyes and emit; emission sources inside them are sampled from the inclusion voxels. Each capture returns ground truth per inclusion: a mask, the depth of the inclusion below the surface, the fraction of the pixel's fluorescence that originates in it, and an *Explain pixel* component per (dye, inclusion). A block without inclusions takes exactly the earlier code path and its results are bit-identical. With four inclusions occupying $1.4\,\%$ of the voxels the run time is about $2.0\times$ (excitation pass) and $2.1\times$ (emission pass) that of the homogeneous block, measured by interleaved runs on a loaded machine, so only the ratio is meaningful.

H. Sensor, tunable-filter bands and noise

The camera model follows a thin lens focused at distance S with f-number N, transmission T, magnification m. For a pixel of area A (including binning) exposed for t, the electrons collected in band b are

$$\begin{aligned} e_b &= k_0 \cos^4\theta \, A\, t \int L(\lambda) R_b(\lambda)\, T_{p,b}\, \mathrm{QE}(\lambda) \frac{\lambda}{hc}\, \mathrm{d}\lambda, \\ k_0 &= \frac{\pi T}{4N^2(1+m)^2}, \end{aligned} \tag{14}$$

with R_b the band response (pass-band times long-pass edge) and $T_{p,b}$ the peak transmission. Liquid-crystal tunable filters (LCTF) [41] are Gaussian pass-bands, filter wheels super-Gaussian of order six; a ray at field angle θ sees the centre wavelength shifted to $\lambda_0\sqrt{1-\sin^2\theta/n_{\mathrm{eff}}^2}$, and the cube is resolved in six field-angle rings. Blur is the thin-lens circle of confusion with an Airy floor and axial chromatic shift, varying per pixel and band. The detector chain is

$$\begin{aligned} S \to S(1+\mathrm{PRNU}) \to +\mathrm{dark} \to \mathrm{Poisson} \to +\mathcal{N}(0,\sigma_r^2) \\ \to \mathrm{clip} \to \div g \to \mathrm{round}, \end{aligned} \tag{15}$$

so for one pixel the variance in electrons squared is

$$\sigma^2(S) = (S+D) + \sigma_r^2 + g^2/12, \tag{16}$$

where D is the dark signal and $g^2/12$ the quantisation noise [70], [71]; for the mean of an ROI with n pixels it is divided by n. A white-reference mode renders a Spectralon plate with the same band set so that sample/white reproduces the laboratory normalisation, and a dark mode removes the source.

V. Research Toolkit

A. Study engine and provenance

A `StudySpec` names a base experiment (or preset), a list of parameters addressed by JSON pointers (continuous, integer or choice, linear or log scale, with optional linked pointers), a design (grid, random, Latin hypercube, Sobol' sequence, Saltelli or Morris), replicate seeds and metrics (ROI band sums and ratios, peak wavelength, EEM integral, Michelson or Weber contrast, contrast-to-noise ratio, ground-truth statistics). Each run is journaled so a study is resumable, every capture can embed its provenance, and a Methods-text generator produces a description of the experiment and study from the specification.

B. Global sensitivity analysis

For inputs scaled to the unit cube and a Saltelli design with matrices A, B and AB_i (A with column i from B) and N base samples, the first-order and total effects are

$$V_i = \frac{1}{N}\sum_j f(B)_j\big[f(AB_i)_j - f(A)_j\big], \tag{17}$$

$$V_{T_i} = \frac{1}{2N}\sum_j \big[f(A)_j - f(AB_i)_j\big]^2, \tag{18}$$

with $S_i = V_i/\mathrm{Var}(Y)$ and $S_{T_i} = V_{T_i}/\mathrm{Var}(Y)$ (estimators of Saltelli *et al.* [46] and Jansen [47]; Sobol' indices [45]). The first-order estimator is applied to outputs centred on $\frac{1}{2}(\bar f(A) + \bar f(B))$ (recomputed inside each bootstrap resample): the estimator is unbiased either way, but its variance scales with the squared output mean, which for band signals far exceeds the spread, and an uncentred version returned $\sum_i S_i = -13$ for the peak wavelength in our data; S_{T_i} is shift-invariant. Confidence intervals are percentile bootstraps over base rows. Morris screening [48], [49] reports $\mu^* = \mathbb{E}|EE_i|$ and $\sigma = \mathrm{sd}(EE_i)$ of the elementary effects. With R replicate seeds per design point, the within-point variance s_w^2 gives the Monte Carlo noise floor $(s_w^2/R)/\mathrm{Var}(\text{replicate means})$, the fraction of output variance that is engine noise; indices below it are not distinguishable from zero.

C. Unmixing

The joint EEM of a pixel (all bands of all excitations, C channels) is modelled as $\mathbf{y} = \mathbf{E}\mathbf{a} + \boldsymbol{\epsilon}$, which is exact for independent emitters seen through the same optical path and approximate when the paths differ, which is precisely the cover effect studied here. Endmembers come from the forward model (the per-pixel component decomposition at the purest pixels, an in-silico equivalent of a pure-component capture through the same instrument), from user ROIs, or

from NMF [39] or VCA [40]. Solvers are least squares, non-negative least squares [38] and fully constrained least squares ($a_k \geq 0,\ \sum_k a_k = 1$) by the augmented-NNLS construction of Heinz and Chang [37], optionally with inverse Poisson–Gaussian channel weights.

D. Surrogates and maximum-likelihood retrieval

A surrogate replaces the Monte Carlo forward model by an interpolant $\mu_c(\boldsymbol{\theta})$ of the noise-free ROI-mean signal of every channel on a grid or Sobol' set of one to three parameters. The noise-free output (expected electrons, before detector noise) is used, and the Monte Carlo seed is fixed across the grid (common random numbers) so finite-difference Jacobians are smooth; residual Monte Carlo noise is removed with a smoothing spline chosen by generalised cross-validation. The estimator minimises

$$\chi^2(\boldsymbol{\theta}) = \sum_c \frac{[y_c - \mu_c(\boldsymbol{\theta})]^2}{\sigma_c^2(\mu_c)}, \qquad \sigma_c^2 = \frac{(\mu_c + D_c) + \sigma_r^2 + g^2/12}{n}, \tag{19}$$

by iteratively re-weighted trust-region least squares, with start points from a scan of a dense surrogate table so that non-monotonic signals do not depend on the user's guess. The fixed point solves the Gaussian maximum-likelihood and, for Poisson data, the Poisson likelihood estimating equation $\mathbf{J}^\top \mathbf{W}(\mathbf{y} - \boldsymbol{\mu}) = 0$. Confidence intervals are profile-likelihood intervals [72], $\{\theta_p : \min_{\text{others}} \chi^2 - \chi^2_{\min} \leq q_{\chi^2}(\text{level}, 1)\}$, and open intervals are flagged. A map retrieval fits one parameter vector per tile of a noisy capture and reports bias, RMSE and interval coverage against the ground truth.

E. Fisher information, CRLB and optimal design

With channels independent and Poisson–Gaussian variance (16), the Fisher information is

$$\mathbf{F} = \sum_{c=1}^{C} \mathbf{I}_c, \qquad \mathbf{I}_c = \frac{\mathbf{J}_c \mathbf{J}_c^\top}{\sigma_c^2(\mu_c)}, \quad J_{c,i} = \frac{\partial \mu_c}{\partial \theta_i}, \tag{20}$$

and the Cramér–Rao bound on any unbiased estimator is $\mathrm{Cov}(\hat{\boldsymbol{\theta}}) \succeq \mathbf{F}^{-1}$. Eq. (20) equals the Poisson information μ'^2/μ when read noise vanishes and the Gaussian information when the signal is read-noise limited; the term from the dependence of the variance on $\boldsymbol{\theta}$ is neglected. The forward model is imperfect, so a purely photon-limited $\mathbf{F}^{-1}$ is optimistic. We therefore add a systematic model-error term to every channel,

$$\sigma_c^2 \ \rightarrow\ \sigma_c^2 + (\epsilon\, \mu_c)^2, \tag{21}$$

with a relative error ϵ that we estimate from the sim-to-real residuals: the median relative RMS residual of the 22 calibrated fits (nine collagen ROIs, 16 droplets, `derive_eps`) is $7.0\,\%$ (range 5.3–$9.6\,\%$), an upper bound because it contains camera noise. The term acts as independent per-channel variance and is optimistic if real errors are correlated across channels; it also does not represent bias. Because $\mathbf{F}$ is a sum of per-channel matrices, choosing k channels is the classical design problem [50]. D-optimality maximises $\log\det \mathbf{F}$, A-optimality minimises $\mathrm{tr}\,\mathbf{F}^{-1}$, and the target (c-optimal) criterion minimises $[\mathbf{F}^{-1}]_{pp}$, the bound of one parameter. The first is monotone and submodular, so the greedy set is within $1 - 1/e$ of the optimum [51]; it is checked against exhaustive search when $\binom{C}{k}$ is small. Exposure allocation over selected channels uses $I_c(t) = \mathbf{J}_c \mathbf{J}_c^\top (t/t_0)^2/\sigma_c^2(t)$ under a total-time budget and a full-well limit: it is convex in the shot-limited regime and solved by sequential quadratic programming otherwise. The selected set can be compared with the channels chosen by the label-free selectors of our earlier work [55].

F. Calibration and digital twin

For an optically thin, dilute layer the electrons in band b at excitation λ_{ex} are

$$S[\lambda_{\text{ex}}, b] = \sum_k a_k\, A_k(\lambda_{\text{ex}})\, M_k(\lambda_{\text{ex}}, b), \tag{22}$$

with $A_k = \sum_\lambda \mathrm{line}_{\text{ex}}(\lambda)\, \mathrm{exc}_k(\lambda)$ the excitation overlap, $M_k = \sum_\lambda R_b(\lambda) e_k(\lambda)$ the emission band projection and $a_k = c_k \varepsilon_k \Phi_k \ln 10/10$ a lumped amplitude that cannot be separated into concentration, extinction and yield from relative data. Gains (a global scale, or one scale per excitation) are eliminated in closed form by variable projection [73], so the optimiser sees only shape parameters (excitation and emission peak and width, red-tail skew, vibronic shoulder). The objective is a robust soft-L1 loss of per-excitation relative residuals from deterministic multi-starts; covariance is $s^2(\mathbf{J}^\top \mathbf{J})^+$ (statistical only), and a pixel bootstrap is available. Optional shared instrument terms are a smooth log-throughput polynomial and an LCTF ripple orthogonal to cubic polynomials. Which parameters are identifiable is reported explicitly, e.g. an emission peak below the first band is not.

VI. Validation

All numbers below are regenerated by scripts in the repository; only literature reference values are hard-coded. Table II summarises the checks; the following subsections give details.

A. Monte Carlo transport against standard benchmarks

The baked white-Monte-Carlo tables and a conventional MCML implementation both reproduce the Wang–Jacques/Prahl slab benchmark ($\mu_a = 1$, $\mu_s = 9$ mm^{-1}, $g = 0.75$, $n = 1$, $d = 0.2$ mm; $R_d = 0.09734$, $T_t = 0.66096$) to better than $0.1\,\%$ (Table III), conserve energy to 10^{-9}, match the exact incoherent multiple-reflection solution for a non-scattering, index-mismatched slab to $0.01\,\%$ and agree with each other for an oblique, index-mismatched turbid slab to $\leq 0.05\,\%$. Semi-infinite diffuse reflectance agrees with the Farrell–Patterson albedo formula [74] to $0.3\,\%$ ($n = 1$) and $1.1\,\%$ ($n = 1.4$; the dipole formula itself is good to a few percent at an index step). For a high-albedo medium ($\mu_a = 0.01$, $\mu_s' = 1$ mm^{-1}, $g = 0.9$) the absorbed-weight profile decays as $\exp(-\mu_{\text{eff}} z)$ with a fitted $\mu_{\text{eff}} = 0.1736$ mm^{-1} against $\sqrt{3\mu_a(\mu_a + \mu_s')} = 0.1741$ mm^{-1} ($-0.29\,\%$) over nine decades [75].

TABLE II
SUMMARY OF VALIDATION CHECKS (ENGINE VERSUS ANALYTIC OR INDEPENDENT REFERENCE).

Check	Reference	Result	Section
Wang–Jacques slab, R_d, T_t	0.09734, 0.66096	$< 0.1\,\%$ deviation	VI-A
Energy conservation (1D and 3D MC)	$R+T+A+\ldots=1$	10^{-9}	VI-A, VI-C
Clear index-mismatched slab	exact incoherent solution	$0.01\,\%$	VI-A
Diffusion decay, μ_{eff}	0.1741 mm^{-1}	0.1736 mm^{-1} ($-0.29\,\%$)	VI-A
Closed form vs. volume quadrature	internal	$0.008\,\%$	VI-B
Dilute limit vs. legacy engine	ratio CV	$0.09\,\%$	VI-B
Photon transfer: gain, read noise	1.5 e$^-$/ADU, 8.0 e$^-$	1.501, 7.99	VI-B
3D MC vs. MCML (benchmark slab)	MCML	R $-0.21\,\%$, T $+0.04\,\%$	VI-C
3D bottom boundary, clear slab	closed form	agreement to 5 digits	VI-C
Peel-off vs. cone	cone estimator	0.993–1.010	VI-C
Radial $R(\rho)$, 3D engine	independent MCML-type MC	within $1.1\,\%$	VI-D
Null / filling inclusion	host-only run	identical (bit level, peel-off off)	VI-D
Absorbing sphere, ΔR	first-order perturbation theory	0.948 of prediction	VI-D
Twin fast model vs. full capture	full engine	$4.1\cdot10^{-4}$ max. rel. dev.	VI-E

TABLE III
BAKED MONTE CARLO TABLES AND A CONVENTIONAL MCML CODE AGAINST THE WANG–JACQUES SLAB BENCHMARK [14] AND FURTHER CLOSED-FORM REFERENCES. R_d, T_t, A: DIFFUSE REFLECTANCE, TOTAL TRANSMITTANCE, ABSORPTANCE.

Case	Quantity	Engine	Reference	Diff.	Reference source
MCML slab (μ_a=1, μ_s=9, g=0.75, n=1, d=0.2 mm)	R_d	0.09742	0.09734	$+0.08\,\%$	Wang–Jacques
	T_t	0.66090	0.66096	$-0.01\,\%$	
	R+T+A	1.00000	1.00000	$+0.00\,\%$	energy conservation
Baked tables, same slab	R_d	0.09736	0.09734	$+0.02\,\%$	as above
	T_t	0.66137	0.66096	$+0.06\,\%$	
Clear slab (n=1.5, μ_a=0.8/mm, d=1 mm)	T_t	0.41426	0.41424	$+0.01\,\%$	exact multiple reflections
	R_d	0.00744	0.00745	$-0.06\,\%$	
Tables vs MCML (n=1.4, oblique, turbid)	R_d	0.16519	0.16525	$-0.04\,\%$	independent code
	T_t	0.23807	0.23804	$+0.01\,\%$	
	A	0.59655	0.59670	$-0.03\,\%$	
Semi-infinite, a'=0.99, n=1.0	R_d	0.74465	0.74671	$-0.28\,\%$	Farrell–Patterson
Semi-infinite, a'=0.99, n=1.4	R_d	0.61927	0.61277	$+1.06\,\%$	Farrell–Patterson

B. Optics, photophysics and sensor

Two independent implementations of the clear-medium physics, the closed form of Eq. (7) and the per-pixel volume quadrature, agree to $0.008\,\%$ from 0.1 µM to 1 mM fluorescein for a laterally uniform beam. In the dilute limit the 3D engine and the legacy 2D engine give the same emission shapes and cross-excitation ratios up to one geometry constant (61 points; coefficient of variation $0.09\,\%$). The inner-filter effect behaves as expected [76] (Fig. 3): in front-face geometry the signal per unit concentration saturates, while in 90° geometry the cuvette-integrated signal peaks near 13 µM ($A_{1\,\mathrm{cm}} \approx 0.9$) and the centre-pixel signal falls as $10^{-A/2}$. Reabsorption shifts the apparent fluorescein maximum from 512 nm (dilute) to 540 nm at 500 µM across 9 mm of liquid, but only to 518 nm in the front-face slab. Photobleaching follows $\exp(-kn)$ with $k = 0.094$ per step in a worked case, and the same total dose delivered as 300→40 mW or 40→300 mW records 76/55 % versus 96/71 % of the un-bleached signals. The photon-transfer curve of the detector chain returns the configured gain (1.501 against 1.5 e$^-$/ADU) and read noise (7.99 against 8.0 e$^-$) within $0.2\,\%$ and $0.1\,\%$. The depth-of-field blur coincides with the analytic thin-lens circle of confusion, and a 600 nm band shifts by -1.5 nm at the corner of the lab-replica field ($\theta_{\max} = 8°$, -13 nm at 25°).

C. Layered-block 3D Monte Carlo

Homogeneous slabs reproduce the MCML implementation to Monte Carlo noise (benchmark slab R $-0.21\,\%$, T $+0.04\,\%$; index-mismatched slab R $-0.05\,\%$, T $+0.07\,\%$; tissue-like slab R $-0.55\,\%$, T $+0.98\,\%$, A $-0.54\,\%$), energy is conserved to 10^{-9}, and two identical stacked layers equal one layer of the summed thickness (R $-0.04\,\%$, T $-0.11\,\%$). A 2 mm fluorescent layer reproduces the 1D kernel result of Eq. (1) to $+0.14\,\%$ for fluorescence and $-4.04\,\%$ for elastic remission (Fig. 4). For clear slabs the bottom-boundary treatment matches the closed forms $T = (1-R_0)^2 e^{-\tau}/(1-R_0^2 e^{-2\tau})$ and $R_{\mathrm{top}} = (1-R_0)^2 R_0 e^{-2\tau}/(1-R_0^2 e^{-2\tau})$ to the digits reported (e.g. $n = 1.5$, $\mu_a d = 0.5$: $T_{\mathrm{bottom}} = 0.55931$ for both). Turbid fluorescent slabs captured as a block and as a 1D box (secondary emission off) agree: block/1D $= 0.982$ at 6 mm and 0.987 at 1 mm. Thick slabs (about 20 mm) differ by up to $20\,\%$ in the red tail for a physical reason: the 1D engine assumes laterally uniform irradiance while the block sees the finite lamp spot.

The peel-off estimator agrees with the cone estimator within 0.993–1.010 across four configurations (Fig. 5); at equal photon count the cell-to-cell standard deviation of a uniform field is reduced by a factor 3.7 at 10^5 photons and 2.0 at $3\cdot10^6$ photons (about 6× for $g = 0$ and 2× for $g = 0.9$), and a sub-




Fig. 3. Inner-filter effect for fluorescein excited at 488 nm: front-face versus 90° geometry. The dotted curve is the excitation-only prediction $10^{-A/2}$ for the cell centre.




Fig. 4. The 3D Monte Carlo against the MCML implementation and the 1D engine for homogeneous slabs and a fluorescent layer.

cell smoothing of 0.35 mm reaches the noise level of the old cone estimator with five times fewer photons.

For a buried fluorescein layer under a wedge cover the signal per unit surface irradiance falls roughly exponentially with local thickness (effective attenuation 0.43 mm^{-1} at 490 nm excitation) and the blue/red emission ratio stays within 1.78–1.83, i.e. this nearly neutral cover attenuates but does not distort (Fig. 6). Under collagen (tissue with NADH/FAD at 365 nm), the buried signal drops by more than an order of magnitude between 0.2 and about 1 mm and the 440 nm band falls less than the 520 nm band, and the cover's own autofluorescence saturates with a decay length of 0.37 mm (Fig. 7).

D. Voxel inclusions and lateral transport

Identities. A null inclusion (same optics as the host) reproduces R, T and A of the host-only run to the last bit with the peel-off off ($2 \cdot 10^5$ photons), and the camera radiance to $6 \cdot 10^{-16}$ with it on. A box that fills a wedge layer equals replacing the layer, again to the last bit without peel-off; with peel-off on, R differs by 0.19 % (0.161004 against 0.161303), the camera radiance by 0.12 % and the layer absorption by 0.07 %. $R + T + A +$ specular $= 1$ to better than 10^{-9} in all six test scenes (edge-straddling, wedge-crossing, clear, vessel




Fig. 5. Peel-off versus acceptance-cone estimator: agreement of region sums and noise reduction at equal photon count.




Fig. 6. Buried fluorophore under a wedge cover, signal against local cover thickness (noise-free sensor, 8 pixels per bin). Top: fluorescein under an intralipid-like wedge (490 nm); (a) signal per unit irradiance with exponential fit, (b) blue (495–525 nm) to red (535–560 nm) ratio, (c) band radiance at two bands, raw and divided by the surface irradiance. Bottom: NADH/FAD tissue under collagen (365 nm); (d), (f) as above, (e) blue (420–480 nm) to green-red (520–600 nm) ratio.

with box), and the results are bit-identical for 1, 4 and 14 threads.

Lateral transport. For a pencil beam on a semi-infinite block ($\mu_a = 0.01$, $\mu_s' = 1$ mm^{-1}, $g = 0$, $n = 1.4$) the radial diffuse reflectance $R(\rho)$ of the 3D engine ($4 \cdot 10^6$ photons) agrees with an independent MCML-type code ($2{\cdot}10^7$ photons, no shared code) to within 1.1 % in all eight annuli between 1 and 10 mm (ratio 0.989–1.000; statistical noise about 0.5 %; total diffuse reflectance 0.6046; Fig. 8). Against the Farrell–Patterson–Wilson dipole [74] the engine is −28 % at 1–1.5 mm, +0.1 % at 2–3 mm and up to +12 % at 3–6 mm; the independent code shows the same ratios, so the difference belongs to the diffusion model (which fails within about two transport mean free paths of the source and uses an empirical boundary parameter), not to the engine.

Perturbation. Nine absorbing spheres ($r = 1.5$ mm, 3 mm deep) in a turbid block ($\mu_a = 0.1$, $\mu_s = 2$ mm^{-1}, $g = 0$, $n = 1.4$) decrease the diffuse reflectance by ΔR when their absorption rises by δ. Using common random numbers (per-photon streams; $2{\cdot}10^{-7}$ absolute noise) we compare ΔR with




Fig. 7. Collagen cover autofluorescence against thickness (saturating with decay length 0.37 mm) while the buried NADH/FAD signal falls.

the first-order prediction $\delta \int \Phi \Phi^{\dagger} \mathrm{d}V$, with Φ from the null run and the adjoint fluence $\Phi^{\dagger}$ obtained by re-emitting the power absorbed in the spheres isotropically. The measured/predicted ratio is 0.937, 0.916, 0.876 and 0.809 for $\delta = 0.01$, 0.02, 0.04 and 0.08 mm^{-1}, and a quadratic fit extrapolates to 0.948 at $\delta \to 0$ (Fig. 9). The 5 % shortfall has the expected sign (the isotropic-importance approximation overestimates the overlap integral because downward fluence is correlated with low upward importance), but we did not verify this with an exact adjoint, so the check is a consistency test, not a 0.1 % validation.

Buried bead. A fluorescein bead ($r = 1$ mm; $\mu_s' = 0.5$, 1.5, 3 mm^{-1}; 490 nm excitation; $3 \cdot 10^6$ photons) gives an on-bead excess radiance and Michelson contrast that fall monotonically with the depth of the bead top, from a contrast of 0.26–0.29 at 0.5 mm to 0.06–0.11 at 1.5–3 mm for $\mu_s' = 0.5$ mm^{-1}; beyond 2–3 mm the excess is below the Monte Carlo noise.

E. Sim-to-real: calibrating the twin

Machinery. The fast forward model of Eq. (22) agrees with a full capture to a maximum relative deviation of $4.1 \cdot 10^{-4}$ (target 3 %). Fitting noisy full captures (eight seeds, global gain, true emission 470/80 nm, excitation 350/70 nm) recovers the emission peak with bias +0.02 nm (sd 0.06 nm) at high signal and +0.31 nm (sd 0.63 nm) at low signal, with 95 % interval coverage of 100 % for four of five parameters; without dark subtraction the low-signal peak is biased by −1.5 nm.

Collagen. For six measured collagen ROIs (Pepsin and Acetic Acid samples) a one-component twin with per-excitation gain and a smooth throughput correction reaches $R^2 = 0.80$–0.92 and a mean spectral angle (SAM) of 6.2–7.8°, against $R^2 = 0.21$–0.78 and 11.0–16.9° for the nominal instrument (Fig. 10; Pepsin ROI 2: $R^2 = 0.919$, SAM 7.0° against 0.638 and 14.9°). Two components are no better and have a higher BIC. The calibration exposes what is not identifiable: the Nuance window starts at 420 nm, so the emission peak sits on its 380 nm bound; the identifiable descriptors are the centroid above 420 nm (498–505 nm) and the red half-maximum (516–526 nm). The fitted per-excitation gains reveal a radiometric inconsistency of the processed collagen data (a step of about 10^3 between 340 and 365 nm), which a global-gain model would hide. Refitting under alternative instruments (LCTF FWHM 10/14 nm, long-pass 405/420 nm, UV-enhanced CCD) moves the centroid only within 497–499 nm, but an unphysical flat LCTF throughput moves descriptors by 20–50 nm, so the calibrated emission is conditional on the LCTF curve.

Hold-out validation (Fig. 11; mean SAM). B1, calibrate on three excitations and predict the other three (six ROIs, four splits): corrected twin 8.4°, twin with ripple 8.6°, two components 12.3°, nominal 14.2°, empirical pooled spectrum 7.7°, flat 23.7° (in-sample refit 6.4°). B2, calibrate on one ROI and predict another (30 pairs): 7.8° (same dataset) and 7.4° (other dataset), empirical 6.9°/6.7°, flat 23.7°. B3, droplets with global gain, train on 310/340/385/415 nm and predict 325/365/400 nm absolutely without refitting a gain (16 drops): median $R^2 = 0.956$ (nominal 0.951), median SAM 6.0°, median $|\ln(\text{predicted}/\text{measured peak})| = 0.17$. The twin therefore predicts held-out excitations, samples and datasets with SAM about 1–2° above its in-sample value and removes about 90 % of the flat-baseline error, but it does *not* beat a non-parametric copy of another measured spectrum, and the excitation-dependent shape changes of collagen are not captured by one or two components. Its advantage is that it transfers to instrument settings that an empirical copy cannot represent. Scattering of the measured samples is not identifiable: all bands start at least 25–30 nm above the excitation line and there is no flat-field white reference.

VII. Experiments

All experiments use the public API on a tissue-under-collagen scene built from the literature materials of Section IV-E: a dense-collagen cover of variable thickness over a papillary-dermis layer carrying NADH, FAD, collagen and elastin, above a reticular-dermis substrate. They ran on an Apple M4 Pro (14 cores) with fixed seeds; each experiment's script, raw runs and provenance record are in the repository. Unless stated, intervals are 95 % bootstrap intervals. Most studies use the preview or 'fast' capture mode, whose per-




Fig. 8. Radial diffuse reflectance of a pencil beam on a semi-infinite medium ($\mu_a = 0.01$, $\mu_s' = 1$ mm^{-1}, $n = 1.4$). Left: 3D engine, independent MCML-type Monte Carlo and Farrell–Patterson–Wilson dipole. Right: ratios; the engine follows the independent Monte Carlo within 1.1 % (blue band ±3 %), while the dipole deviates by −28 to −16 % inside two reduced mean free paths (grey shading) and by up to +12 % at 3–6 mm.




Fig. 9. Perturbation check: decrease of diffuse reflectance caused by absorbing spheres against the increase of their absorption, from the Monte Carlo with common random numbers and from first-order theory.

pixel Monte Carlo noise matters for the pixel-level analyses (Section VIII).

A. E2: Cover-thickness response

Question. How fast does the buried signal decay with cover thickness, when does the cover's own autofluorescence take over, and how does the cover distort the emission spectrum?

Setup. Collagen thickness t = 0–2 mm (13 values, three seeds), excitation 340, 365, 400 and 430 nm, preview scale 0.25, $2 \cdot 10^6$ photons per pass, central 3×3 pixels; a blood study with $B = 0, 1, 3$ % at $t = 0.1$–1 mm (400 and 430 nm); and a gentle (40 mm long) and a steep (8 mm) wedge.

Results. The buried-layer signal falls approximately exponentially, by about three decades over 0.5 mm at 340 nm and five decades over 2 mm at 430 nm. The effective depth δ_{eff} of the fit $I = I_0 e^{-t/\delta_{\text{eff}}}$ rises from 0.073 to 0.358 mm with excitation wavelength (Table IV). Diffusion theory, $\delta = 1/\sqrt{3\mu_a(\mu_a + \mu_s')}$ evaluated for the two-way path, agrees at 340 and 365 nm, where absorption is strong ($\mu_a/\mu_s' \approx 0.5$), and overestimates δ by 1.4–1.5× at 400 and 430 nm ($\mu_a/\mu_s' \approx 0.04$–$0.1$), where its assumptions fail; the decay is also not exactly mono-exponential there ($R^2 = 0.90$–0.97), so δ_{eff} depends on the fit window. The cover's autofluorescence equals that of the buried layer (50 % crossover) at $t = 0.027$, 0.034 and 0.045 mm for 340, 365 and 400 nm, so at UV excitation any cover thicker than about 50 µm is dominated by its own emission; at 430 nm, where collagen barely absorbs, the crossover is at 0.55 mm (Fig. 12a). Once the cover dominates, the emission centroid has moved by −12 to −13 nm (340, 365 nm), −29 nm (400 nm) and −65 nm (430 nm), i.e. to the collagen emission (Fig. 12b), and the blue/red ratio, which grows by 1.5–2.5× already at 25–50 µm, reaches $2 \cdot 10^2$–$3 \cdot 10^2$ in the UV as the buried layer's red and green emission is lost first. Blood in the cover at $t = 0.25$ mm reduces the buried fluorescence to 0.30 ($B = 1$ %) and 0.09 (3 %) at 400 nm and to 0.32 and 0.057 at 430 nm, within the Soret band; the ratio to the blood-free spectrum shows the HbO_2 dips near 542 and 577 nm. The magnitude of the blue/red ratio with blood is uncertain (standard errors up to 100 % at $t \geq 0.5$ mm with three seeds); only its monotone rise with B and t is robust. For the wedges, the ratio of the 3D wedge response to the 1D uniform-block response of the same local thickness has a mean log-ratio of -0.03 ± 0.03 (400 nm) and -0.10 ± 0.04 (430 nm) for the gentle wedge and $+0.12 \pm 0.03$ and $+0.13 \pm 0.05$ for the steep one, with no trend with thickness. The uniform-block response therefore predicts the wedge to about 10–15 %. The lateral blur could not be identified: the fitted width sat at a bound of the search grid in most cases, so this is a null result at two seeds and $2 \cdot 10^6$ photons, not a measurement.




Fig. 10. Calibrated twin for collagen (Pepsin ROI 2; $R^2 = 0.919$, mean SAM 7.0° against 0.638 and 14.9° for the nominal instrument): (a–c) measured and simulated EEM (fraction of the excitation peak) and residual; (d) per-pixel spectral angle to the model; below, spectra at each excitation (calibrated and nominal instrument).

TABLE IV
E2: EFFECTIVE PENETRATION DEPTH OF THE BURIED SIGNAL (MC FIT, 95 % BOOTSTRAP INTERVAL) AGAINST TWO-WAY DIFFUSION THEORY.

λ_{ex} (nm)	δ_{eff} (mm)	diffusion (mm)	MC/diffusion
340	0.073 [0.057, 0.084]	0.068	1.07
365	0.105 [0.070, 0.132]	0.096	1.10
400	0.199 [0.164, 0.199]	0.300	0.66
430	0.358 [0.219, 0.436]	0.496	0.72

Take-home. At 340–365 nm the buried signal is confined to the top 0.1 mm and the cover's autofluorescence takes over within 50 µm, while at 430 nm the cover is nearly transparent to the excitation but absorbs the Soret band if it contains blood.

B. E3: Global sensitivity and confounders

Question. How much of the variance of the optical redox ratio FAD/(FAD+NADH) is caused by the cover, as opposed to the metabolic state, and does ratioing against an elastic or cover-autofluorescence channel remove it?

Setup. Eight inputs: cover thickness (0.1–1 mm), $\mu_s'(500)$ (1.5–4.5 mm^{-1}) and b_{Mie} (0.5–1.5) of the cover, blood fraction (0.2–5 %), StO_2 (0.4–1), buried NADH (20–150 µM) and FAD (10–100 µM) and collagen concentration (300–1500 µM). Saltelli design with $N = 64$ base rows (640 runs, none failed; mean capture 2.7 s, 10^6 photons per pass), 1000 bootstrap resamples, Morris with eight trajectories (72 runs) and a noise study of six points with four replicates. Outputs: the NADH band (365 nm excitation, 440–480 nm), the FAD band (450 nm, 510–550 nm), the redox ratio, and the peak wavelength. $N = 64$ is below the $N \geq 256$ we planned, which limits the tightness of the intervals; the Sobol' indices




Fig. 11. Hold-out validation. B1: held-out excitations. B2: held-out ROIs. B3: absolute amplitude prediction for droplets at held-out lines.

below use the centred estimator of Section V.

Results. The Monte Carlo noise floor is 0.007 for the redox ratio, 0.011 for the NADH and FAD bands, and 0.447 for the peak wavelength, which is therefore inconclusive. For the redox ratio, cover thickness alone has $S_i = 0.49$ $[0.23, 0.81]$ and $S_{T_i} = 1.04$ $[0.68, 1.57]$, FAD concentration $S_i = 0.22$ $[0.09, 0.37]$, and NADH concentration $S_i = -0.06$ $[-0.18, 0.06]$, indistinguishable from zero; the other inputs have $S_i \leq 0.10$. Summed over groups, the cover inputs (thickness, μ_s', b_{Mie}, blood, $\mathrm{StO_2}$, collagen fluorescence) carry a first-order share of 0.66 against 0.16 for NADH and FAD (Table V, Fig. 13), and a Gaussian-process surrogate (cross-validated $R^2 = 0.73$) gives group total indices of 0.91 and 0.16. The sum of first-order indices of the redox ratio is 0.82; for the NADH and FAD bands the cover share is 0.73 and 0.87. Morris screening agrees (rank correlation of μ^* with S_T of 0.83; thickness, then FAD, then μ_s' by μ^*), as does the scatter (Spearman $\rho = -0.67$ against thickness, $+0.04$ against NADH). Ratioing against an elastic channel (FAD/elastic), a cover-autofluorescence channel (NADH/AF) or both does not reduce the cover share: the group total index stays at 0.93–0.95 against 0.91 without correction.

Reliability. Single-input total indices are noisy at $N = 64$ and some exceed one (thickness for the redox ratio, 1.04; blood for the NADH band, 1.13), so we use S_{T_i} only as ranks with intervals. The NADH/AF ratio has $\sum_i S_i = 1.85$ and is unreliable for absolute shares. The bootstrap intervals of the group indices are degenerate (lower bound of the metabolic S_T interval, 0.17, above its estimate of 0.16), so only the point estimates are usable. An earlier analysis with an uncentred first-order estimator gave a redox-ratio $\sum_i S_i = -0.04$ and peak-wavelength $\sum_i S_i = -13$; the numbers here come from re-analysing the stored runs with the centred estimator (ST and Morris were unaffected).

TABLE V
E3: GROUPED SOBOL' INDICES OF THE COVER INPUTS AND OF THE METABOLIC INPUTS (NADH, FAD), $N = 64$. S_1 FROM THE CENTRED SALTELLI ESTIMATOR, S_T FROM JANSEN; S_T VALUES ARE NOISY (SEE TEXT).

output	cover S_1	metab. S_1	cover S_T	metab. S_T
NADH band	0.73	−0.03	1.40	0.04
FAD band	0.87	0.22	2.09	0.52
redox ratio	0.66	0.16	2.21	0.40
peak wavelength*	0.31	0.21	3.51	1.23

*Noise floor 0.447; inconclusive.

Take-home. In this model, the redox ratio is mainly a measure of the cover (thickness alone $S_i = 0.49$), and the simple ratio corrections tested do not remove this.

C. E4: Inclusion detectability

Question. Down to which depth under the collagen cover can a buried fluorescent nodule or an absorbing blood vessel be detected, and how does the contrast decay with depth?

Setup. A 16×16 mm block with a dense-collagen cover of thickness d on autofluorescent tissue (the buried-optics scene of E2); the top of the inclusion touches the cover, so d is the depth of the inclusion. Inclusions are voxel spheres (nodule, $r = 0.5$, 1, 2 mm, of NADH 400 µM and FAD 150 µM, i.e. 6.7× and 3.8× the host; an author-chosen composition, not a literature value) and vessels (cylinders of oxygenated blood, $r = 0.25$, 0.5, 1 mm) with 0.1 mm voxels, at 340,




Fig. 12. E2, cover-thickness response (three seeds; error bars show the uncertainty over seeds). *Top:* (a) buried-layer signal against collagen thickness with exponential fits (points below 10^{-5} of the thin-cover signal, the Monte Carlo floor, are clipped); (b) fitted δ_{eff} against two-way and band-resolved diffusion theory; (c) their ratio. *Bottom:* (a) fraction of the signal that is cover autofluorescence, with 50 % crossovers marked; (b) emission-centroid shift relative to $t = 0$; (c) blue/red band ratio relative to $t = 0$ (log scale).

365, 400 and 430 nm excitation with the laboratory protocol exposures (1000, 250, 50, 80 ms) and the Nuance LCTF (31 bands) at the native 348×256 pixels (0.21 mm on the block). Depths are 0.05–3 mm (nodule) and 0.05–2 mm (vessel), with $1.2 \cdot 10^6$ photons per pass and three seeds. The contrast is the difference of the region-of-interest (ROI) and annulus signals (bands at least 25 nm above the excitation) between the capture with the inclusion and a matched capture without it, which uses the same random numbers (common random numbers). The CNR divides this by the sensor noise of a native pixel (44–55 e^-; the analytic variance agrees with 40 detector realisations); the ROI-averaged CNR uses the noise of the ROI mean. Detectability is CNR $= 3$.

Results. Absorbing vessels are detectable per pixel only close to the surface and only in the UV: the CNR$= 3$ depth at 340 nm is 0.24, 0.39 and 0.44 mm for $r = 0.25$, 0.5 and 1 mm, while at 365, 400 and 430 nm the per-pixel CNR is already below 3 at the shallowest depth (0.05 mm; Figs. 14, 15a). Averaging over the ROI helps with larger inclusions, to 0.4–1.5 mm (e.g. 0.67, 1.47, 1.34 mm at 340 nm), but the values beyond about 0.75 mm are within twice the Monte Carlo floor and are upper bounds (Fig. 15b). The nodule never reaches CNR $= 3$ at any depth, size or line: the peak per-pixel CNR is 2.6 ($r = 0.5$ mm, 340 nm), of the order of the Monte Carlo floor (1.0–2.1), and its contrast changes sign with depth and seed. Exogenous labels (EBFP2, DAPI, fluorescein, PpIX at 10–50 µM, $d = 0.5$–1 mm, $r = 1$ mm) gave no positive contrast either. The reason in this model is physical: with the




Fig. 13. E3, Sobol' indices of eight inputs for four outputs ($N = 64$ base rows, 95 % bootstrap intervals). Light bars: first-order S_i; dark bars: total S_{T_i}; orange: cover and confounders, blue: metabolic state. Thickness and blood dominate the NADH and FAD bands and the redox ratio.

dense cover the excitation is absorbed within $1/\mu_{\mathrm{eff}} = 0.07$–$0.1$ mm at 340–365 nm, and the autofluorescence of the cover itself is the dominant background; the nodule mainly changes the image by modulating the cover fluorescence through its different scattering.

The contrast of the vessel decays with a length $\delta = 0.5$–1.7 mm (fits of $|\Delta| \propto e^{-d/\delta}$ to four to six points, intervals 0.4 to above 4 mm; Fig. 15d). This is 5–10× longer than the buried-emitter prediction $1/(\mu_{\mathrm{eff,ex}} + \mu_{\mathrm{eff,em}})$ of the cover (0.07–0.5 mm) and 2–4× shorter than the emission-only length $1/\mu_{\mathrm{eff,em}}$ (2.0–3.2 mm). The vessel is therefore not seen as an emitter attenuated twice by the cover; it acts as an absorber of the diffusely transported emission of the cover and the tissue, and its contrast decays on a scale set by the emission side. The CNR ordering is 340 > 365 > 430 > 400 nm: the 400 nm line is limited by its 50 ms exposure (background 250 e^- against 1300–2400 e^- in the UV).

Caveats. Only three seeds and $1.2 \cdot 10^6$ photons: the Monte Carlo noise of the contrast (50–110 e^- at 340 and 365 nm) is comparable to the signal beyond 0.5 mm and exceeds the sensor noise, so the maximum depths are upper bounds with crude bootstrap intervals; common random numbers reduce the noise of the difference by 1.5–6× but do not cancel it, because the emission source list changes when an inclusion is added. The results are specific to the dense-collagen cover; a clearer cover would give millimetre-scale depths. The nodule composition is an assumption, the exposures differ between lines, and a vessel of $r = 0.25$ mm is only 2.5 voxels in radius.

Take-home. Under a dense collagen cover, absorbing vessels are detectable per pixel only to about 0.2–0.45 mm (340 nm) and ROI-averaged to about 0.4–1.5 mm, and an endogenous-contrast nodule is not detectable at all; the contrast decays on the emission-side diffusion scale.

D. E5: Information-optimal acquisition

Question. Which channels of the $8 \times 31 = 248$-channel lab-replica protocol carry the information on cover thickness, how many reach the Cramér–Rao bound (CRLB) of the full protocol, and how robust is this to model error?

Setup. Thickness surrogate on eight grid points over 0.05–1.5 mm (leave-one-out relative RMS 6.4 %, maximum 9.2 %), scale 0.25, ROI of 3744 full-resolution pixels, full-protocol exposure 233 s. The model-error level ϵ of Eq. (21) is 0, 0.5, 1, 2, 5 and 7 % (the last is the 7.0 % derived from the sim-to-real fits). Greedy and exchange-refined D-optimal selection are compared with 40 random subsets (median, IQR), uniform subsets, and the label-free selection of our earlier work [55] ($k = 8$, two seeds). Efficiency is the ratio of the standard deviation of 200 ML fits to the CRLB.

Results. At $t = 0.6$ mm the photon-limited CRLB for thickness is 0.26 µm alone, 0.90 µm with an unknown amplitude and 1.05 µm with per-excitation gains. Model error lifts this floor roughly linearly in ϵ (about 1.3 µm per percent), to 2.2–9.5 µm for $\epsilon = 0.5$–7 % (Table VI, Fig. 16): already at 0.5 % the floor is 2.4× the photon-limited one, so precision is systematics-limited. The photon-limited values are bounds on information and not achievable precisions; the floors are still one to two orders of magnitude below the 25–200 µm errors of E6, which the variance-only CRLB cannot contain. For thickness and amplitude, 28 channels come within 10 % of the full-protocol CRLB at $\epsilon = 0$ (81 % less exposure time) and 28–45 channels for $\epsilon > 0$ (88–94 % less), i.e. about 12 % of 248; for the joint (thickness, NADH, μ_s') problem 13 channels suffice photon-limited and 20–27 under model error. Random and uniform subsets reach this level only with $k \geq 200$. At $k = 8$ the information-optimal set gives a CRLB 1.24 µm against a random median of 12.6 µm (10×) at $\epsilon = 0$, and 6.7 against 34 µm (5×) at $\epsilon = 2$ %. The optimal set depends




Fig. 14. E4, contrast-to-noise ratio of the per-pixel instrument contrast against the depth of the inclusion top for nodules (top) and vessels (bottom) at three radii and four excitation lines: solid, mean over three seeds with the seed range shaded; dotted, single capture without a matched baseline. The dashed line is CNR = 3. No nodule reaches it; vessels do so only at 340 nm and below about 0.45 mm.

on ϵ (Fig. 16a): photon-limited selection picks UV-excited blue channels (310–325 nm excitation, 420–480 nm emission), which have high absolute signal, while for $\epsilon \geq 0.5\,\%$ it moves to 415–430 nm excitation and 455–570 nm emission with larger relative thickness sensitivity (Jaccard index with the photon-limited set 0.14 at $0.5\,\%$ and 0.07 for $\epsilon \geq 2\,\%$). Using the photon-optimal set when $\epsilon = 2\,\%$ costs 28 µm instead of 6.7 µm. The label-free selection is 2–2.5 µm photon-limited ($2\times$ worse than optimal, $5\times$ better than random) but 39–44 µm at $\epsilon = 2\,\%$, comparable to random (34 µm). The photon-only efficiency of the ML fit is 1.01 and 0.99 at $t = 0.25$ and 0.6 mm (0.95–1.03 for $k = 8$); at 1.1 mm the surrogate bias (−230 µm) invalidates the comparison. With $\epsilon = 2\,\%$ the ratio is 1.13 to 1.64 for all channels and 0.99 to 4.2 for $k = 8$, larger where bias exceeds the standard deviation.

Caveats. The finite-difference and surrogate Jacobians disagree at $t = 0.9$ mm (standard deviations 0.39 and 0.92 µm, Monte Carlo noise in the differences); the exposure allocation of the design tool did not beat equal allocation in this scene and is not used; and this is one scene and one ROI.

Take-home. About 30 of 248 channels carry the information of the full protocol, but the choice depends on the assumed model error and photon-limited CRLBs overstate the precision.

TABLE VI
E5: THICKNESS PRECISION FLOOR AT $t = 0.6$ MM (THICKNESS AND AMPLITUDE UNKNOWN) AND NUMBER k OF GREEDY-SELECTED CHANNELS WITHIN 10 % OF THE FULL-PROTOCOL BOUND, AGAINST MODEL ERROR ϵ.

ϵ (%)	0	0.5	1	2	5	7
floor (µm)	0.90	2.2	2.8	4.0	7.3	9.5
k (of 248)	28	35	28	30	40	45

E. E6: Thickness mapping

Question. How accurately can the cover thickness be mapped on a wedge, and are the retrieval intervals calibrated?

Setup. Tile retrieval (4×4 output pixels, scale 0.25) with the surrogate fit of Section V on a tissue wedge (0.1–1.4 mm, 144 tiles, 248 channels) and a phantom wedge (2–0 mm, 108 tiles, six excitations), with free amplitude, three exposure levels (0.003, 0.03 and 1 of nominal), photon-only or $\epsilon = 2\,\%$




Fig. 15. E4, summary. (a, b) Depth at which the vessel reaches CNR = 3, per pixel and ROI-averaged (triangles: already below 3 at 0.05 mm). (c) Peak CNR of the nodule over depth (dots) and the Monte Carlo floor of the three-seed mean (crosses). (d) Fitted decay length of the vessel contrast against the diffusion length of the cover: filled, emission-only $1/\mu_{\mathrm{eff}}(\lambda_{\mathrm{em}})$; open, two-way $1/(\mu_{\mathrm{eff,ex}} + \mu_{\mathrm{eff,em}})$.

weights and standard errors, bootstrap intervals over tiles, and channel sets "all", $k = 8$ information-optimal and five random $k = 8$ draws. The surrogate is trained on uniform-thickness blocks, so lateral transport at the wedge slope is unmodelled. One noise realisation.

Results. At nominal exposure the tissue wedge is retrieved with an RMSE of 405 µm (343–469) using photon-only weights and 262 µm (220–316) with $\epsilon = 2\,\%$ (bias +89 and +25 µm); the phantom reaches 184 and 160 µm. The phantom RMSE rises only from 184 to 246 µm (photon-only) when exposure falls from 1 to 0.003, while the tissue RMSE reaches 776 µm at 0.003 (Fig. 17), so at nominal exposure accuracy is limited by model bias, not photons. The photon-limited tile CRLB is 17 µm (tissue) and 2.7 µm (phantom) at nominal exposure, so the observed RMSE exceeds it by 15–70×. Nominal 95 % coverage is 8 % (tissue) and 5 % (phantom) with photon-only standard errors (mean 23 and 3 µm against errors of 100–400 µm), and 19 and 13 % with $\epsilon = 2\,\%$; coverage looks acceptable at the lowest exposure (0.88 for the phantom) only because the photon standard error is large. Eight information-optimal channels give 274 µm on the tissue (coverage 0.54) against 405 (photon-only) and 262 µm (all channels), and 498 ± 91 µm for random $k = 8$ sets; on the phantom, however, all channels (160–184 µm) beat the optimal $k = 8$ set (247–335 µm), which was optimised on the tissue Fisher information for thickness, not on mapping accuracy, so CRLB-optimality does not guarantee low bias. In the tissue the RMSE is nearly flat with thickness (260, 261 and 266 µm in bins 0–0.5, 0.5–1 and 1–2.1 mm).

Take-home. Wedge thickness is retrievable to about 160–260 µm RMSE, limited by model bias; photon-based intervals are over-confident by an order of magnitude, and model-error weighting improves the RMSE by about 35 % and the coverage 2.5× without restoring nominal coverage.

F. E7: Depth-dependent unmixing failure

Question. How does cover depth bias the linear unmixing of buried NADH, FAD and the matrix, and does a depth-matched library remove the bias?

Setup. Uniform cover $t = 0.05$–1.3 mm (eight values), noisy 'fast' capture (scale 0.25), NNLS on the mean spectrum of 24 random ROI pixels, groups NADH, FAD, matrix (collagen and elastin) and scatter (elastic, glint, leakage), joint 8-line EEM (also 4-line and 365 nm only). Signatures and ground truth are taken from the *Explain pixel* decomposition of the *same* 24 pixels (oracle libraries); a held-out variant builds signatures from 12 pixels and tests on the other 12. Errors are normalised by the true signal at 0.05 mm. A first version took endmembers from three pixels and ground truth from other pixels, and the per-pixel Monte Carlo noise of the 'fast' capture (about 39 % per channel, median over channels) then produced errors of −33 to +60 % at 0.05 mm; this was an artefact and is removed here (noisy floor at 0.05 mm: NADH +1 %, FAD +6 %; noise-free: exactly zero).

Results. The detected NADH signal falls 10^4-fold between 0.05 and 1.3 mm (53 % left at 0.1, 2.2 % at 0.4, 0.34 % at 0.6 mm), FAD 50-fold, while matrix and scatter rise by 1.5–1.8×. A library measured on thinly covered material assigns NADH = 0 for every $t \geq 0.1$ mm although the true NADH is 53 % of its shallow value at 0.1 mm (noise-free data give the same, so it is bias), which makes the redox-ratio error +0.59 at 0.1 mm (Table VII, Fig. 18). With endmembers simulated at the right depth, NADH and FAD are recovered to within ±6 % of the shallow signal up to 0.6 mm (redox error ≤ 0.05 for 0.2–0.6 mm; the +0.54 at 0.1 mm is a noise-limited outlier of the joint-8 fit). Beyond about 0.6 mm failure is a detection limit, not a bias: the NADH estimate sits at a noise floor of 1–2.5 % of the shallow signal while the true value is ≤ 0.3 %, so the redox ratio saturates wrongly (−0.73 at 1.3 mm), and




Fig. 16. E5. *Top:* CRLB standard deviation of the cover thickness, (a) for all 248 channels against thickness and model error $\epsilon = 0$ to $7\,\%$, (b) with nuisance parameters (amplitude, per-excitation gain, joint fit with NADH and μ_s'), (c) at $t = 0.6$ mm against ϵ; the grey band marks the retrieval errors of E6. *Bottom:* (a) information on thickness per channel at $t = 0.6$ mm with the eight channels selected at $\epsilon = 0$ (circles) and $5\,\%$ (squares); (b, c) CRLB against the number of channels k for information-optimal, random (median, IQR) and uniform selection and the label-free autoencoder selection, at $\epsilon = 0$ and $2\,\%$ (stars mark the k within $10\,\%$ of the full-protocol bound).

per-layer endmembers (condition number 5–34) do not help. In the held-out variant, signature noise alone gives redox errors of -0.4 to -0.5 at 0.4–0.6 mm and -1.0 at 1.3 mm, which is partly an artefact of the 'fast' noise. The 4-line set behaves like the 8-line set (condition number 6–11 against 4–7); 365 nm alone is ill-conditioned (condition number 107 at 0.05 mm, $8 \cdot 10^3$ at 0.6 mm).

Take-home. Unmixing with a thin-cover library fails by spectral distortion between 0.1 and 0.6 mm, where a depth-matched library repairs it to a few percent; beyond that, NADH is below the noise floor and no library recovers it.

TABLE VII
E7: ERROR OF THE UNMIXED REDOX RATIO FAD/(FAD+NADH) (ESTIMATED MINUS TRUE; TRUE VALUE 0.30 AT 0.05 MM, 0.88 AT 0.4 MM, 1.00 FOR $\geq$ 1 MM), JOINT 8-LINE EEM, NOISY DATA.

library, t (mm)	0.05	0.1	0.4	0.6	1.3
thin-cover	+0.009	+0.59	+0.119	+0.017	0.000
depth-matched	+0.009	+0.54	−0.031	+0.017	−0.726
depth, per layer	+0.011	+0.04	+0.077	−0.010	−0.702
depth, held-out	−0.003	+0.37	−0.467	−0.398	−1.000

G. E8: Sim-to-real consolidation

Question. How well does the calibrated twin predict measured spectra, and does the calibrated collagen emission keep its shape when it is transported through the 3D scene?

Setup. Table VIII consolidates the hold-out results of Section VI-E (mean ± s.d. of the spectral angle in degrees over ROIs and splits; B1 six ROIs with held-out excitations, B2 30 ROI pairs; the droplet test B3 has 16 drops). In addition, the calibrated collagen emission replaces the library collagen in




Fig. 17. E6, thickness mapping on a tissue wedge and a phantom wedge. (a) Retrieved against true thickness at nominal exposure (all channels); (b) RMSE, (c) nominal 95 % interval coverage and (d) bias against relative exposure for all channels and eight information-optimal channels, with photon-only ($\epsilon = 0$) or $\epsilon = 2\,\%$ weights. Accuracy saturates at about 160–260 µm while the photon-only intervals cover only 5–8 % of the tiles.




Fig. 18. E7, unmixing against cover thickness. (a) Detected signal of each group relative to 0.05 mm; (b, c) absolute error of NADH and FAD abundance relative to the shallow true signal for the four libraries (the black line is the true signal); (d) redox-ratio error for joint 8-line, 4-line and 365 nm-only channel sets. The thin-cover library fails from 0.1 mm; depth-matched endmembers help until the NADH signal falls below the noise floor.

the 3D Monte Carlo for a 3 mm collagen slab and a 0.5 mm collagen cover on dermis (lab source and Nuance twin, 10^6 photons, central $40\,\% \times 40\,\%$ ROI, bands at least 20 nm above the excitation, 1200 µM), and the simulated emission shape is compared with the measured Pepsin ROI 2 spectrum at each excitation. The measured samples are thin, so this tests whether the calibrated shape survives transport, not whether the simulation predicts a measured thick cover.

Results. The one-component twin with throughput correction has $8.4° \pm 1.7°$ on held-out excitations (nominal 14.2°, empirical copy 7.7°, flat 23.7°), 7.8° and 7.4° on held-out ROIs of the same and the other dataset, and a droplet median $R^2 = 0.956$ at held-out lines without refitting a gain (Table VIII); it does not beat an empirical copy. In the cover scenes, the library collagen emission has a mean spectral angle of 52.1° (slab) and 51.6° (cover on dermis) from the measurement and the calibrated emission 17.2° in both scenes; once the ROI-fitted instrument throughput polynomial, a property of the real instrument that the nominal twin lacks, is applied, the angles are 57.4° and 58.2° (library) and 9.4° and 8.6° (calibrated; Fig. 19), against 7.0° for the thin-layer fit itself. The calibrated shape therefore survives transport with about 1.6°–2.4° of added deviation after throughput correction. One ROI, one calibration, single seed.

TABLE VIII
E8: MEAN SPECTRAL ANGLE (°, MEAN ± S.D.) FOR HELD-OUT EXCITATIONS (B1) AND HELD-OUT ROIS OF THE SAME (B2 SAME) AND ANOTHER (B2 OTHER) DATASET.

Model	B1	B2 same	B2 other
In-sample refit	6.4	–	–
Twin, corrected (1 c.)	8.4 ± 1.7	7.8 ± 1.4	7.4 ± 1.0
Twin + ripple	8.6 ± 1.8	7.8 ± 1.4	7.6 ± 1.1
Twin, 2 comp.	12.3 ± 9.3	8.0 ± 1.4	8.1 ± 1.2
Twin, nominal	14.2 ± 2.5	13.7 ± 2.2	13.6 ± 2.1
Empirical copy	7.7 ± 2.0	6.9 ± 1.5	6.7 ± 1.3
Flat spectrum	23.7 ± 2.1	–	–

Take-home. With the instrument throughput applied, calibration moves the simulated collagen from about 57° to about 9° of the measured shape, close to the thin-layer fit (7°), but




Fig. 19. E8, calibrated collagen emission in the simulated 3 mm collagen slab (normalised): measured spectrum (dots), simulation with the calibrated emission (solid) and with the library collagen (dashed), at three excitations; the titles give the spectral angles (calibrated / library) with the fitted instrument throughput applied.

the empirical-copy baseline is not beaten.

H. E9: Label-free segmentation under a cover

Question. How does burial under a collagen cover limit unsupervised separation of tissue types, and does the information-optimal channel selection of E5 help?

Setup. 14 captures (Latin-hypercube design, seed 9) of three side-by-side tissue types (NADH-rich, FAD-rich and stroma buried layers, 1 mm) under a dense-collagen wedge ($t_0 \in [0, 0.3] \rightarrow t_1 \in [1.0, 1.5]$ mm), seven excitations (310–430 nm) by 31 bands = 217 channels, 'fast' mode, 87×64 pixels, $3 \cdot 10^5$ photons per step, about 1900 tissue pixels per capture, $K = 3$. Methods: k-means and Gaussian mixture on z-scored PCA features (10 components, raw or 3×3 spatially binned) and k-means on the latent of the convolutional autoencoder of our repository (30 epochs, default sparsity weight, no tuning). Adjusted Rand index (ARI) and NMI against tissue type, as mean $\pm$ s.e. over the 14 captures, per bin of local cover thickness; fits are made separately in each bin to remove the thickness gradient. A supervised shrinkage-LDA (four-fold cross-validated accuracy) is shown as a non-label-free reference. A noise check on one scene with two seeds gives a relative pixel noise of $8.2\,\%$ at the dataset photon budget and $7.3\,\%$ at four times the photons, so the Monte Carlo share is about $4.3\,\%$ and the sensor share about $7.0\,\%$ in this (channel-mean-normalised) metric.

Results. With 3×3 binning, the per-bin ARI of k-means falls monotonically with burial (Table IX, Fig. 20, top a): 0.92 ± 0.05 below 0.05 mm, 0.58 ± 0.04 at 0.1–0.2 mm, 0.19 ± 0.02 at 0.6–1 mm and 0.08 ± 0.01 at 1–1.5 mm; unbinned pixels are lower at every depth (e.g. 0.72 to 0.05). Binning helps most at 0.1–0.6 mm ($+0.1$ to $+0.15$) and not at the extremes. The tissue information is still present: supervised LDA with binning stays at 0.96–1.00 at all depths (unbinned, $0.98 \rightarrow 0.43$), so it is the dominant unsupervised variance that is lost, not the information. Clustering the whole capture fails: ARI against tissue type is 0.04 ± 0.01 (k-means, raw), 0.07 ± 0.01 (k-means, binned), 0.15 ± 0.02 (GMM, binned) and 0.07 ± 0.00 (autoencoder), and the clusters follow the thickness gradient and illumination edges, not tissue type (NMI with thickness bin 0.08–0.19). The autoencoder does not beat PCA with a mixture model (0.07 against 0.15 whole-capture ARI), a null result on this benchmark without tuning. For channel reduction (per-bin k-means, binned, mean over 0–0.35 mm; all 217 channels 0.71), the information-optimal sets of E5 are *worse* than uniform or random sets at $k = 5$ (0.18 against 0.26 and 0.29) and $k = 15$ (0.23 against 0.48 and 0.48) and similar at $k = 26$ (0.55 against 0.63 and 0.55; Fig. 20). Sets optimised for thickness discard the excitation and emission diversity that separates tissue types, and the channel count matters more than the selection rule ($k = 15$ uniform 0.48 against 0.71).

Caveats. One scene family with three strongly different tissue types; thickness is perfectly correlated with image column, so global fits are confounded by design; the binning slightly mixes class borders; the thinnest bins contain few pixels (about 40–90 per capture) and are noisy (GMM standard error up to 0.18).

Take-home. Label-free separation of tissue types degrades monotonically with burial (ARI $0.92 \rightarrow 0.08$) although the information remains in the spectra, and channel sets designed for thickness are not suited to tissue classification.

I. E10: Performance and convergence

Question. How does the cost and the noise of the 3D Monte Carlo scale with the photon count and the number of threads, what do the peel-off estimator and the voxel inclusions cost, and is the preview mode consistent with a full capture?

Setup. The E2 scene (a 0.5 mm collagen cover on a 40×40 mm block, 400 nm, preview grid 87×64), timing the Monte Carlo stage (one excitation pass and nine emission

TABLE IX
E9: ARI OF PER-BIN k-MEANS AGAINST TISSUE TYPE (MEAN OVER 14 CAPTURES) AGAINST LOCAL COVER THICKNESS; LDA IS THE SUPERVISED ACCURACY WITH BINNING.

thickness bin (mm)	0–0.05	0.05–0.1	0.1–0.2	0.2–0.35	0.35–0.6	0.6–1.0	1.0–1.5
k-means, 3×3 binning	0.92	0.91	0.58	0.43	0.33	0.19	0.08
k-means, raw pixels	0.72	0.60	0.43	0.30	0.19	0.12	0.05




Fig. 20. E9. *Top:* clustering against burial, (a) per-bin ARI of k-means with and without binning and supervised LDA accuracy, (b) per-bin ARI of fits on the whole capture, (c) whole-capture ARI and NMI against tissue type and thickness bin. *Bottom:* channel reduction, (a) label-free ARI and (b) supervised accuracy against thickness for 15 channels chosen by the E5 criterion, uniformly or at random, (c) mean ARI over 0–0.35 mm against k; the dashed line is the 217-channel value 0.71.

nodes) for $N = 10^4$–$6.4 \cdot 10^5$ packets per pass. Per-pixel noise is the spatial standard deviation of a plane-detrended 24×24 ROI; the ROI-mean error uses eight seeds. Speed-ups are paired over five interleaved repetitions, the peel-off and raw-cone estimators are compared at equal packets, the preview (10 % of the photons, 87×64) is compared with a full

capture (348×256, 4×4-binned) on the 0–2 mm wedge, and the inclusion overhead is measured on the E4 scene (sphere and vessel of $r = 1$ mm, and a sphere filled with host optics, $N = 10^5$, nine interleaved rounds). The machine was a shared workstation with a high load throughout, so absolute times are inflated; the ratios come from interleaved repetitions with interquartile ranges of a few percent.

Results. The per-pixel noise falls as $N^{-0.47}$ (95 % interval $[-0.48, -0.46]$) for the peel-off estimator and $N^{-0.50}$ for the raw cone estimator, i.e. as the expected $N^{-1/2}$ (Fig. 21b). The ROI-mean error does not follow it: its slope is -0.32 $[-0.45, -0.16]$ and it flattens at about 1 % above $N \approx 1.6{\cdot}10^5$; we hypothesise a seed-to-seed component that does not shrink with the packet count, which we did not test, so the $N^{-1/2}$ law is shown for per-pixel noise only. The runtime is 0.96 s +12.3 s per 10^5 packets for peel-off. Multithreading gives speed-ups of 1.88, 3.60, 6.44 and 7.10 for 2, 4, 8 and 14 threads (efficiency 0.94, 0.90, 0.80, 0.51; the last point mixes performance and efficiency cores on a loaded machine), and the output is bit-identical for 1, 2, 4, 8 and 14 threads in all repetitions. At equal packets the peel-off estimator is 2.1–2.4× slower per packet but has 2.9–3.4× lower per-pixel noise, so the noise-variance efficiency is 3.6–5.3× higher (Fig. 22a,b); at $N = 6.4{\cdot}10^5$ the cone estimator's noise is matched by peel-off with about $5.7 \cdot 10^4$ packets in 7 s against 33 s. The ROI mean of the cone estimator differs from that of peel-off by -0.09 % ($|z| < 1.3$ at every N), so no bias is detectable. The preview differs from a full capture by a relative rms of 0.32 against 0.13 between two full captures with different seeds; with the preview's own seed-to-seed difference (0.43) this is explained by Monte Carlo noise alone (predicted 0.32), with a ratio of sums of 0.998 (Fig. 21d). Voxel inclusions cost 1.99× (sphere), 1.74× (vessel) and 1.59× (host-filled sphere with no boundary physics) of the capture without an inclusion (Fig. 22c). The design target of 1.6× is therefore missed for real inclusions. The number of steps per photon is unchanged, and the host-filled sphere already costs 1.59×, so the per-step voxel lookup is the main cost; we did not profile the compiled kernels, and a coarse occupancy grid is a likely remedy.

Caveats. Timings were taken on one machine (Apple M4 Pro, 14 cores) under load; the per-pixel efficiency gain of peel-off refers to pixel noise, not to ROI-mean noise, where the data are too noisy to interpret; the preview comparison uses one scene, two seeds and one line; and the inclusion overhead was measured at one photon count and voxel size.

Take-home. The engine scales as $N^{-1/2}$ in per-pixel noise, is deterministic across thread counts and gains a factor of 3.6–5.3 in efficiency from peel-off; voxel inclusions cost about 1.6–2×, above the 1.6× target.

VIII. Discussion and Limitations

What the tool is for. The simulator is most useful where ground truth and controlled variation are needed: quantifying how a cover confounds a spectral ratio (E3), designing channel sets (E5), and benchmarking estimators and clustering methods (E7, E9). Whether its predictions transfer to real data depends on the fidelity of the twin, which Section VI-E documents, including that a non-parametric measured spectrum is as good a predictor as the calibrated model.

Bounds are not precision. The photon-limited CRLBs of E5 (sub-micron thickness) are bounds on information. Model bias dominates real retrieval: in E6 the RMSE is 160–260 µm at nominal exposure, 15–70× the photon-limited tile bound, and the nominal 95 % intervals cover only 5–19 % of the tiles. The model-error term of Eq. (21) brings the floor to 2–10 µm, still one to two orders below the observed errors, because it is a variance and not a bias. The relative error ϵ is an upper bound taken from sim-to-real residuals of collagen and droplets and is assumed independent across channels. The channel sets that are optimal depend on ϵ and on the estimand: sets chosen for thickness were worse than random for tissue classification (E9) and, on the phantom, for mapping (E6).

Monte Carlo noise. The 3D result is a Monte Carlo estimate (about 10 % rms per 0.5 mm face cell before smoothing; 2–3 % in a 3 mm region of interest), which is why studies report replicate noise floors. The 'fast' mode used for datasets has a per-pixel tally noise of about 39 % per channel (median over channels), which affects the pixel-level analyses: it produced the artefactual floor of the first E7 version and part of the held-out E7 errors, and it contributes to the degradation of unbinned clustering in E9 (a separate check gives about 4 % Monte Carlo and 7 % sensor noise in a channel-mean-normalised metric, so the two figures are not comparable). Sobol' indices were estimated with $N = 64$ base rows, so total indices are rank-level evidence only, and the group-index intervals are degenerate.

Experimental design caveats. E7 uses oracle endmember libraries built from the same pixels as the data; the held-out variant, which includes library-estimation noise, is the realistic one and is substantially worse beyond 0.4 mm. Most experiments use a single scene family and a single noise realisation (E6, E7), one ROI (E5, E8) or one design of 14 captures (E9), and the surrogate used in E5 and E6 is trained on uniform blocks, so lateral transport at the wedge slope is unmodelled. The lateral blur of the wedge (E2) could not be resolved above the noise. The 'autofluorescence fraction' of E2 uses the engine's own component bookkeeping and cannot be measured in real data without the cover alone. Voxel inclusions slow the 3D Monte Carlo by about 2× (the 1.6× design target was missed in E10), and they are index-matched to the host.

Physics limitations. Absorption and emission spectra of the dyes are parametric (Gaussians with skew and an optional vibronic shoulder), so there is no $S_0 \rightarrow S_2$ band (fluorescein absorbs nothing at 405 nm in the model against about 3 % of its peak in reality), no exact mirror image and no solvent shift. Scattering is a power law with one Henyey–Greenstein g per material, with no wavelength-dependent g, polarisation or coherence. Fluorescence has a single Stern–Volmer constant, FRET without overlap integral or orientation factor, one-generation isotropic secondary emission, only the 3400 cm^{-1} water Raman band, and no lifetimes, depolarisation, temperature dependence or triplet saturation. Bleaching in the 1D engine uses the surface fluence. Block pixels have no secondary emission (fluorescence re-absorbed in the block is lost, making it a few percent to about 10 % dimmer for strongly




Fig. 21. E10, performance and convergence. (a) Runtime against packets per pass for the peel-off and raw-cone estimators; (b) relative per-pixel Monte Carlo noise against N with the fitted log–log slopes in the legend (dashed lines: $N^{-1/2}$) and the ROI-mean error, which flattens at about 1 %; (c) thread speed-up against one thread (median and interquartile range, labels give the parallel efficiency); (d) preview against full capture, pixel by pixel.




Fig. 22. E10, estimators and inclusions. (a) Peel-off against raw cone at equal packets: per-pixel noise ratio (cone/peel) and time ratio (peel/cone); (b) efficiency gain of peel-off, $\sigma^2 T(\text{cone})/\sigma^2 T(\text{peel})$; (c) time ratio of the capture with an inclusion to that without (median and interquartile range); the dashed line is the 1.6× design target.

overlapping dyes), a single camera direction per block, planar interfaces and a rectangular footprint. Tissue pigments measured in organic solvents are proxies for the bound pigments. The bottom of the block is a Fresnel interface to air and the stage never reflects. Indirect diffuse interreflection between objects is not implemented. Fisher information is a local bound that assumes an unbiased estimator and neglects fixed-pattern noise; calibration intervals are statistical, and the twin is calibrated on thin, dilute-equivalent samples, so collagen and droplet parameters are effective rather than intrinsic.

Validation scope. The validation is against analytic results, MCML-type benchmarks, an independent Monte Carlo and one laboratory's measurements; we have not compared 3D inclusion or tissue-phantom predictions with phantom measurements, and the cover-scene check of E8 uses thin measured samples.

IX. Conclusion

SpectralForge Studio couples a literature-grade layered-tissue 3D Monte Carlo with voxel inclusions, dye photophysics and a camera model with a study engine, inverse

tools, information-theoretic design and calibration in one schema-driven, reproducible environment. The transport engine matches established benchmarks and an independent Monte Carlo to about one percent or better, and the calibrated twin predicts held-out excitation lines with a mean spectral angle of $8.4°$, though not better than a copy of a measured spectrum. In simulations of collagen-covered tissue, the cover rather than metabolism dominates the optical redox ratio in this model (first-order share 0.66 against 0.16) and simple ratio corrections do not remove it; about 30 of 248 channels carry the information of the full protocol on cover thickness, but model bias, not photon noise, limits retrieval, so photon-limited bounds should be reported together with a model-error floor; and unmixing needs depth-matched endmembers between 0.1 and 0.6 mm and fails beyond, where the signal is below the noise. Under a dense collagen cover, absorbing vessels are detectable per pixel only to about 0.2–0.45 mm at 340 nm and an endogenous-contrast fluorescent nodule is not detectable at all, and the engine's per-pixel noise follows $N^{-1/2}$ with bit-identical results across thread counts, although voxel inclusions cost 1.6–$2\times$ of an inclusion-free capture.

Code Availability

The source code, documentation, validation scripts and presets are available at https://github.com/Nard248/spectral-select.

References

[1] W. R. Zipfel, R. M. Williams, R. Christie, A. Y. Nikitin, B. T. Hyman, and W. W. Webb, "Live tissue intrinsic emission microscopy using multiphoton-excited native fluorescence and second harmonic generation," *Proceedings of the National Academy of Sciences*, vol. 100, no. 12, pp. 7075–7080, 2003.

[2] I. Georgakoudi and K. P. Quinn, "Optical Imaging Using Endogenous Contrast to Assess Metabolic State," *Annual Review of Biomedical Engineering*, vol. 14, no. 1, pp. 351–367, 2012.

[3] A. Croce and G. Bottiroli, "Autofluorescence spectroscopy and imaging: a tool for biomedical research and diagnosis," *European Journal of Histochemistry*, 2014.

[4] N. Ramanujam, "Fluorescence Spectroscopy of Neoplastic and Non-Neoplastic Tissues," *Neoplasia*, vol. 2, no. 1-2, pp. 89–117, 2000.

[5] M. Mercader, L. Swift, S. Sood, H. Asfour, M. Kay, and N. Sarvazyan, "Use of endogenous NADH fluorescence for real-time in situ visualization of epicardial radiofrequency ablation lesions and gaps," *American Journal of Physiology-Heart and Circulatory Physiology*, vol. 302, no. 10, pp. H2131–H2138, 2012.

[6] D. A. Gil, L. M. Swift, H. Asfour, N. Muselimyan, M. A. Mercader, and N. A. Sarvazyan, "Autofluorescence hyperspectral imaging of radiofrequency ablation lesions in porcine cardiac tissue," *Journal of Biophotonics*, vol. 10, no. 8, pp. 1008–1017, 2017.

[7] S. Guan, H. Asfour, N. Sarvazyan, and M. Loew, "Application of unsupervised learning to hyperspectral imaging of cardiac ablation lesions," *Journal of Medical Imaging*, vol. 5, no. 04, pp. 1–12, 2018.

[8] N. Chilingaryan, F. Villarruel, T. Soghomonyan, V. Yeghiazaryan, and N. Sarvazyan, "4D hyperspectral imaging for intraoperative tissue classification," in *Medical Imaging 2025: Clinical and Biomedical Imaging*, 2025, p. 35.

[9] N. Chilingaryan, A. Gasparyan, and N. Sarvazyan, "Multi-Excitation Hyperspectral Imaging Toward Improved Lichen Identification," in *2025 15th Workshop on Hyperspectral Imaging and Signal Processing: Evolution in Remote Sensing (WHISPERS)*, 2025, pp. 1–5.

[10] A. Hunanyan, N. Ter-Petrosyan, F. Villarruel, T. Soghomonyan, N. Sarvazyan, A. Butavyan, and V. Yeghiazaryan, "Deep Learning Classification of Ablated Atrial Tissue in Multi-Excitation Hyperspectral Images," in *2025 15th Workshop on Hyperspectral Imaging and Signal Processing: Evolution in Remote Sensing (WHISPERS)*, 2025, pp. 1–5.

[11] M. C. Skala, K. M. Riching, A. Gendron-Fitzpatrick, J. Eickhoff, K. W. Eliceiri, J. G. White, and N. Ramanujam, "In vivo multiphoton microscopy of NADH and FAD redox states, fluorescence lifetimes, and cellular morphology in precancerous epithelia," *Proceedings of the National Academy of Sciences*, vol. 104, no. 49, pp. 19 494–19 499, 2007.

[12] J. Wu, M. S. Feld, and R. P. Rava, "Analytical model for extracting intrinsic fluorescence in turbid media," *Applied Optics*, vol. 32, no. 19, p. 3585, 1993.

[13] M. G. Müller, I. Georgakoudi, Q. Zhang, J. Wu, and M. S. Feld, "Intrinsic fluorescence spectroscopy in turbid media: disentangling effects of scattering and absorption," *Applied Optics*, vol. 40, no. 25, p. 4633, 2001.

[14] L. Wang, S. L. Jacques, and L. Zheng, "MCML—Monte Carlo modeling of light transport in multi-layered tissues," *Computer Methods and Programs in Biomedicine*, vol. 47, no. 2, pp. 131–146, 1995.

[15] Q. Fang and D. A. Boas, "Monte Carlo Simulation of Photon Migration in 3D Turbid Media Accelerated by Graphics Processing Units," *Optics Express*, vol. 17, no. 22, p. 20178, 2009.

[16] H. Shen and G. Wang, "A tetrahedron-based inhomogeneous Monte Carlo optical simulator," *Physics in Medicine and Biology*, vol. 55, no. 4, pp. 947–962, 2010.

[17] A. A. Leino, A. Pulkkinen, and T. Tarvainen, "ValoMC: a Monte Carlo software and MATLAB toolbox for simulating light transport in biological tissue," *OSA Continuum*, vol. 2, no. 3, p. 957, 2019.

[18] A. A. Goodenough and S. D. Brown, "DIRSIG5: Next-Generation Remote Sensing Data and Image Simulation Framework," *IEEE Journal of Selected Topics in Applied Earth Observations and Remote Sensing*, vol. 10, no. 11, pp. 4818–4833, 2017.

[19] N. Meloyan and N. Sarvazyan, "SpectralForge: A physics-based simulator for fluorescence hyperspectral imaging," in *Proc. 16th Workshop on Hyperspectral Imaging and Signal Processing: Evolution in Remote Sensing (WHISPERS)*, 2026, iEEE, accepted.

[20] L. C. Henyey and J. L. Greenstein, "Diffuse radiation in the Galaxy," *The Astrophysical Journal*, vol. 93, p. 70, 1941.

[21] S. L. Jacques, "Monte Carlo Modeling of Light Transport in Tissue (Steady State and Time of Flight)," in *Optical-Thermal Response of Laser-Irradiated Tissue*. Springer Netherlands, 2011, pp. 109–144.

[22] S. L. Jacques and B. W. Pogue, "Tutorial on diffuse light transport," *Journal of Biomedical Optics*, vol. 13, no. 04, pp. 1–19, 2008.

[23] E. Alerstam, T. Svensson, and S. Andersson-Engels, "Parallel computing with graphics processing units for high-speed Monte Carlo simulation of photon migration," *Journal of Biomedical Optics*, vol. 13, no. 06, pp. 1–3, 2008.

[24] D. A. Boas, J. P. Culver, J. J. Stott, and A. K. Dunn, "Three dimensional Monte Carlo code for photon migration through complex heterogeneous media including the adult human head," *Optics Express*, vol. 10, no. 3, p. 159, 2002.

[25] Q. Fang, "Mesh-based Monte Carlo method using fast ray-tracing in Plücker coordinates," *Biomedical Optics Express*, vol. 1, no. 1, p. 165, 2010.

[26] S. Yan and Q. Fang, "Hybrid mesh and voxel based Monte Carlo algorithm for accurate and efficient photon transport modeling in complex bio-tissues," *Biomedical Optics Express*, vol. 11, no. 11, p. 6262, 2020.

[27] A. J. Welch, C. Gardner, R. Richards-Kortum, E. Chan, G. Criswell, J. Pfefer, and S. Warren, "Propagation of fluorescent light," *Lasers in Surgery and Medicine*, vol. 21, no. 2, pp. 166–178, 1997.

[28] S. Jacques, "Monte Carlo Simulations Of Fluorescence In Turbid Media," in *Handbook of Biomedical Fluorescence*. CRC Press, 2003.

[29] J. Swartling, A. Pifferi, A. M. K. Enejder, and S. Andersson-Engels, "Accelerated Monte Carlo models to simulate fluorescence spectra from layered tissues," *Journal of the Optical Society of America A*, vol. 20, no. 4, p. 714, 2003.

[30] A. Liebert, H. Wabnitz, N. Zolek, and R. Macdonald, "Monte Carlo algorithm for efficient simulation of time-resolved fluorescence in layered turbid media," *Optics Express*, vol. 16, no. 17, p. 13188, 2008.

[31] J. Tobin, R. Fong, A. Ray, J. Schneider, W. Zaremba, and P. Abbeel, "Domain randomization for transferring deep neural networks from simulation to the real world," in *2017 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS)*, 2017, pp. 23–30.

[32] R. Bro, "PARAFAC. Tutorial and applications," *Chemometrics and Intelligent Laboratory Systems*, vol. 38, no. 2, pp. 149–171, 1997.

[33] K. R. Murphy, C. A. Stedmon, D. Graeber, and R. Bro, "Fluorescence spectroscopy and multi-way techniques. PARAFAC," *Analytical Methods*, vol. 5, no. 23, p. 6557, 2013.

[34] T. Zimmermann, "Spectral Imaging and Linear Unmixing in Light Microscopy," in *Advances in Biochemical Engineering/Biotechnology*. Springer Berlin Heidelberg, 2005, pp. 245–265.

[35] N. Keshava and J. Mustard, "Spectral unmixing," *IEEE Signal Processing Magazine*, vol. 19, no. 1, pp. 44–57, 2002.

[36] J. M. Bioucas-Dias, A. Plaza, N. Dobigeon, M. Parente, Q. Du, P. Gader, and J. Chanussot, "Hyperspectral Unmixing Overview: Geometrical, Statistical, and Sparse Regression-Based Approaches," *IEEE Journal of Selected Topics in Applied Earth Observations and Remote Sensing*, vol. 5, no. 2, pp. 354–379, 2012.

[37] D. C. Heinz and C.-I. Chang, "Fully constrained least squares linear spectral mixture analysis method for material quantification in hyperspectral imagery," *IEEE Transactions on Geoscience and Remote Sensing*, vol. 39, no. 3, pp. 529–545, 2001.

[38] C. L. Lawson and R. J. Hanson, *Solving Least Squares Problems*. Society for Industrial and Applied Mathematics, 1995.

[39] D. D. Lee and H. S. Seung, "Learning the parts of objects by non-negative matrix factorization," *Nature*, vol. 401, no. 6755, pp. 788–791, 1999.

[40] J. Nascimento and J. Dias, "Vertex component analysis: a fast algorithm to unmix hyperspectral data," *IEEE Transactions on Geoscience and Remote Sensing*, vol. 43, no. 4, pp. 898–910, 2005.

[41] H. R. Morris, C. C. Hoyt, and P. J. Treado, "Imaging Spectrometers for Fluorescence and Raman Microscopy: Acousto-Optic and Liquid Crystal Tunable Filters," *Applied Spectroscopy*, vol. 48, no. 7, pp. 857–866, 1994.

[42] G. Lu and B. Fei, "Medical hyperspectral imaging: a review," *Journal of Biomedical Optics*, vol. 19, no. 01, pp. 1–24, 2014.

[43] A. M. Valm, S. Cohen, W. R. Legant, J. Melunis, U. Hershberg, E. Wait, A. R. Cohen, M. W. Davidson, E. Betzig, and J. Lippincott-Schwartz, "Applying systems-level spectral imaging and analysis to reveal the organelle interactome," *Nature*, vol. 546, no. 7656, pp. 162–167, 2017.

[44] R. Richards-Kortum and E. Sevick-Muraca, "Quantitative optical spectroscopy for tissue diagnosis," *Annual Review of Physical Chemistry*, vol. 47, no. 1, pp. 555–606, 1996.

[45] I. Sobol', "Global sensitivity indices for nonlinear mathematical models and their Monte Carlo estimates," *Mathematics and Computers in Simulation*, vol. 55, no. 1-3, pp. 271–280, 2001.

[46] A. Saltelli, P. Annoni, I. Azzini, F. Campolongo, M. Ratto, and S. Tarantola, "Variance based sensitivity analysis of model output. Design and estimator for the total sensitivity index," *Computer Physics Communications*, vol. 181, no. 2, pp. 259–270, 2010.

[47] M. J. Jansen, "Analysis of variance designs for model output," *Computer Physics Communications*, vol. 117, no. 1-2, pp. 35–43, 1999.

[48] M. D. Morris, "Factorial Sampling Plans for Preliminary Computational Experiments," *Technometrics*, vol. 33, no. 2, pp. 161–174, 1991.

[49] F. Campolongo, J. Cariboni, and A. Saltelli, "An effective screening design for sensitivity analysis of large models," *Environmental Modelling & Software*, vol. 22, no. 10, pp. 1509–1518, 2007.

[50] F. Pukelsheim, *Optimal Design of Experiments*. Society for Industrial and Applied Mathematics, 2006.

[51] G. L. Nemhauser, L. A. Wolsey, and M. L. Fisher, "An analysis of approximations for maximizing submodular set functions—I," *Mathematical Programming*, vol. 14, no. 1, pp. 265–294, 1978.

[52] C.-I. Chang, Q. Du, T.-L. Sun, and M. L. G. Althouse, "A joint band prioritization and band-decorrelation approach to band selection for hyperspectral image classification," *IEEE Transactions on Geoscience and Remote Sensing*, vol. 37, no. 6, pp. 2631–2641, 1999.

[53] W. Sun and Q. Du, "Hyperspectral Band Selection: A Review," *IEEE Geoscience and Remote Sensing Magazine*, vol. 7, no. 2, pp. 118–139, 2019.

[54] Y. Cai, X. Liu, and Z. Cai, "BS-Nets: An End-to-End Framework for Band Selection of Hyperspectral Image," *IEEE Transactions on Geoscience and Remote Sensing*, vol. 58, no. 3, pp. 1969–1984, 2020.

[55] N. Meloyan and N. Sarvazyan, "Unsupervised deep learning for wavelength selection in multi-excitation fluorescence imaging," *Scientific Reports*, 2025, under revision.

[56] A. Rasheed, O. San, and T. Kvamsdal, "Digital Twin: Values, Challenges and Enablers From a Modeling Perspective," *IEEE Access*, vol. 8, pp. 21 980–22 012, 2020.

[57] E. Katsoulakis, Q. Wang, H. Wu, L. Shahriyari, R. Fletcher, J. Liu, L. Achenie, H. Liu, P. Jackson, Y. Xiao, T. Syeda-Mahmood, R. Tuli, and J. Deng, "Digital twins for health: a scoping review," *npj Digital Medicine*, vol. 7, no. 1, p. 77, 2024.

[58] R. L. Cook and K. E. Torrance, "A Reflectance Model for Computer Graphics," *ACM Transactions on Graphics*, vol. 1, no. 1, pp. 7–24, 1982.

[59] B. Smith, "Geometrical shadowing of a random rough surface," *IEEE Transactions on Antennas and Propagation*, vol. 15, no. 5, pp. 668–671, 1967.

[60] L. Song, E. Hennink, I. Young, and H. Tanke, "Photobleaching kinetics of fluorescein in quantitative fluorescence microscopy," *Biophysical Journal*, vol. 68, no. 6, pp. 2588–2600, 1995.

[61] C. Eggeling, J. Widengren, R. Rigler, and C. A. M. Seidel, "Photobleaching of Fluorescent Dyes under Conditions Used for Single-Molecule Detection: Evidence of Two-Step Photolysis," *Analytical Chemistry*, vol. 70, no. 13, pp. 2651–2659, 1998.

[62] E. Alerstam, S. Andersson-Engels, and T. Svensson, "White Monte Carlo for time-resolved photon migration," *Journal of Biomedical Optics*, vol. 13, no. 04, pp. 1–10, 2008.

[63] S. L. Jacques, "Optical properties of biological tissues: a review," *Physics in Medicine and Biology*, vol. 58, no. 11, pp. R37–R61, 2013.

[64] R. M. Pope and E. S. Fry, "Absorption spectrum (380–700 nm) of pure water II Integrating cavity measurements," *Applied Optics*, vol. 36, no. 33, p. 8710, 1997.

[65] G. M. Hale and M. R. Querry, "Optical Constants of Water in the 200-nm to 200-μm Wavelength Region," *Applied Optics*, vol. 12, no. 3, p. 555, 1973.

[66] R. L. van Veen, H. J. Sterenborg, A. Pifferi, A. Torricelli, E. Chikoidze, and R. Cubeddu, "Determination of visible near-IR absorption coefficients of mammalian fat using time- and spatially resolved diffuse reflectance and transmission spectroscopy," *Journal of Biomedical Optics*, vol. 10, no. 05, pp. 1–6, 2005.

[67] H. J. van Staveren, C. J. M. Moes, J. van Marie, S. A. Prahl, and M. J. C. van Gemert, "Light scattering in Intralipid-10% in the wavelength range of 400–1100 nm," *Applied Optics*, vol. 30, no. 31, p. 4507, 1991.

[68] R. Michels, F. Foschum, and A. Kienle, "Optical properties of fat emulsions," *Optics Express*, vol. 16, no. 8, p. 5907, 2008.

[69] E. Veach and L. J. Guibas, "Optimally combining sampling techniques for Monte Carlo rendering," in *Proceedings of the 22nd annual conference on Computer graphics and interactive techniques - SIGGRAPH '95*, 1995, pp. 419–428.

[70] J. R. Janesick, *Photon Transfer*. SPIE, 2007.

[71] G. Healey and R. Kondepudy, "Radiometric CCD camera calibration and noise estimation," *IEEE Transactions on Pattern Analysis and Machine Intelligence*, vol. 16, no. 3, pp. 267–276, 1994.

[72] D. J. Venzon and S. H. Moolgavkar, "A Method for Computing Profile-Likelihood-Based Confidence Intervals," *Applied Statistics*, vol. 37, no. 1, p. 87, 1988.

[73] G. H. Golub and V. Pereyra, "The Differentiation of Pseudo-Inverses and Nonlinear Least Squares Problems Whose Variables Separate," *SIAM Journal on Numerical Analysis*, vol. 10, no. 2, pp. 413–432, 1973.

[74] T. J. Farrell, M. S. Patterson, and B. Wilson, "A diffusion theory model of spatially resolved, steady-state diffuse reflectance for the noninvasive determination of tissue optical properties in vivo," *Medical Physics*, vol. 19, no. 4, pp. 879–888, 1992.

[75] M. S. Patterson, B. Chance, and B. C. Wilson, "Time resolved reflectance and transmittance for the noninvasive measurement of tissue optical properties," *Applied Optics*, vol. 28, no. 12, p. 2331, 1989.

[76] M. Kubista, R. Sjöback, S. Eriksson, and B. Albinsson, "Experimental correction for the inner-filter effect in fluorescence spectra," *The Analyst*, vol. 119, no. 3, pp. 417–419, 1994.