الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No. 2055011643
س. ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/116836

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	**Sudhir Equipment Rental Company**
Location	Dammam, Kingdom of Saudi Arabia
Description	**Pressure Relief Valve**
Manufacturer	HoneyWell
Type	N/A
Serial No.	N/A
ID. No.	PRV-116836
Location	Calibration Lab (Dammam)

Receipt Date	:	09-Nov-26
Calibration Date	:	09-Nov-26
Calibration Due Date	:	08-May-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224313	2232411017	02-Dec-26

Calibrated By

Calibration Engineer

QA/QC

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa
ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+

CERTIFICATE NO: UIC/DAM/116836

Customer	**Sudhir Equipment Rental Company**
Location	Dammam, Kingdom of Saudi Arabia
Description	**Pressure Relief Valve**
Manufacturer	HoneyWell
Type	N/A
Serial No.	N/A
ID. No.	PRV-116836
Calibration Date	09-Nov-26
Calibration Due Date	08-May-27
Temperature Range	23.5°C
Relative Humidity	RH 48%

Test Results

Set Pressure	:	**12.00**	**BAR**
Back Pressure	:	**0.00**	**BAR**
CDTP	:	**12.00**	**BAR**
Setting Accuracy	:	**± 3 % of Setting**	
Pop Test : PASSED			

Spring Range : N/A

Size Inlet/Outlet : 1"

Type of Valve : Metal Seated

Test Medium : Air

BASIS OF TEST:

Ref Standard ASME Sec VIII,Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pnenumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceabilty to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

QA/QC — الشركة العالمية لخدمات التفتيش المحدودة — UNIVERSAL INSPECTION CO. LTD.

++++ End of the Certificate ++++

UIC 67 CAL ENGR

Calibrated By: _____
Calibration Engineer

Approved by: _____
QA/QC Engineer