الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No. 2055011643
س. ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/116837

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	**Sudhir Equipment Rental Company**			
Location	Dammam, Kingdom of Saudi Arabia			
Description	**Pressure Gauge**			
Manufacturer	TIWAC			
Model	N/A			
Serial No.	N/A	**Receipt Date**	:	09-Nov-26
ID. No.	SAC-500-447	**Calibration Date**	:	09-Nov-26
Location	Calibration Lab (Dammam)	**Calibration Due Date**	:	08-May-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224313	2232411017	02-Dec-26

Calibrated By

67 CAL ENGR

Calibration Engineer

QA/QC

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa
ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+

CERTIFICATE NO: UIC/DAM/116837

Customer	**Sudhir Equipment Rental Company**		
Location	Dammam, Kingdom of Saudi Arabia		
Description	**Pressure Gauge**		
Manufacturer	TIWAC	**Calibration Date**	09-Nov-26
Model	N/A	**Calibration Due Date**	08-May-27
Serial No.	N/A	**Temperature Range**	23.5°C
ID. No.	SAC-500-447	**Relative Humidity**	RH 48%

Calibration Test Results for Pressure Gauge

Min. Range: **0.0 BAR** **Max. Range:** **25.0 BAR**

Function Tested	Gauge Pressure		
UUT Setting BAR	Actual Reading in Master BAR	Error BAR	Error (FS) (% of Range)
0.00	0.00	0.000	0.000
5.00	5.02	0.020	0.080
10.00	10.05	0.050	0.200
15.00	15.06	0.060	0.240
20.00	20.05	0.050	0.200
25.00	25.08	0.080	0.320

Calibration Procedure:

UIC/P/PPGCP Rev. 00

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated.
The item was received in good condition.

QA/QC

\+ \+ \+ \+ \+ End of the Certificate \+ \+ \+ \+ \+

UIC 67 CAL ENGR

Calibrated By: _____
Calibration Engineer

Approved by: _____
QA/QC Engineer