



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/RIY/37948

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company.		
Address	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Gauge		
Manufacturer	DOOSAN	Receipt Date	: 15-Jan-26
Model	N/A	Calibration Date	: 15-Jan-26
Serial No.	000392BCAGJ01	Report Date	: 15-Jan-26
ID. No.	SAC-185-045	Issue date	: 15-Jan-26
Location	Calibration Lab	Calibration Due Date	: 14-Jul-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model/Cert No	Serial No	Due Date
High Pressure Pneumatic Calibrator	SA178189-223867	2011H13013	23-Nov-26
Digital Pressure Calibrator	SA178189-223866	223320200723	26-Nov-26

Calibrated By
13
CAL ENGR
Calibration Engineer



Approved By
QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ / فاكس ٨٨٨١ ٣٤٤ ١٣ +٩٦٦
هاتف ٥١٥ ٣٦٧ ١٣ +٩٦٦

www.ui.com.sa

CERTIFICATE NO: UIC/RIY/37948

Customer **Sudhir Equipment Rental Company.**

Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Gauge**

Manufacturer DOOSAN

Model N/A

Serial No. 000392BCAGJ01

ID. No. SAC-185-045

Calibration Date 15-Jan-26

Calibration Due Date 14-Jul-26

Temperature $23.0 \pm 2^{\circ}\text{C}$

Relative Humidity RH 45% to 75%

CALIBRATION RESULTS FOR PRESSURE GAUGE

Min. Range: 0.0 (PSI)

Max. Range: 200.0 (PSI)

Function Tested		Gauge Pressure	
UUT Setting (PSI)	Actual reading in Master (PSI)	Error (PSI)	Error (% of Range) (FS)
0.00	0.00	0.00	0.000
20.00	20.45	0.45	0.225
100.00	101.66	1.66	0.830
160.00	162.45	2.45	1.225
200.00	202.74	2.74	1.370

Calibration Procedure:

UIC/P/PPGCP Rev. 02

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated.

The item was received in good condition.

Calibrated By: _____
Calibration EngineerApproved by: _____
QA/QC Engineer