



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/RIY/36430

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company.		
Address	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Relief Valve		
Manufacturer	N/A	Receipt Date	: 19-Nov-25
Model	SN501/2	Calibration Date	: 19-Nov-25
Serial No.	N/A	Report Date	: 19-Nov-25
Id No.	SERC-PRV-001	Issue date	: 19-Nov-25
Location	Calibration Lab	Calibration Due Date	: 18-May-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No.	Due Date
Water Pressure Comparator	VIC-25-42247	990S2412008	04-May-26
Digital Pressure Gauge	VIC-25-57095	2232412038	04-May-26

Calibrated By
13
CAL-ENGR
Calibration Engineer



UIC-QP-7.8.2-17-REV.02

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١
هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO : UIC/RIY/36430

Customer
Address**Sudhir Equipment Rental Company.**

Dammam, Kingdom of Saudi Arabia

Description

Pressure Relief Valve

Calibration Date

19-Nov-25

Manufacturer

N/A

Calibration Due Date

18-May-26

Model

SN501/2

Temperature

23.0 ± 2°C

Serial No.

N/A

Relative Humidity

RH 45% to 75%

Id No.

SERC-PRV-001

Test Results

Set Pressure : 180.0 PSI

Spring Range : N/A

Back Pressure: 0.0 PSI

Size Inlet / Outlet : 1/2

CDTP : 180.0 PSI

Setting Accuracy : ± 3 % of Setting

Type of Valve : Metal Seated

Pop Test : PASSED

Test Medium : Air/Water/Oil/Gas

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

No

The instrument was adjusted

Yes

The instrument was repaired

No

++++ End of the Certificate +++++

Calibrated By:



Calibration Engineer

UIC-OP-7.8.2-17,REV 02

Approved by:

QA/QC Engineer