



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت. ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/RIY/28199

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company.**

Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer	N/A	Receipt Date	: 22-Feb-25
Part No.	N/A	Calibration Date	: 25-Feb-25
Serial No.	MRZ-PRV-7915	Report Date	: 25-Feb-25
ID No.	N/A	Issue date	: 25-Feb-25
Location	Calibration Lab	Calibration Due Date	: 24-Aug-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No.	Due Date
High Pressure Pneumatic Calibrator	VIC-24-41692	2011H13013	15-Sep-25
Digital Pressure Calibrator	VIC-24-56304	223320200723	23-Nov-25

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب. ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٠٥١٥ ١٣ ٣٦٧ ٩٦٦

CERTIFICATE NO : UIC/RIY/28199

PAGE

1 of 2

Customer **Sudhir Equipment Rental Company.**
Address Dammam, Kingdom of Saudi Arabia

Description	Pressure Relief Valve	Calibration Date	25-Feb-25
Manufacturer	N/A	Calibration Due Date	24-Aug-25
Model	N/A	Temperature	23.0 ± 2°C
Serial No.	MRZ-PRV-7915	Relative Humidity	RH 45% to 75%

Test Results

Set Pressure :	200.0	PSI	Spring Range	: 101 - 225 PSI
Back Pressure:	0.0	PSI	Size Inlet / Outlet	: 3/4" X 1"
CDTP :	200.0	PSI	Type of Valve	: Metal Seated
Setting Accuracy :	± 3 % of Setting		Test Medium	: Air/Water/Oil/Gas
Pop Test :	PASSED			

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

++++ End of the Certificate +++++



UIC/OP/7.5.2-17, REV 02

Calibrated By: _____

Approved by: _____

Calibration Engineer