



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No.2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/JED/63270

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	SUDHIR EQUIPMENT RENTAL COMPANY.		
Address	JEDDAH - KINGDOM OF SAUDI ARABIA		
Description	SAFETY RELIEF VALVE		
Manufacturer	DOOSAN		
Part	54731450	Receipt Date	: 18-May-25
Serial No	2209003	Calibration Date	: 18-May-25
Location	Calibration Lab (Jeddah)	Calibration Due Date	: 17-Nov-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Certificate No.	Serial No.	Due Date
High Pressure Pneumatic Calibrator	VIC-24-41690	213E170502	14-Sep-25
Digital Pressure Gauge	VIC-24-55765	700140406	14-Sep-25

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

PAGE : 1 of 2

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٣٤٤ / فاكس ٨٨٨١ ١٣ ٣٤٤ +٩٦٦
هاتف ٠٥١٥ ١٣ ٣٦٧ +٩٦٦

CERTIFICATE NO :		UIC/JED/63270		PAGE	2 of 2
Customer	SUDHIR EQUIPMENT RENTAL COMPANY.				
Address	JEDDAH - KINGDOM OF SAUDI ARABIA				
Description	SAFETY RELIEF VALVE	Calibration Date	18-May-25		
Manufacturer	DOOSAN	Calibration Due Date	17-Nov-25		
Part	54731450	Temperature	23.6°C		
Serial No	2209003	Relative Humidity	47%		
<u>Test Results</u>					
Set Pressure :	250.0 PSI	Spring No./ Press. Range: -----			
Back Pressure:	0.0 PSI	Size Inlet / Orifice : ¾"× 1"			
CDTP :	250.0 PSI	Type of Valve : Metal Seated			
Setting Accuracy : ± 3 % of Setting		Test Medium : Air/Gas/Water/Oil			
Pop Test : PASSED					
<u>BASIS OF TEST:</u>					
Ref Standard ASME Sec VIII, Div I					
<u>Summary of Test Procedure:</u>					
The Calibration of above Safety relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.					
<u>Traceability:</u>					
The reported tests were undertaken using equipment detailed below for which the company maintains Traceability to the third party.					
<u>Condition of Instrument</u>					
With in the tolerance on receipt(as found)	No				
The instrument was adjusted	Yes				
The instrument was repaired	No				
***** End of the Certificate *****					
					
Calibrated By:		UIC/OP-7-8-24-F02, REV 0		Approved by:	
Calibration Engineer				QA/QC Engineer	