



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/RIY/41565

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company.		
Address	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Relief Valve		
Manufacturer	N/A	Receipt Date	: 24-Mar-26
Model No.	N/A	Calibration Date	: 24-Mar-26
Serial No.	N/A	Report Date	: 24-Mar-26
ID No.	SAC-825-065	Issue date	: 24-Mar-26
Location	Calibration Lab	Calibration Due Date	: 23-Sep-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model/Cert No	Serial No	Due Date
High Pressure Pneumatic Calibrator	SA178189-223867	2011H13013	23-Nov-26
Digital Pressure Calibrator	SA178189-223866	223320200723	26-Nov-26

Calibrated By
13
CAL ENGR
Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax.: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٠٥١٥ ١٣ ٩٦٦

CERTIFICATE NO : UIC/RIY/41565		PAGE	1 of 2
Customer Sudhir Equipment Rental Company. Address Dammam, Kingdom of Saudi Arabia			
Description	Pressure Relief Valve	Calibration Date	24-Mar-26
Manufacturer	N/A	Calibration Due Date	23-Sep-26
Model	N/A	Temperature	23.0 ± 2°C
Serial No.	N/A	Relative Humidity	RH 45% to 75%
ID No.	SAC-825-065		
<u>Test Results</u>			
Set Pressure :	10.3 BAR	Spring Range	: N/A
Back Pressure:	0.0 BAR	Size Inlet / Outlet	: N/A
CDTP :	10.3 BAR	Type of Valve	: Metal Seated
Setting Accuracy : ± 3 % of Setting		Test Medium	: Air
Pop Test : PASSED			
<u>BASIS OF TEST:</u> Ref Standard ASME Sec VIII, Div I			
<u>Summary of Test Procedure:</u> The Calibration of above Pressure relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.			
<u>Traceability:</u> The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceability to the third party.			
<u>Condition of Instrument</u>			
With in the tolerance on receipt(as found)		No	
The instrument was adjusted		Yes	
The instrument was repaired		No	
<u>Result:</u> Calibration Result is pass.			
 End of the Certificate		Approved by:  QA/QC Engineer	
Calibrated By:  Calibration Engineer			