



# الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س. ت ٢٠٥٥٠١١٦٤٣

## INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

### CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/RIY/37932

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

|              |                                         |                      |             |
|--------------|-----------------------------------------|----------------------|-------------|
| Customer     | <b>Sudhir Equipment Rental Company.</b> |                      |             |
| Location     | Dammam, Kingdom of Saudi Arabia         |                      |             |
| Description  | <b>Pressure Relief Valve</b>            |                      |             |
| Manufacturer | SMT                                     | Receipt Date         | : 14-Jan-26 |
| Model No.    | SN50                                    | Calibration Date     | : 14-Jan-26 |
| Serial No.   | 068703                                  | Report Date          | : 14-Jan-26 |
| ID No.       | N/A                                     | Issue date           | : 14-Jan-26 |
| Location     | Calibration Lab                         | Calibration Due Date | : 13-Jul-26 |

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

#### Reference Standard Equipment Used

| Description               | Model / Cert. No. | Serial No.  | Due Date |
|---------------------------|-------------------|-------------|----------|
| Water Pressure Comparator | VIC-25-42247      | 990S2412008 | 4-May-26 |
| Digital Pressure Gauge    | VIC-25-57095      | 2232412038  | 4-May-26 |

Calibrated By



Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO : UIC/RIY/37932

Customer  
Address**Sudhir Equipment Rental Company.**  
Dammam, Kingdom of Saudi ArabiaDescription  
Manufacturer  
Model  
Serial No.  
ID No.**Pressure Relief Valve**  
SMT  
SN50  
068703  
N/ACalibration Date  
Calibration Due Date  
Temperature  
Relative Humidity  
14-Jan-26  
13-Jul-26  
23.0 ± 2°C  
RH 45% to 75%Test ResultsSet Pressure : 13.0 BAR  
Back Pressure: 0.0 BAR  
CDTP : 13.0 BAR

Spring Range : 200 PSI

Size Inlet/Orif Dia : 1/2"

Setting Accuracy : ± 3 % of Setting

Type of Valve : Metal Seated

Pop Test : PASSED

Test Medium : WATER

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceability to the third party.

Condition of Instrument

With in the tolerance on receipt( as found )

No

The instrument was adjusted

No

The instrument was repaired

No

++++ End of the Certificate +++++

Calibrated By:



Calibration Engineer

UIC-QP-718.2-17-REV 02



Approved by:

QA/QC Engineer