



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/RIY/30087

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company.**

Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer STLMO

Model SN501/2NPT

Serial No. N/A

ID No. 22885

Location Calibration Lab

Receipt Date : 11-May-25

Calibration Date : 12-May-25

Report Date : 12-May-25

Issue date : 12-May-25

Calibration Due Date : 11-Nov-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model	Serial No.	Due Date
High Pressure Pneumatic Calibrator	VIC-24-41692	2011H13013	15-Sep-25
Digital Pressure Calibrator	VIC-24-56304	223320200723	23-Nov-25

Calibrated By

Calibration Engineer

UIC-QP-7.8.2-17, REV 02

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO : UIC/RIY/30087

PAGE

1 of 2

Customer	Sudhir Equipment Rental Company.		
Address	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Relief Valve	Calibration Date	12-May-25
Manufacturer	STLMO	Calibration Due Date	11-Nov-25
Model	SN501/2NPT	Temperature	23.0 ± 2°C
Serial No.	N/A	Relative Humidity	RH 45% to 75%
ID No.	22885		

Test Results

Set Pressure :	190.0	PSI	Spring Range	: N/A
Back Pressure:	0.0	PSI	Size Inlet	: 1/4"
CDTP :	190.0	PSI	Type of Valve	: Metal Seated
Setting Accuracy : ± 3 % of Setting			Test Medium	: Air
Pop Test : PASSED				

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	No
The instrument was repaired	No



Calibrated By: _____
Calibration Engineer

++++ End of the Certificate +++++



UIC-QP-7.8.2-17-REV 02

Approved by: _____

QA/QC Engineer