



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س. ت ١١٦٤٣ ٢٠٥٥

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/ALH/15738

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer Sudhir Equipment Rental Company

Address Al Hasa, Kingdom Of Saudi Arabia.

Description Pressure Relief Valve

Manufacturer Atlas Cobco

part No. 0832100066

Serial No. 2218054

Location Calibration Lab (Al Hasa)

Receipt Date : 6-Apr-25

Calibration Date : 7-Apr-25

Calibration Due Date : 6-Oct-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-210076	214E2207012	20-Apr-25
Digital Pressure Calibrator	SA178189-210077	223210304	20-Apr-25

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

[Signature]

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦

هاتف ٥١٥ ١٣ ٩٦٦

CERTIFICATE NO: UIC/ALH/15738		Page 2 of 2
Customer Address	Sudhir Equipment Rental Company Al Hasa, Kingdom Of Saudi Arabia.	
Description	Pressure Relief Valve	Calibration Date 7-Apr-25
Manufacturer	Atlas Cobco	Calibration Due Date 6-Oct-25
part No.	0832100066	Temperature 23.2°C
Serial No.	2218054	Relative Humidity RH 48 %
<u>Test Results</u>		
Set Pressure	: 230.00 PSI	Spring Range : N/A
Back Pressure	: 0.00 PSI	Size Inlet x Outlet : 1" x 1½"
CDTP	: 230.00 PSI	Type of Valve : Metal Seated
Setting Accuracy	: ± 3 % of Setting	Test Medium : ☒ Air/Water/Oil/Gas
Pop Test : PASSED		
<u>BASIS OF TEST:</u>		
Ref Standard ASME Sec VIII, Div I		
<u>Summary of Test Procedure:</u>		
The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.		
<u>Traceability:</u>		
The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.		
<u>Condition of Instrument</u>		
With in the tolerance on receipt(as found)	Yes	
The instrument was adjusted	No	
The instrument was repaired	No	
*** End of the Certificate ***		
 Calibrated By: <u> </u> Calibration Engineer	 UIC-QP-7.8-24 F02, REV 0	Approved by: <u> </u> QA/QC Engineer