



الشركة العالمية لخدمات التفتيش المحدودة

UNIVERSAL INSPECTION CO. LTD.

CR No.2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

## CERTIFICATE OF CALIBRATION

**CERTIFICATE NO: UIC/DAM/94495**

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

**Customer** **Sudhir Equipment Rental Company**  
**Location** Dammam, Kingdom of Saudi Arabia

**Description** **Pressure Relief Valve**  
**Manufacturer** HARSHAD  
**Type** N/A  
**Serial No.** 2301153  
**ID. No.** SAC-500-123  
**Location** Calibration Lab (Dammam)

**Receipt Date** : 07-Jun-26  
**Calibration Date** : 08-Jun-26  
**Calibration Due Date** : 07-Dec-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

### Reference Standard Equipment Used

| Description                             | Model / Cert. No. | Serial No   | Due Date  |
|-----------------------------------------|-------------------|-------------|-----------|
| High Pressure Pneumatic Pressure Source | SA178189-224312   | 214E2408011 | 02-Dec-26 |
| Digital Pressure Calibrator             | SA178189-224313   | 2232411017  | 02-Dec-26 |

Calibrated By  
**90**  
Calibration Engineer



Approved By  
  
QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results are valid only if used except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.  
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881  
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية  
هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨١ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١  
هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC/DAM/94495

Page

2 of 2

**Customer** **Sudhir Equipment Rental Company****Location** Dammam, Kingdom of Saudi Arabia**Description** **Pressure Relief Valve****Manufacturer** HARSHAD**Model** N/A**Serial No.** 2301153**ID. No.** SAC-500-123**Calibration Date** 08-Jun-26**Calibration Due Date** 07-Dec-26**Temperature Range** 23.5°C**Relative Humidity** RH 47%Test Results

Set Pressure : 12.00 BAR

Maximum Pressure : N/A

Back Pressure : 0.00 BAR

Size Inlet/Outlet : 1" x 2"

CDTP : 12.00 BAR

Setting Accuracy :  $\pm 3\%$  of Setting

Type of Valve : Metal Seated

Pop Test : PASSED

Test Medium : AIR

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt( as found )

No

The instrument was adjusted

Yes

The instrument was repaired

No

Calibrated By: \_\_\_\_\_

Calibration Engineer



Approved by: \_\_\_\_\_

QA/QC Engineer