



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/16708

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company**
Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**
Manufacturer Atlas Copco
Part No. 0832100066
Serial No. 2213488
ID No. SAC-500-128
Location Calibration Lab (Dammam)


Receipt Date : 05-Jun-23
Calibration Date : 05-Jun-23
Calibration Due Date : 04-Jun-24

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA 178189 - 198663	214E191209	04-Apr-24
Digital Pressure Calibrator	SA 178189 - 198664	223162191108	04-Apr-24

Calibrated By



Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٥٥١٥ ١٣ ٩٦٦

CERTIFICATE NO: UIC/DAM/16708

Page

2 of 2

Customer **Sudhir Equipment Rental Company**

Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer Atlas Copco

Part No. 0832100066

Serial No. 2213488

ID No. SAC-500-128

Calibration Date 05-Jun-23

Calibration Due Date 04-Jun-24

Temperature 23.7°C

Relative Humidity RH 48%

Test Results

Set Pressure : 16.00 Bar

Spring Range : N/A

Back Pressure : 0.00 Bar

Size Inlet / Outlet : 1" × 1½"

CDTP : 16.00 Bar

Type of Valve : Metal Seated

Setting Accuracy : ± 3 % of Setting

Pop Test : PASSED

Test Medium : Air/Water/Oil/Gas

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

No

The instrument was adjusted

Yes

The instrument was repaired

No

Calibrated By:



Calibration Engineer

***** End of the Certificate *****



Approved by:

QA/QC Engineer