



الشركة العالمية لخدمات التفتيش المحدودة

UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/69850

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Location	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Relief Valve		
Manufacturer	Harshad		
Part No.	N/A		
Serial No.	2417386	Receipt Date	: 13-Nov-25
ID. No.	SAC-250-328	Calibration Date	: 15-Nov-25
Location	Calibration Lab (Dammam)	Calibration Due Date	: 14-May-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No.	Due Date
High Pressure Pneumatic Pressure Source	AME/24/TC/1212-A7	214E2408011	13-Dec-25
Digital Pressure Calibrator	AME/24/TC/1212-A12	2232411017	13-Dec-25

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

[Signature]

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results are not to be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦

هاتف ٠٥١٥ ١٣ ٩٦٦

CERTIFICATE NO: UIC/DAM/69850

Page

2 of 2

Customer **Sudhir Equipment Rental Company****Address** Dammam, Kingdom of Saudi Arabia**Description** **Pressure Relief Valve****Manufacturer** Harshad**Model** N/A**Serial No.** 2417386**ID. No.** SAC-250-328**Calibration Date** 15-Nov-25**Calibration Due Date** 14-May-26**Temperature Range** 23.5°C**Relative Humidity** RH 48%**Test Results****Set Pressure** : 9.50 BAR**Spring Range** : N/A**Back Pressure** : 0.00 BAR**Size Inlet** : 3/4"**CDTP** : 9.50 BAR**Setting Accuracy** : $\pm 3\%$ of Setting**Type of Valve** : Metal Seated**Pop Test** : PASSED**Test Medium** : AIR**BASIS OF TEST:**

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

Yes

The instrument was adjusted

No

The instrument was repaired

No



***** End of the Certificate *****

**Calibrated By:** _____
Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by: _____
QA/QC Engineer