



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/87665

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer Sudhir Equipment Rental Company
Location Dammam, Kingdom of Saudi Arabia

Description Pressure Relief Valve
Manufacturer HARSHAD
Part No. N/A
Serial No. 2500135
ID. No. SAC-500-352
Location Calibration Lab (Dammam)

Receipt Date : 21-Apr-26
Calibration Date : 21-Apr-26
Calibration Due Date : 20-Oct-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224313	2232411017	02-Dec-26

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results are valid only if used in full without the prior permission of UIC-OP-718-24-F02 (REV.10)

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٠٥١٥ ١٣ ٩٦٦

CERTIFICATE NO: UIC/DAM/87665

Page

2 of 2

Customer **Sudhir Equipment Rental Company**
Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**
Manufacturer HARSHAD
Model N/A
Serial No. 2500135
ID. No. SAC-500-352

Calibration Date 21-Apr-26
Calibration Due Date 20-Oct-26
Temperature Range 23.5°C
Relative Humidity RH 48%

Test Results

Set Pressure : 12.00 BAR Spring Range : N/A
 Back Pressure : 0.00 BAR Size Inlet/Outlet : 1"x2"
 CDP : 12.00 BAR Type of Valve : Metal Seated
 Setting Accuracy : $\pm 3\%$ of Setting Test Medium : AIR
 Pop Test : PASSED

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)
 The instrument was adjusted
 The instrument was repaired

No
 Yes
 No

++++ End of the Certificate +++++

Calibrated By:



Calibration Engineer

UIC-QP-7.8-24-F02, REV 0

Approved by:

QA/QC Engineer