



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / 362807

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer Sudhir Equipment Rental Company

Address Dammam, Kingdom of Saudi Arabia.

Description Pressure Relief Valve

Manufacturer HARSHAD

Model/Part No. 9095799400

Serial No. 2508981

Receipt Date : 22-Jun-25

ID No. N/A

Calibration Date : 24-Jun-25

Location Calibration Lab (Jubail)

Calibration Due Date : 23-Dec-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
High Pressure Pneumatic Comparator	SA 178189-216695	214E2207005	19-Dec-25
Digital Pressure Calibrator	2024005367	710190419	14-Jul-25

Calibrated By



Calibration Engineer



Approved By

Quality/ Lab Manager

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC / 362807

Customer
Address**Sudhir Equipment Rental Company**

Dammam, Kingdom of Saudi Arabia.

Description

Pressure Relief Valve

Calibration Date

24-Jun-25

Manufacturer

HARSHAD

Calibration Due Date

23-Dec-25

Model/Part No.

9095799400

Lab Temperature

24.2 °C

Serial No.

2508981

Relative Humidity

RH 50 %

ID No.

N/A

Test Results

Set Pressure : 5.00 BARG

Spring Range : N/A

Back Pressure: 0.00 BARG

Size Inlet : 3/4"

CDTP : 5.00 BARG

Type of Valve : Metal Seated

Setting Accuracy : $\pm 3\%$ of Setting

Test Medium : Air

Pop Test : PASSED

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

No

The instrument was adjusted

No

The instrument was repaired

No

***** End of the Certificate *****

Calibrated By: _____

Calibration Engineer

UIC-QP-7.8 24 F02, REV 0

Approved by: _____

Quality / Lab Manager