



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت. ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/77400

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer Sudhir Equipment Rental Company
Location Dammam, Kingdom of Saudi Arabia

Description Pressure Gauge
Manufacturer Wika
Model N/A
Serial No. 250704-10-PG05
ID. No. N/A
Location Calibration Lab (Dammam)

Receipt Date : 15-Jan-26
Calibration Date : 17-Jan-26
Calibration Due Date : 16-Jul-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224313	2232411017	02-Dec-26



Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٣٤٤ / فاكس ٨٨٨١ ١٣ ٣٤٤
هاتف ٥١٥ ١٣ ٣٦٧

CERTIFICATE NO: UIC/DAM/77400

Page

2 of 2

Customer **Sudhir Equipment Rental Company**

Location Dammam, Kingdom of Saudi Arabia

Description **Pressure Gauge**

Manufacturer Wika

Model N/A

Serial No. 250704-10-PG05

ID. No. N/A

Calibration Date 17-Jan-26

Calibration Due Date 16-Jul-26

Temperature Range 23.4°C

Relative Humidity RH 48%

Calibration Test Results For Pressure Gauge

Min. Range: 0.0 PSI

Max. Range: 100.0 PSI

Function Tested		Gauge Pressure	
UUT Setting PSI	Actual Reading in Master PSI	Error PSI	Error (FS) (% of Range)
0.00	0.00	0.000	0.000
20.00	20.20	0.200	0.200
40.00	40.30	0.300	0.300
60.00	60.40	0.400	0.400
90.00	90.50	0.500	0.500
100.00	100.60	0.600	0.600

Calibration Procedure:

UIC/P/PPGCP Rev. 00

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated.
The item was received in good condition.



Calibrated By: _____
Calibration Engineer



++++ End of the Certificate +++++

UIC-QP-7.8-24 F02, REV 0

Approved by: _____
QA/QC Engineer