



الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/HAR/ 42977

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Address	Haradh, Kingdom Of Saudi Arabia.		
Description	Pressure Relief Valve		
Manufacturer	HARSHAD		
Part No	9095799400		
Serial No.	2408723	Receipt Date	: 24-Feb-26
ID.No.	N/A	Calibration Date	: 25-Feb-26
Location	Calibration Lab (Haradh)	Calibration Due Date	: 24-Aug-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
High Pressure Pneumatic Pressure Source	2025002700	2011H141203	29-Apr-26
Digital Pressure Calibrator	2024004506	70014391075	30-Jun-26

Calibrated By

Calibration Engineer



UIC-QP-7.8-24-F02-REV 0



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧٠٥١٥

CERTIFICATE NO: UIC/HAR/ 42977

Customer **Sudhir Equipment Rental Company**
Address Haradh, Kingdom Of Saudi Arabia.

Description **Pressure Relief Valve**

Calibration Date 25-Feb-26

Manufacturer HARSHAD

Calibration Due Date 24-Aug-26

Part No 9095799400

Temperature 23.0 ± 2°C

Serial No. 2408723

Relative Humidity RH 45% to 75%

ID.No. N/A

Test Results

Set Pressure : 9.00 BAR

Spring Range : 9.5 BAR

Back Pressure: 0.00 BAR

Size Inlet : 3/4"

CDTP : 9.00 BAR

Setting Accuracy : ± 3 % of Setting

Type of Valve : Metal Seated

Pop Test : PASSED

Test Medium : AIR

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

Yes

The instrument was adjusted

No

The instrument was repaired

No

Calibrated By:



Calibration Engineer



Approved by:

[Signature]
QA/ QC Engineer