



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت. ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / 445999

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Address	Al -Jubail, Kingdom of Saudi Arabia		
Description	Pressure Relief Valve		
Manufacturer	KUNKLE		
Type/Model.	6010		
Serial No.	242736	Receipt Date	: 18-May-26
ID No.	N/A	Calibration Date	: 18-May-26
Location	Calibration Lab (Jubail)	Calibration Due Date	: 17-Nov-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
High Pressure Pneumatic Comparator	AME/25/TC-0711-A1	214E2408017	10-Jul-26
Digital Pressure Calibrator	AME/25/CC-0703-A2	2232209008	02-Jul-26

Calibrated By

Calibration Engineer



Approved By

Quality/ Lab Manager

This certificate and the attached measurement results pertain only to the instrument calibrated.
This certificate and the attached measurement results shall not be reproduced except in full without the prior permission
P.O. Box No. 2072, Jubail - 31951, K.S.A. UIC-OP-7.8-24-F02-REV 0
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
ص.ب. ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١
هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥
www.uic.com.sa

Customer	Sudhir Equipment Rental Company		
Address	Al -Jubail, Kingdom of Saudi Arabia		
Description	Pressure Relief Valve	Calibration Date	18-May-26
Manufacturer	KUNKLE	Calibration Due Date	17-Nov-26
Type/Model.	6010	Lab Temperature	24.6°C
Serial No.	242736	Relative Humidity	RH 50 %
ID No.	N/A		

Test Results

Set Pressure :	10.00 BARG	Spring Range :	N/A
Back Pressure:	0.00 BARG	Size Inlet/Outlet :	1 1/4" X 1"
CDTP :	10.00 BARG	Type of Valve :	Metal Seated
Setting Accuracy :	± 3 % of Setting	Test Medium :	Air
Pop Test :	PASSED		

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic /pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

***** End of the Certificate *****

Calibrated By:



Calibration Engineer



UIC-OP-78-24 E02, REV 0

Approved by:

[Signature]
Quality/ Lab Manager