



الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.

س.ن. ٢٠٥٥٠١٦٤٣
CR No. 2055011643

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/79021

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company**

Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer AFS

Type A28X-16T

Serial No. 240187

ID No. SAC-185-429

Location Calibration Lab (Dammam)

Receipt Date : 31-Jan-26

Calibration Date : 01-Feb-26

Calibration Due Date : 31-Jul-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224313	2232411017	02-Dec-26

Calibrated By

Calibration Engineer

Approved By

QA/QC Engineer



The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
ص.ب. ٢٠٧٢، الجبيل - ٣١٩٥١ - المملكة العربية السعودية
هاتف: ٨٨٨٢ ١٣٣٤٤ / فاكس: ٨٨٨١ ١٣٣٤٤ +٩٦٦

www.uic.com.sa

Tel.: +966 13 367 0515

CERTIFICATE NO: UIC/DAM/79021		Page 2 of 2
Customer Address	Sudhir Equipment Rental Company Dammam, Kingdom of Saudi Arabia	
Description	Pressure Relief Valve	
Manufacturer	AFS	01-Feb-26
Type	A23X-16T	31-Jul-26
Serial No.	240187	23.5°C
ID No.	SAC-185-429	RH 46%
Test Results		
Set Pressure	: 7.00 BAR	Spring Range : N/A
Back Pressure	: 0.00 BAR	Size Inlet : 3/4"
CDTP	: 7.00 BAR	Type of Valve : Metal Seated
Setting Accuracy	: $\pm 3\%$ of Setting	Test Medium : Air
Pop Test	: PASSED	
BASIS OF TEST: Ref Standard ASME Sec VIII, Div I		
Summary of Test Procedure: The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.		
Traceability: The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.		
Condition of Instrument With in the tolerance on receipt (as found) The instrument was adjusted The instrument was repaired		
Calibrated By: Calibration Engineer		
Approved by: QA/QC Engineer		