



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/83305

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company**

Location Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer AFS

Part No. N/A

Serial No. 240133

ID. No. SAC-185-467

Location Calibration Lab (Dammam)

Receipt Date : 12-Mar-26

Calibration Date : 14-Mar-26

Calibration Due Date : 13-Sep-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
Hydraulic Pressure Comparator (Water)	SA178189-224314	990S2410010	02-Dec-26
Digital Pressure Calibrator	SA178189-224315	2232411020	02-Dec-26

Calibrated By

Calibration Engineer

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results are valid and may be used except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC/DAM/83305

Page

2 of 2

Customer Location **Sudhir Equipment Rental Company**
Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**
Manufacturer AFS
Model N/A
Serial No. 240133
ID. No. SAC-185-467

Calibration Date 14-Mar-26
Calibration Due Date 13-Sep-26
Temperature Range 23.2°C
Relative Humidity RH 49%

Test Results

Set Pressure	:	7.50	BAR	Spring Range	:	N/A
Back Pressure	:	0.00	BAR	Size Inlet	:	3/4"
CDTP	:	7.50	BAR	Type of Valve	:	Metal Seated
Setting Accuracy	:	± 3 % of Setting		Test Medium	:	Air
Pop Test : PASSED						

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

++++ End of the Certificate +++++



Calibrated By: _____
Calibration Engineer



UIC-QP-7.8-24 F02, REV 0

Approved by: _____
QA/QC Engineer