



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/14478

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Address	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Gauge		
Manufacturer	Brannan		
Model	N/A		
Serial No.	PG-4478	Receipt Date	: 19-Mar-23
ID. No.	N/A	Calibration Date	: 19-Mar-23
Location	Calibration Lab (Dammam)	Calibration Due Date	: 18-Mar-24

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	DSC-22-02517	214E201209	04-Jul-23
Digital Pressure Calibrator	DSC-22-02373	710190419	21-Jun-23

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

[Signature]

The measurement results pertain to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦

CERTIFICATE NO: UIC/DAM/14478

Page

2 of 2

Customer **Sudhir Equipment Rental Company**
Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Gauge**
Manufacturer Brannan
Model N/A
Serial No. PG-4478

Calibration Date 19-Mar-23
Calibration Due Date 18-Mar-24
Temperature Range 23.9°C
Relative Humidity Range RH 49%

CALIBRATION RESULTS FOR PRESSURE GAUGE

Min. Range: 0.0 (Psi)

Max. Range: 250.0 (Psi)

Function Tested		Gauge Pressure	
UUT Setting (Psi)	Actual Reading in Master (Psi)	Error (Psi)	Error (FS) (% of Range) (Psi)
0.00	0.00	0.00	0.000
50.00	51.70	1.70	0.680
100.00	101.10	1.10	0.440
150.00	150.60	0.60	0.240
200.00	200.40	0.40	0.160
250.00	249.80	-0.20	-0.080

Calibration Procedure:

UIC/P/PPGCP Rev. 00

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The measured value stated above is the mean of 2 readings taken.
 The results stated in this certificate relate only to the item calibrated.
 The item was received in good condition.

***** End of the Certificate *****

Calibrated By:



Calibration Engineer



UIC-QP-7.8-24 F02, REV 0

Approved by:

QA/QC Engineer