



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س. ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/ALH/54732

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer Sudhir Equipment Rental Company

Location Dammam, Kingdom of Saudi Arabia

Description Pressure Relief Valve

Manufacturer AFS

Type N/A

Serial No. 240133

ID. No. SAC-185-467

Location Calibration Lab (Al Hasa)

Receipt Date : 7-Sep-26

Calibration Date : 7-Sep-26

Calibration Due Date : 6-Mar-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-227063	214E2207012	27-Apr-27
Digital Pressure Calibrator	SA178189-227064	223210304	27-Apr-27

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer



The measurement results pertain only to the instrument calibrated.

UIC-OP-08-24-F02-REV 0

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission.

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC/ALH/54732

Page

2 of 2

Customer **Sudhir Equipment Rental Company**

Location Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer AFS

Model N/A

Serial No. 240133

ID. No. SAC-185-467

Calibration Date 7-Sep-26

Calibration Due Date 6-Mar-27

Temperature Range 23.5°C

Relative Humidity RH 47%

Test Results

Set Pressure : 7.50 BAR

Maximum Presssure : N/A

Back Pressure : 0.00 BAR

Size Inlet/Outlet : 3/4"

CDTP : 7.50 BAR

Setting Accuracy : $\pm 3\%$ of Setting

Type of Valve : Metal Seated

Pop Test : PASSED

Test Medium : AIR

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

Yes

The instrument was adjusted

No

The instrument was repaired

No



Calibrated By: _____

Calibration Engineer



++++ End of the Certificate +++++

Approved by: _____

QA/QC Engineer

UIC-OP-18-24-102, REV 0