



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/ALH/53695

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer Sudhir Equipment Rental Company

Location Al Hasa, Kingdom of Saudi Arabia.

Description Pressure Gauge

Manufacturer Veethree

Model N/A

Serial No. N/A

ID. No. SAC-825-388

Location Calibration Lab (Al Hasa)

Receipt Date : 25-Aug-26

Calibration Date : 25-Aug-26

Calibration Due Date : 24-Feb-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-227063	214E2207012	27-Apr-27
Digital Pressure Calibrator	SA178189-227064	223210304	27-Apr-27

Calibrated By



Calibration Engineer

UIC-QP-7.8-24 F02, REV 0



Approved By

QA/QC Engineer

These measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC/ALH/53695

Page

2 of 2

Customer **Sudhir Equipment Rental Company**

Location Al Hasa, Kingdom of Saudi Arabia.

Description **Pressure Gauge**

Manufacturer Veethree

Model N/A

Serial No. N/A

ID. No. SAC-825-388

Calibration Date 25-Aug-26

Calibration Due Date 24-Feb-27

Temperature Range 24.3°C

Relative Humidity RH 46%

Calibration Test Results For Pressure Gauge

Min. Range: 0.0 BAR

Max. Range: 25.0 BAR

Function Tested		Gauge Pressure	
UUT Setting	Actual Reading in Master	Error	Error (FS) (% of Range)
BAR	BAR	BAR	
0.00	0.00	0.000	0.000
5.00	5.02	0.020	0.080
10.00	10.05	0.050	0.200
15.00	15.08	0.080	0.320
20.00	20.11	0.110	0.440
25.00	25.16	0.160	0.640

Calibration Procedure:

UIC/P/PPGCP Rev. 00

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per DKD-R-6-01 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated.
The item was received in good condition.



Calibrated By: _____
Calibration Engineer



++++ End of the Certificate +++++

Approved by: _____
QA/QC Engineer

UIC QP-7.8-24 P02, REV 0