



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/JED/79904

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.


All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	SUDHIR EQUIPMENT RENTAL COMPANY.		
Address	JEDDAH - KINGDOM OF SAUDI ARABIA		
Description	SAFETY RELIEF VALVE		
Manufacturer	APPOLO VALVES		
Model	19KFEK200A		
Serial No.	N/A	Receipt Date	: 30-Apr-26
ID No.	SAC-825-021	Calibration Date	: 30-Apr-26
Location	Calibration Lab (Jeddah)	Calibration Due Date	: 29-Oct-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Certificate No.	Serial No.	Due Date
High Pressure Pneumatic Calibrator	SA19-9047-75821	214E2408004	24-Dec-26
Digital Pressure Gauge	SA19-9047-75820	2232406117	24-Dec-26

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer
PAGE : 1 of 2

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ +٩٦٦ / فاكس ٨٨٨١ ٣٤٤ ١٣ +٩٦٦
هاتف ٠٥١٥ ٣٦٧ ١٣ +٩٦٦

CERTIFICATE NO : UIC/JED/79904

PAGE

2 of 2

Customer	SUDHIR EQUIPMENT RENTAL COMPANY.		
Address	JEDDAH - KINGDOM OF SAUDI ARABIA		
Description	SAFETY RELIEF VALVE		
Manufacturer	APPOLO VALVES	Calibration Date	30-Apr-26
Model	19KFEK200A	Calibration Due Date	29-Oct-26
Serial No.	N/A	Temperature	21.6°C
ID No.	SAC-825-021	Relative Humidity	47%

Test Results

Set Pressure :	10.0 Bar	Spring No./ Press. Range : ---
Back Pressure:	0.0 Bar	Size Inlet / Outlet : 1" x 1¼"
CDTP :	10.0 Bar	Type of Valve : Metal Seated
Setting Accuracy : ± 3 % of Setting		Test Medium : Air/Gas/Water/Oil
Pop Test : PASSED		

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Safety relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed below for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

++++ End of the Certificate +++++



Calibrated By:



Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by:

QA/QC Engineer



الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/JED/79905

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **SUDHIR EQUIPMENT RENTAL COMPANY.**
Address JEDDAH - KINGDOM OF SAUDI ARABIA

Description **SAFETY RELIEF VALVE**
Manufacturer HARSHAD
Part No. 0832100066
Serial No. 2327608
ID No. SAC-825-268
Location Calibration Lab (Jeddah)

Receipt Date : 30-Apr-26
Calibration Date : 30-Apr-26
Calibration Due Date : 29-Oct-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements


Reference Standard Equipment Used

Description	Certificate No.	Serial No.	Due Date
High Pressure Pneumatic Calibrator	SA19-9047-75821	214E2408004	24-Dec-26
Digital Pressure Gauge	SA19-9047-75820	2232406117	24-Dec-26

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer
PAGE : 1 of 2

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ / فاكس ٨٨٨١ ٣٤٤ ١٣ +٩٦٦
هاتف ٠٥١٥ ٣٦٧ ١٣ +٩٦٦

CERTIFICATE NO : UIC/JED/79905		PAGE	2 of 2
Customer	SUDHIR EQUIPMENT RENTAL COMPANY.		
Address	JEDDAH - KINGDOM OF SAUDI ARABIA		
Description	SAFETY RELIEF VALVE		
Manufacturer	HARSHAD	Calibration Date	30-Apr-26
Part No.	0832100066	Calibration Due Date	29-Oct-26
Serial No.	2327608	Temperature	21.6°C
ID No.	SAC-825-268	Relative Humidity	47%
<u>Test Results</u>			
Set Pressure :	16.0 Bar	Spring No./ Press. Range : ---	
Back Pressure:	0.0 Bar	Size Inlet / Outlet : 1" x 1½"	
CDTP :	16.0 Bar	Type of Valve : Metal Seated	
Setting Accuracy : ± 3 % of Setting		Test Medium : Air/Gas/Water/Oil	
Pop Test : PASSED			
<u>BASIS OF TEST:</u>			
Ref Standard ASME Sec VIII, Div I			
<u>Summary of Test Procedure:</u>			
The Calibration of above Safety relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.			
<u>Traceability:</u>			
The reported tests were undertaken using equipment detailed below for which the company maintains Traceability to the third party.			
<u>Condition of Instrument</u>			
With in the tolerance on receipt(as found)	No		
The instrument was adjusted	No		
The instrument was repaired	No		
***** End of the Certificate *****			
			
Calibrated By:	UIC-QP-7.8-24 F02, REV 0		Approved by:
 Calibration Engineer			 QA/QC Engineer



الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/JED/79906

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **SUDHIR EQUIPMENT RENTAL COMPANY.**
Address JEDDAH - KINGDOM OF SAUDI ARABIA

Description **SAFETY RELIEF VALVE**
Manufacturer HARSHAD
Part No. 9095799400
Serial No. 2408692
ID No. SAC-825-245
Location Calibration Lab (Jeddah)

Receipt Date : 30-Apr-26
Calibration Date : 30-Apr-26
Calibration Due Date : 29-Oct-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Certificate No.	Serial No.	Due Date
High Pressure Pneumatic Calibrator	SA19-9047-75821	214E2408004	24-Dec-26
Digital Pressure Gauge	SA19-9047-75820	2232406117	24-Dec-26

Calibrated By

2

Calibration Engineer



Approved By

QA/QC Engineer

PAGE : 1 of 2

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٣٤٤ / فاكس ٨٨٨١ ١٣ ٣٤٤
هاتف ٠٥١٥ ١٣ ٣٦٧ +٩٦٦

CERTIFICATE NO : UIC/JED/79906

PAGE

2 of 2

Customer	SUDHIR EQUIPMENT RENTAL COMPANY.		
Address	JEDDAH - KINGDOM OF SAUDI ARABIA		
Description	SAFETY RELIEF VALVE		
Manufacturer	HARSHAD	Calibration Date	30-Apr-26
Part No.	9095799400	Calibration Due Date	29-Oct-26
Serial No.	2408692	Temperature	21.6°C
ID No.	SAC-825-245	Relative Humidity	47%

Test Results

Set Pressure :	9.5 Bar	Spring No./ Press. Range : ----
Back Pressure:	0.0 Bar	Size Inlet : 3/4"
CDTP :	9.5 Bar	Type of Valve : Metal Seated
Setting Accuracy : ± 3 % of Setting		Test Medium : Air/Gas/Water/Oil
Pop Test : PASSED		

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Safety relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed below for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

++++ End of the Certificate +++++



Calibrated By: _____
Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by: _____
QA/QC Engineer



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT/EQUIPMENT RECEIPT

No. 158967

التاريخ

Date 30-04-2026

Sl.No.	Description	No.of.Qty.	Inst / Equip Serial No	Make & Model	Method / Standard	Accredited /Un Accredited	validity period
1,	pressure relife valve	3	240 SAC-825-245 SAC-825-268 SAC-825-021				
الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD. DELIVERY DATE 30 APR 2026 JEDDAH BRANCH C.R: 2055011643 ٤:٥٥ P.M Total = 3							

Test or Services Required: for calibration

Approval of Authorized Representative

Name: Nadim
Company: suhail rental
Address:
Mob./ Tel. No.: 0532400728
Fax:
Sign.:

Acknowledgment of Receipt

Received by:

REVIEW:	YES	NO
STATEMENT OF CONFORMITY	☐	☐
DELIVERY DATE ACCEPTED	☐	☐
RESOURCES AVAILABLE	☐	☐
METHOD ACCEPTED	☐	☐
INST / EQUIP CONDITION ACCEPTED	☐	☐
DECISION RULE	☐	☐
REMARKS IF ANY:		
	Sign	

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881 www.ui.com.sa

Tel.: +966 13 367 0515

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف +٩٦٦١٣٣٤٤٨٨٨٢ / فاكس +٩٦٦١٣٣٤٤٨٨٨١

هاتف +٩٦٦١٣٣٦٧٠٥١٥

UIC - QP - 7.4 - 22 F03