



الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/RIY/27417

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company.**
Address Dammam, Kingdom of Saudi Arabia

Description	Pressure Relief Valve		
Manufacturer	APOLLO VALVES	Receipt Date	: 6-Jan-25
Model	19KFEK200A	Calibration Date	: 6-Jan-25
Serial No.	N/A	Report Date	: 6-Jan-25
ID. No.	SAC-825-028	Issue date	: 6-Jan-25
Location	Calibration Lab	Calibration Due Date	: 5-Jul-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model/Cert No	Serial No	Due Date
Hydraulic Ultra High Pressure Compactor	2024005369	214K2303001	30-Jul-25
Digital Pressure Gauge	2024002686	223307200706	17-Apr-25

Calibrated By
13
CAL-ENGR
Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١
هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO : UIC/RIY/27417

Customer **Sudhir Equipment Rental Company.**
Address Dammam, Kingdom of Saudi Arabia

Description	Pressure Relief Valve	Calibration Date	6-Jan-25
Manufacturer	APOLLO VALVES	Calibration Due Date	5-Jul-25
Model	19KFEK200A	Temperature	23.0 ± 2°C
Serial No.	N/A	Relative Humidity	RH 45% to 75%

Test Results

Set Pressur	12.0	BAR	Spring Range	: N/A
Back Press	0.0	BAR	Size Inlet / Outlet	: 1" X 1-1/2"
CDTP :	12.0	BAR	Type of Valve	: Metal Seated
Setting Accuracy	: ± 3 % of Setting		Test Medium	: Air/Water/Oil/Gas
Pop Test	: PASSED			

BASIS OF TEST:

Ref Standard : ASME SEC VIII DIV I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed below for which the company maintains Traceability to the third party.

Condition of Instrument


With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

++++ End of the Certificate +++++

Calibrated By: 
 Calibration Engineer



UIC-OP-7.8-2-17 REV 02

Approved by: 
 QA/QC Engineer