



الشركة العالمية لخدمات التفتيش المحدودة

UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / 436994

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer Sudhir Equipment Rental Company

Address Al Jubail, Kingdom of Saudi Arabia

Description Pressure Relief Valve

Manufacturer AFS SAFETY VALVE

Type/Model. AK281X-50T

Serial No. 220108

ID No. SAC-350-040

Location Calibration Lab (Jubail)

Receipt Date : 12-Apr-26

Calibration Date : 12-Apr-26

Calibration Due Date : 11-Oct-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
High Pressure Pneumatic Comparator	AME/25/TC-0711-A1	214E2408017	10-Jul-26
Digital Pressure Calibrator	AME/25/CC-0703-A2	2232209008	02-Jul-26

Calibrated By

57
Calibration Engineer



Approved By

Quality/ Lab Manager

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results are not to be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١
هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC / 436994

PAGE

2 OF 2

Customer **Sudhir Equipment Rental Company**
Address Al Jubail, Kingdom of Saudi Arabia

Description	Pressure Relief Valve	Calibration Date	12-Apr-26
Manufacturer	AFS SAFETY VALVE	Calibration Due Date	11-Oct-26
Type/Model.	AK281X-50T	Lab Temperature	24.6°C
Serial No.	220108	Relative Humidity	RH 50 %
ID No.	SAC-350-040		

Test Results

Set Pressure :	9.50 BARG	Spring Range :	N/A
Back Pressure:	0.00 BARG	Size Inlet/Outlet :	1/2" X 1"
CDTP :	9.50 BARG	Type of Valve :	Metal Seated
Setting Accuracy :	± 3 % of Setting	Test Medium :	Air
Pop Test :	PASSED		

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic /pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

The instrument was adjusted

The instrument was repaired

No


Yes

No

Calibrated By: 
 Calibration Engineer

***** End of the Certificate *****



Approved by: 
 Quality Lab Manager