



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/82152

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Location	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Gauge		
Manufacturer	VEETHREE		
Model	N/A		
Serial No.	N/A	Receipt Date	28-Feb-26
ID. No.	PG-825-179	Calibration Date	28-Feb-26
Location	Calibration Lab (Dammam)	Calibration Due Date	27-Aug-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements


Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224315	2232411020	02-Dec-26

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٣٤٤ / فاكس ٨٨٨١ ١٣ ٣٤٤ +٩٦٦
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦ +٩٦٦

CERTIFICATE NO: UIC/DAM/82152

Page

2 of 2

Customer **Sudhir Equipment Rental Company**

Location Dammam, Kingdom of Saudi Arabia

Description **Pressure Gauge**

Manufacturer VEETHREE

Model N/A

Serial No. N/A

ID. No. PG-825-179

Calibration Date 28-Feb-26

Calibration Due Date 27-Aug-26

Temperature Range 23.6°C

Relative Humidity RH 45%

Calibration Test Results for Pressure Gauge

Min. Range: 0.0 BARG

Max. Range: 25.0 BARG

Function Tested		Gauge Pressure	
UUT Setting BARG	Actual Reading in Master BARG	Error BARG	Error (FS) (% of Range)
0.00	0.00	0.000	0.000
5.00	4.96	-0.040	-0.160
10.00	9.95	-0.050	-0.200
15.00	14.95	-0.050	-0.200
20.00	19.93	-0.070	-0.280
25.00	24.90	-0.100	-0.400

Calibration Procedure:

UIC/P/PPGCP Rev. 00

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated.
The item was received in good condition.



Calibrated By: _____
Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by:
QA/QC Engineer