الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No. 2055011643
س. ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/115797

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	**Sudhir Equipment Rental Company**			
Location	Dammam, Kingdom of Saudi Arabia			
Description	**Pressure Relief Valve**			
Manufacturer	Atlas Copco			
Type	N/A			
Serial No.	2304705	**Receipt Date**	:	01-Nov-26
ID. No.	SAC-500-350	**Calibration Date**	:	01-Nov-26
Location	Calibration Lab (Dammam)	**Calibration Due Date**	:	30-Apr-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224313	2232411017	02-Dec-26

Calibrated By
UIC 67 CAL ENGR
Calibration Engineer

QA/QC

Approved By
QA/QC Engineer

The measurement results pertain only to the instrument calibrated.
This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa
ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+

CERTIFICATE NO: UIC/DAM/115797

Customer	**Sudhir Equipment Rental Company**		
Location	Dammam, Kingdom of Saudi Arabia		
Description	**Pressure Relief Valve**		
Manufacturer	Atlas Copco	**Calibration Date**	01-Nov-26
Type	N/A	**Calibration Due Date**	30-Apr-27
Serial No.	2304705	**Temperature Range**	23.5°C
ID. No.	SAC-500-350	**Relative Humidity**	RH 48%

Test Results

Set Pressure	:	**12.00**	**BAR**	**Spring Range : N/A**
Back Pressure	:	**0.00**	**BAR**	**Size Inlet/Outlet : 1" × 1½"**
CDTP	:	**12.00**	**BAR**	
Setting Accuracy	:	**± 3 % of Setting**		**Type of Valve : Metal Seated**
Pop Test : PASSED				**Test Medium : Air**

BASIS OF TEST:

Ref Standard ASME Sec VIII,Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pnenumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceabilty to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

+ + + + + End of the Certificate + + + + +

QA/QC — UNIVERSAL INSPECTION CO. LTD. — الشركة العالمية لخدمات التفتيش المحدودة

UIC 67 CALENGR

Calibrated By: _____
Calibration Engineer

Approved by: _____
QA/QC Engineer