الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No. 2055011643
س. ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

# CERTIFICATE OF CALIBRATION

## CERTIFICATE NO: UIC/DAM/116041

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

| | | | | |
|---|---|---|---|---|
| **Customer** | **Sudhir Equipment Rental Company** | | | |
| **Location** | Dammam, Kingdom of Saudi Arabia | | | |
| **Description** | **Pressure Gauge** | | | |
| **Manufacturer** | Powerink | | | |
| **Model** | N/A | | | |
| **Serial No.** | N/A | **Receipt Date** | : | 03-Nov-26 |
| **ID. No.** | SAC-185-425 | **Calibration Date** | : | 03-Nov-26 |
| **Location** | Calibration Lab (Dammam) | **Calibration Due Date** | : | 02-May-27 |

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

**Reference Standard Equipment Used**

| Description | Cert. No. | Serial No | Due Date |
|---|---|---|---|
| High Pressure Pneumatic Pressure Source | SA178189-224312 | 214E2408011 | 02-Dec-26 |
| Digital Pressure Calibrator | SA178189-224313 | 2232411017 | 02-Dec-26 |

Calibrated By

UIC 67 CAL ENGR

Calibration Engineer

QA/QC

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission



P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa
ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+

# CERTIFICATE NO: UIC/DAM/116041



| | | | |
|---|---|---|---|
| **Customer** | **Sudhir Equipment Rental Company** | | |
| **Location** | Dammam, Kingdom of Saudi Arabia | | |
| **Description** | **Pressure Gauge** | | |
| **Manufacturer** | Powerink | **Calibration Date** | 03-Nov-26 |
| **Model** | N/A | **Calibration Due Date** | 02-May-27 |
| **Serial No.** | N/A | **Temperature Range** | 23.5°C |
| **ID. No.** | SAC-185-425 | **Relative Humidity** | RH 48% |

## Calibration Test Results for Pressure Gauge

**Min. Range:** **0.0 PSI** **Max. Range:** **300.0 PSI**

| Function Tested | Gauge Pressure | | |
|---|---|---|---|
| UUT Setting PSI | Actual Reading in Master PSI | Error PSI | Error (FS) (% of Range) |
| 0.00 | 0.00 | 0.000 | 0.000 |
| 50.00 | 50.20 | 0.200 | 0.067 |
| 100.00 | 100.50 | 0.500 | 0.167 |
| 150.00 | 150.80 | 0.800 | 0.267 |
| 250.00 | 251.20 | 1.200 | 0.400 |
| 300.00 | 301.50 | 1.500 | 0.500 |

**Calibration Procedure:**

UIC/P/PPGCP Rev. 00

**Reference Standard:**

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

**Remarks / Deviations IF ANY:**

The results stated in this certificate relate only to the item calibrated.
The item was received in good condition.

QA/QC

UNIVERSAL INSPECTION CO. LTD.

+ + + + + End of the Certificate + + + + +

UIC 67 CAL ENGR

**Calibrated By:** _____
Calibration Engineer



**Approved by:** _____
QA/QC Engineer