

OWNERSHIP CERTIFICATE

To Whom It May Concern

This is to certify that the following equipment listed below is the sole property of Sudhir Equipment Rental Company:

Equipment	Air Compressor
Brand / Make	ATLAS COPCO
Asset ID	SAC-500-394
Model Number	XAS500-175
Serial Number	PE2560712
Engine Number	SNHM107775
Rating	500 CFM
Manufacturing Year	2025

This certification is issued to confirm the ownership status of the aforementioned equipment. Sudhir Equipment Rental Company holds full legal ownership rights over the listed equipment and retains all associated benefits and responsibilities. Should there be any queries or requests for verification regarding the ownership of the mentioned equipment, please do not hesitate to contact us at : Info@sudhirrentals.com

Sincerely,

Rajesh Thakur
Plant Manager
Sudhir Equipment Rental Company





بروج للتأمين التعاوني
Buruj Cooperative Insurance

شهادة التأمين - CERTIFICATE OF INSURANCE

Policy Holder Name	:	Sudhir Equipment Rental Company	:	اسم حامل الوثيقة
Policy Number	:	01/CPM/ 2025/000002	:	رقم الوثيقة
Policy Period	:	05/04/ 2025 to 04/04/2026	:	فترة الوثيقة
Description of Machinery / Equipment	:	ATLAS COPCO-Air Compressor-XAS500-175	:	وصف المعدة/ال لة
Chassis / Serial No.	:	PE2560712	:	رقم الهيكل/ التسلسلي
Year of Manufacturing	:	2025	:	سنة الصنع
Asset ID No.	:	SAC-500-394	:	رقم ال صل
Location	:	Kingdom of Saudi Arabia.	:	الموقع

BURUJ Cooperative Insurance Company certifies that above client is holding a Machinery/Equipment Insurance Policy for the above-mentioned Equipment in accordance with the above-mentioned Insurance Policy.

With effect from: 22-07-25 to 04/04/2026

This Certificate has been issued upon Insured's request as per all terms and conditions, exclusions, warranties mentioned in the policy.

تشهد شركة بروج للتأمين التعاوني بأن السادة أعل ه مؤمنين لديها تأمين المكان/المعدات المذكورة أعل ه وذلك حسب وثيقة التأمين المذكورة أعل ه

والتي يبدأ تاريخ سريانها من ٢٠٢٥/٠٧/٢٢ إلى ٢٠٢٦/٠٤/٠٤

تم إصدار هذه الشهادة بناءً على طلب المؤمن له وذلك حسب الشروط وال استثناءات وال شتراطات المذكورة في الوثيقة.

Important Notices:

1. This Certificate does not pay any fines, penalties, bonds or cautions that may be imposed.
2. Original certificates are only accepted.
3. Due to the importance of correct information being stated in this certificate, kindly examine the document carefully and advise the Company of any difference.
4. In the event of changes in the above information, kindly contact +966-11-2938383

مل حظات هامة:

- ١) ل تطبق هذه الشهادة التأمين على الغرامات والعقوبات المالية بما في ذلك المخالفات أو الضمانات أو الكفالات التي قد تفرض.
- ٢) تقبل الشهادة ال صلية فقط.
- ٣) نظراً ل همية صحة البيانات الخاصة بكم في هذه الشهادة، الرجاء التأكد منها وإشعار الشركة في حالة وجود أي اختل فات.
- ٤) عند الحاجة إلى تعديل البيانات أعل ه يرجى ال اتصال على شركة بروج للتأمين التعاوني ٢٩٣٨٣٨٣-١١-٩٦٦+



BURUJ Cooperative Insurance Company

P.O. Box 51855, Riyadh 11553,
National Address 3385, Al Thomamah Road, An-Nada District.
800 124 0201 | burujinsurance.com

Issue Date: 22 July 2025



YASTRONICS

مختبر ياسر ترونيكس للفحص والمعايرة

YASTRONICS LABORATORY FOR INSPECTION AND CALIBRATION



STICKER NO: **12837**

EQUIPMENT NAME	Air compressor
SWL	12 Bar
EQUIPMENT NO.	SAC-500-394
EQUIPMENT SERIAL NO.	PE2560712
DATE OF INSPECTION	18-01-26
NEXT INSPECTION DATE	17-07-26
INSPECTED BY	Ahamed. W

✉ inspection@yastronics.net

☎ 920007296

📍 KSA, Riyadh, Jarir Dist. Abi Al Haram Al Katib St.

INSPECTION CERTIFICATE

AIR COMPRESSOR

CLIENT NAME : SUDHIR EQUIPMENT RENTAL CO		WORK ORDER : WR-YAS-285	
EQUIPMENT DESCRIPTION : AIR COMPRESSOR WITH CAPACITY OF 500 CFM		CERTIFICATE NO : YAS-CRT-949	
		STICKER NO : 12837	
DATE OF INSPECTION : 18-Jan-2026		NEXT INSPECTION DUE : 17-Jul-2026	
PREVIOUS INSPECTION :		PREVIOUS INSPECTION DETAILS : NA	
MANUFACTURER : ATLAS COPCO		MODEL/TYPE : XAS500-175	
SERIAL NO : PE2560712		OWNER ID : SAC-500-394	
CHECKLIST NO. : 0775		MANUFACTURED YEAR : 2025	
TYPE OF INSPECTION : PERIODIC AND VISUAL INSPECTION		LOCATION : RIYADH	
REFERENCE STANDARDS : ISO 22484:2024			
CAPACITY		500 CFM	
WORKING PRESSURE		12 BAR	
ENGINE SPEED		2200 RPM	
ENGINE POWER		127 KW	
Is this the first examination after installation or assembly at a new site or location ?		☐ Yes ☒ No	
If the answer to the above question is YES has the equipment been installed correctly ?		☐ Yes ☒ No	
After the occurrence of exceptional circumstances ?		☐ Yes ☒ No	
Identification of any part found to have a defect which is or could become a danger to persons and a description of the defect: (If none state NONE)		NONE	
Is the above a defect which is of immediate danger to persons		☐ Yes ☒ No	
Is the above a defect which is not yet but could become a danger to persons:(If YES state the date by when)		NONE	
Particulars of any tests carried out as part of the examination: (If none state NONE)		NONE	
RESULT:		IS THIS EQUIPMENT FIT FOR PURPOSE?	
☒ PASS ☐ FAIL		☒ Yes ☐ No	
REMARKS/RECOMMENDATION THIS IS TO CERTIFY THAT THE AIR COMPRESSOR HAS BEEN FUNCTIONALLY INSPECTED IN ACCORDANCE WITH INTERNATIONALLY RECOGNIZED STANDARDS ON THE INSPECTION DATE. THE EQUIPMENT IS FREE FROM EXTERNAL DAMAGE, AND THE EQUIPMENT IS FOUND TO BE IN SATISFACTORY CONDITION.			

Approved By : ATHAR AZEEM




Authorized Inspector: AHAMED



Note : This certificate becomes void if the equipment is altered, repaired, or modified without authorization. The equipment owner is solely responsible for safe operation and maintenance. The company accepts no liability for damages or failures arising from negligence, poor maintenance, or conditions not detectable during routine inspection.

Branch of Electronics Components, co - Closed Joint Stock.co

inspection@yastronics.net
info@yastronics.net

920007296
yastronics.net



فرع شركة الرقائق الإلكترونية - شركة مساهمة مقفلة

سجل فرعي: 1010325658 رقم السجل الضريبي: 300047897800003

128379 ابي الحرم الكاتب حي جريز - الرياض 4499 المملكة العربية السعودية



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/RIY/37987

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company.		
Address	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Gauge		
Manufacturer	N/A	Receipt Date	: 18-Jan-26
Model	N/A	Calibration Date	: 18-Jan-26
Serial No.	N/A	Report Date	: 18-Jan-26
ID. No.	SAC-500-145	Issue date	: 18-Jan-26
Location	Calibration Lab	Calibration Due Date	: 17-Jul-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model/Cert No	Serial No	Due Date
High Pressure Pneumatic Calibrator	SA178189-223867	2011H13013	23-Nov-26
Digital Pressure Calibrator	SA178189-223866	223320200723	26-Nov-26

Calibrated By
13
CAL-ENGR
Calibration Engineer



Approved By

[Signature]
QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٣٤٤ / فاكس ٨٨٨١ ١٣ ٣٤٤ +٩٦٦
هاتف ٠٥١٥ ١٣ ٣٦٧ +٩٦٦

CERTIFICATE NO: UIC/RIY/37987

Customer **Sudhir Equipment Rental Company.**

Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Gauge**

Manufacturer N/A

Model N/A

Serial No. N/A

ID. No. SAC-500-145

Calibration Date 18-Jan-26

Calibration Due Date 17-Jul-26

Temperature 23.0 ± 2 °C

Relative Humidity RH 45% to 75%

CALIBRATION RESULTS FOR PRESSURE GAUGE

Min. Range: 0.0 (BAR)

Max. Range: 25.0 (BAR)

Function Tested		Gauge Pressure	
UUT Setting (BAR)	Actual reading in Master (BAR)	Error (BAR)	Error (% of Range) (FS)
0.00	0.00	0.00	0.000
5.00	4.99	-0.01	-0.040
15.00	15.10	0.10	0.400
20.00	20.18	0.18	0.720
25.00	25.33	0.33	1.320

Calibration Procedure:

UIC/P/PPGCP Rev. 02

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated
The item was received in good condition.



++++ End of the Certificate +++++

Calibrated By:



Calibration Engineer

UIC-QP-7.8-2-17,REV 02

Approved by:

QA/QC Engineer