AIR QUALITY TEST REPORT

REPORT NO: UIC/AIR/2026/2047

DATE OF REPORT: 21 OCT 2026
DATE OF TESTING: 21 OCT 2026

INTRODUCTION:

A cylinder containing breathing air was received from **M/S SUDHIR EQUIPMENT RENTAL COMPANY,** on 20th OCT 2026. It was requested by the client to carryout purity testing as per CGA Standard Grade "D" breathing air.

Customer Name	:	**SUDHIR EQUIPMENT RENTAL COMPANY,**
Location	:	**DAMMAM, KSA**
Sample Identification	:	**AIR COMPRESSOR**
Sticker Number	:	**AQT/2349**
Make	:	**ATLAS COPCO**
Model	:	**XAS500-175**
SL.No	:	**PE2260968**
ID.No	:	**SAC-500-112**

TEST PROCEDURE:

The major components of the air sample were determined by using gas chromatograph equipped with thermal conductivity and flame ionization detectors. The other impurities are analyzed by Chemical reaction method using detection tubes.

SPECIFICATIONS:

CGA Commodity Specification for Air, ANSI/CGA G-7.1-2018

RESULTS

Test Description	Unit	Results	CGA Grade - D Limits (Max)
Oxygen (O_2)	(Vol. %)	21.81	19.5 – 23.5
Carbon Dioxide (CO_2)	(Vol. %)	0.04	0.10
Carbon Monoxide (CO)	(ppm)	< 2.0	10.0
Oil Condensed	(mg/m^3)	< 1.0	< 5.0
Moisture Content	(ppm)	59	67.0
Odour	---	None	None

REMARKS:

The above test result of the air sample complies with the Compressed Gas Association Standard for Grade "D" Breathing Air for SCBA or Airline respirators. Further precautionary measure time to time, we would recommend to clean the filter tap prior to use in breathing purposes.

Tested By:

Materials Analysis Dept.

Approved By:

QA/QC Manager

UIC-AQT-R1-01/REV.NO:00

This report relates only to the sample tested and only shall be reproduced in full with written approval of UIC testing laboratory