CERTIFICATE NO : PAIC/CAL-1090

Customer	**Sudhir, Equipment Rentals Company**			
Address	Jeddah Kingdom of saudi Arabia			
Description	**Pressure Relief Valve**			
Manufacturer	AFS			
Model	AK28X-50T			
Serial No.	210349	**Calibration Date**	:	9-Sep-26
Location	Jeddah-PAK ARAB LAB	**Calibration Due Date**	:	8-Sep-27

The instrument was calibrated against the reference standard listed below, which is maintained by the PAIC Instrument & Equipment Calibration Division in compliance with ISO/IEC 17025 requirements.

Reference Standard Equipment Used

Description	Manufacture	Serial No.	Due Date
High Pressure Pneumatic Pump	SPMK	LM 11585854	1-Apr-27

Calibrated By

PAIC 2 TECHNICIAN

Calibration Engineer

Approved By

Lab Manager

Disclaimer:

Pak Arab Inspection (PAIC) performs calibration in accordance with applicable standards and best practices. PAIC shall not be liable for damage, malfunction, or incidents resulting from improper use, handling, maintenance, environmental conditions, or unauthorized adjustment after calibration. This certificate is valid only when authenticated by the official PAIC stamp and authorized signature, and any alteration or defacement shall render it invalid.

شركة باك آراب انسبكشن ليميتد
QA / QC
PAK ARAB INSPECTION LTD. CO

ISO ISO ISO

شركة باك آراب إنسبكشن ليميتد

PAK ARAB
INSPECTION LTD. CO

CERTIFICATE NO : PAIC/CAL-1090

Customer	Sudhir, Equipment Rental Co.		
Customer Adress	Jeddah-Kingdom of Saudi Arabia		
Description	**Pressure Relief Valve**		
Manufacture	AFS	**Calibration Date**	9-Sep-26
Model	AK28X-50T	**Calibration Due Date**	8-Sep-27
Serial number	210349	**Lab Temperature**	24.9 °C
Location	Jeddah-PAK ARAB LA	**Relative Humidity**	RH 50 %

CALIBRATION TEST RESULT

Set pressure :	**9.0 Bar**	**Spring Range :**	**Set Pressure: 17.0 Bar**
Back Pressure :	**0 Bar**	**Size Inlet/Outlet:**	**3/4" x 1"**
CDTP :	**9.0 Bar**		
Setting Accuracy :	**± 3 % of setting**	**Type of Valve:**	**Metal Seated**
Pop Test : Passed		**Test Medium :**	**Air**

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed below for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	**Yes**
The instrument was adjusted	**Yes**
The instrument was repaired	**No**

PAIC 2 TECHNICIAN

Calibrated By: _____
Calibration Engineer

شركة باك آراب انسبكشن ليميتد
QA / QC
PAK ARAB INSPECTION LTD. CO

Approved by: _____
Lab Manager