CERTIFICATE NO : PAIC/CAL-1082

Customer	**Sudhir, Equipment Rentals Company**			
Address	Jeddah Kingdom of saudi Arabia			
Description	**Pressure Relief Valve**			
Manufacturer	NF			
SAC No.	SAC-435-53			
Serial No.	2208997	**Calibration Date**	:	07-Sep-26
Location	Jeddah-PAK ARAB LAB	**Calibration Due Date**	:	06-Sep-27

The instrument was calibrated against the reference standard listed below, which is maintained by the PAIC Instrument & Equipment Calibration Division in compliance with ISO/IEC 17025 requirements.

Reference Standard Equipment Used

Description	Manufacture	Serial No.	Due Date
High Pressure Pneumatic Pump	SPMK	LM 11585854	01-Apr-27

Calibrated By

PAIC 3 TECHNICIAN

Calibration Engineer

Approved By

Lab Manager

Disclaimer:

Pak Arab Inspection (PAIC) performs calibration in accordance with applicable standards and best practices. PAIC shall not be liable for damage, malfunction, or incidents resulting from improper use, handling, maintenance, environmental conditions, or unauthorized adjustment after calibration. This certificate is valid only when authenticated by the official PAIC stamp and authorized signature, and any alteration or defacement shall render it invalid.

شركة باك آراب انسبكشن ليميتد
QA / QC
PAK ARAB INSPECTION LTD. CO

ISO ISO ISO

CERTIFICATE NO : PAIC/CAL-1082

Customer	Sudhir, Equipment Rental Co.		
Customer Adress	Jeddah-Kingdom of Saudi Arabia		
Description	Pressure Relief Valve		
Manufacture	NF	Calibration Date	06-Sep-26
Sac No.	SAC-435-53	Calibration Due Date	05-Sep-27
Serial number	2208997	Lab Temperature	24.4 °C
Location	Jeddah-PAK ARAB LA	Relative Humidity	RH 50 %

CALIBRATION TEST RESULT

Set pressure :	16 BARG	Spring Range :	Set Pressure: 16.0 Bar
Back Pressure :	0 BARG		
CDTP :	16 BARG		
Setting Accuracy :	± 3 % of setting	Type of Valve:	Metal Seated
Pop Test : Passed		Test Medium :	Air

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed below for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	Yes
The instrument was adjusted	Yes
The instrument was repaired	No

PAIC 3 TECHNICIAN

Calibrated By: _____
Calibration Engineer

شركة باك آراب انسبكشن ليميتد
QA / QC
PAK ARAB INSPECTION LTD. CO

Approved by: _____
Lab Manager