الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No. 2055011643
س. ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / TUR / 26224

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	**Sudhir Equipment Rental Company**			
Address	DAMMAM, KINGDOM OF SAUDI ARABIA.			
Description	**Pressure Relief Valve**			
Manufacturer	ATLAS COPCO INDIA LTD			
Model/Part No.	0832 00066			
Serial No.	2304720			
ID. No.	SAC-825-159	**Calibration Date**	:	22-Jul-26
Location	Calibration Lab (Turaif)	**Calibration Due Date**	:	21-Jul-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
PORTABLE HIGH PRESSURE PNEUMATIC HAND PUMP	VIC-25-42537	2011H13013	18-Sep-26
DIGITAL PRESSURE GAUGE	VIC-25-57999	S132001185	18-Sep-26

Calibrated By

UIC 41 CAL ENGR

Calibration Engineer

QA/QC

UIC-QP-7.8-24 F02, REV 0

Approved By

For, T. Tawfik Hussain

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa

ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+

UNIVERSAL INSPECTION CO. LTD.

الشركة العالمية لخدمات التفتيش المحدودة

CERTIFICATE NO: UIC / TUR / 26224

Customer	**Sudhir Equipment Rental Company**		
Address	DAMMAM, KINGDOM OF SAUDI ARABIA.		
Description	**Pressure Relief Valve**	**Calibration Date**	22-Jul-26
Manufacturer	ATLAS COPCO INDIA LTD	**Calibration Due Date**	21-Jul-27
Model/Part No.	0832 00066	**Lab Temperature**	23°C
Serial No.	2304720	**Relative Humidity**	45%
ID. No.	SAC-825-159		

Test Results

Set Pressure :	**9 BARG**	**Spring Range :**	----------
Back Pressure :	**0 BARG**		
CDTP :	**9 BARG**	**Size Inlet/Outlet :**	**1"ISO7RM x 1.5"BSPF**
Setting Accuracy : ± 3 % of Setting		**Type of Valve :**	**Metal Seated**
Pop Test : PASSED		**Test Medium :**	**Air**

BASIS OF TEST:

Ref Standard ASME Sec VIII,Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pnenumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed below for which the company maintains Traceabilty to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	Yes
The instrument was adjusted	No
The instrument was repaired	No

UIC 41 CAL ENGR

Calibrated By: _____
Calibration Engineer

QA/QC
UIC-QP-7.8-24 F02, REV 0

UNIVERSAL INSPECTION CO. LTD.

For, T. Tasleem Hussain

Approved by: _____
QA/QC Engineer