الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No. 2055011643 س. ت ٢٠٥٥٠١١٦٤٣

AIR QUALITY TEST REPORT

REPORT NO: UIC/AIR/2026/1685

DATE OF REPORT: 02 SEP 2026
DATE OF TESTING: 02 SEP 2026

INTRODUCTION:

A cylinder containing breathing air was received from **M/S SUDHIR EQUIPMENT RENTAL COMPANY,** on 01st SEP 2026 it was requested by the client to carryout purity testing as per CGA Standard Grade "D" breathing air.

Customer Name	:	**SUDHIR EQUIPMENT RENTAL COMPANY,**
Location	:	**TURAIF, KSA.**
Sample Identification	:	**AIR COMPRESSOR**
Sticker Number	:	**AQT/02029**
Make	:	**ATLAS COPCO**
Model	:	**XAS 825-125 SKID**
SL.No	:	**PE2460302**
ID.No	:	**SAC-825-267.**

TEST PROCEDURE:

The major components of the air sample were determined by using gas chromatograph equipped with thermal conductivity and flame ionization detectors. The other impurities are analyzed by Chemical reaction method using detection tubes.

SPECIFICATIONS:

CGA Commodity Specification for Air, ANSI/CGA G-7.1-2018

RESULTS

Test Description	Unit	Results	CGA Grade - D Limits (Max)
Oxygen (O_2)	(Vol. %)	22.27	19.5 – 23.5
Carbon Dioxide (CO_2)	(Vol. %)	0.06	0.10
Carbon Monoxide (CO)	(ppm)	< 1.0	10.0
Oil Condensed	(mg/m^3)	< 2.0	< 5.0
Moisture Content	(ppm)	59	67.0
Odour	---	None	None

REMARKS:

The above test result of the air sample complies with the Compressed Gas Association Standard for Grade "D" Breathing Air for SCBA or Airline respirators. Further precautionary measure time to time, we would recommend to clean the filter tap prior to use in breathing purposes.

Tested By:

Materials Analysis Dept.

Approved By:

QA/QC Manager

UIC-AQT-RT-01/REV.NO:00

This report relates only to the sample tested and only shall be reproduced in full with written approval of UIC testing laboratory

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa
ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+

الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
TESTING DIVISION

AIR QUALITY TESTING FOR AIR COMPRESSOR

NO.AQT/02029

CERTIFICATE NO	: UICAIR/2026/1685
EQUIPMENT ID NO	: SAC-825-267
SERIAL NUMBER	: PE2460302
SAMPLE COLLECTION DATE	: 01-SEP-2026
REPORT DATE	: 02-SEP-2026
TESTING REPORT	: PASSED
TESTED BY	: SADAM HUSSAIN

Contact No. Tel.: +966 13 344 8882, Fax: +966 13 344 8881
Website : www.uic.com.sa