الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No. 2055011643
س. ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC /MED/15167

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	**SUDHIR RENTAL**			
Address	SITE KM/73, Kingdom of Saudi Arabia			
Description	**DIESEL WELDING MACHINE**			
Manufacturer	MILLER			
Model	BIG BLUE 500X CC			
Serial No.	NF290094R			
ID No.	SWM-500-011	**Calibration Date**	:	17-Sep-26
Location	Pipeline Project (KM/73)	**Calibration Due Date**	:	16-Mar-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Certificate No.	Serial No.	Due Date
AC/DC Clamp Meter	VIC-25-58370	65900001	10-Nov-26

Calibrated By

Calibration Engineer

UIC-QP-7.8-24.F02, REV 0

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa
ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+

CERTIFICATE NO : UIC /MED/15167

Customer	**SUDHIR RENTAL**		
Address	SITE KM/73, Kingdom of Saudi Arabia		
Description	**DIESEL WELDING MACHINE**		
Manufacturer	MILLER	**Calibration Date**	17-Sep-26
Model	BIG BLUE 500X CC	**Calibration Due Date**	16-Mar-27
Serial No.	NF290094R	**Temperature**	33°C
ID No.	SWM-500-011	**Relative Humidity**	RH 13%

CALIBRATION TEST RESULTS FOR DIESEL WELDING MACHINE

DC Current:

Range (A)	UUT Setting (A)	Master Reading (A)	Error	% Error (FS)
500	100.00	100.9	0.9	0.18
	200.00	201.3	1.3	0.26
	300.00	302.8	2.8	0.56
	400.00	403.5	3.5	0.70

Power Source No-Load Voltage Accuracy Test:

Measured Output Voltage (VDC)		
Increasing	Decreasing	Mean
91.8	91.5	91.65

Power Source Load Voltage Accuracy Test:

Function Tested	UUT Setting (V)	Master Reading (V)	Error	% Error (FS)
Load Voltage	24.00	24.04	0.036	0.090
	36.00	36.14	0.140	0.350

General Information about the Calibration:

The DiEsel Welding Machine has been Calibrated on the function of Voltage and Current.

Tolerance:

As per manufacturer's specification.

Reference Standards:

IEC 60974-14

Other Tests:		
Description	**Test**	**Remarks**
Functional Checks:		
Welding Process	Check	Pass
Power Supply	Check	Pass

✦✦✦✦✦ End of the Certificate ✦✦✦✦✦

Calibrated By: Calibration Engineer

Approved by: QA/QC Engineer